

Specification Sheet

Date Issued:

Specification Number: 9015H-170S Revision: 01

24-Nov-21

1. Part Number: 9015H-170S

2. Photograph:



3. Product Information:

Description: SteriWare® ULTRA Micro Spatula 2ml – Sterile

Use: Double bagged, single use, disposable powder spatula

Material of Construction: 100% High Density Polyethylene (HDPE)

• Conforms to FDA CFR 177.1520

Conforms to EU Regulations 10/2011
Conforms to EC Regulations 1935/2004

• Conforms to USP Class VI

• Conforms to USP 661

Certified BSE/TSE free

Method of Construction: Single piece, injection moulded

Nominal Volume: 2ml

Overall Length: 175mm

Spatula Blade Length: 40mm

Blade Width: 13mm

Nominal Weight: 4g

Moulding Environment: Class 100,000 Cleanroom
Packing Environment: Class 10,000 Cleanroom

Bags: - Double Bagged, heat-sealed PE/PET bag.

- Inner bag printed with batch information

Outer Packaging: Cardboard box

Box Liner: LDPE bag

Quantity per Box: 50

Method of Sterilisation: Gamma Irradiation (25-45kGy)
Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture

ATEX: This device has not been ATEX tested



Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet

Specification Number: 9015H-170S
Revision: 01
Date Issued: 24-Nov-21

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to change.control@sampling.com

5. Country of Origin: Made and packed in the UK.

6. Repeated Use: This is a single-use, disposable product. Repeated use will lead to

failure.

7. Approval: Signed:

Name: Kuldeep Flora
Position: Quality Manager

Date: 24-Nov-21