

Certificate of Conformity

Konformitätsbescheinigung
Certificat de Conformité
Certificato di conformità
Certificado de conformidad
Certificado de Conformidade



Page 1 of 1

Part Number	9015HV-170S
Description	SteriWare® ULTRA Micro Spatula 0.3ml
Material	HDPE (High Density Polyethylene)
Certificate Issued	24 November 2021

- We confirm that the goods are manufactured with HDPE that conforms to:
 - FDA 21 CFR 177.1520
 - USP Class VI
 - EU 10/2011
 - EC 1935/2004
- Products manufactured from the HDPE polymer used to manufacture this batch have been tested and meet the requirements of USP 661.
- The HDPE resin used in the manufacture of the above products is BSE/TSE free. During manufacture this item does not come into contact with any products or processes that may pose a BSE/TSE risk.
- These goods are manufactured in a class 100,000 Cleanroom.
- These goods are packed in a class 10,000 Cleanroom.
- Each item is double bagged.
- This product is designed for single use only. It must not be used multiple times.
- The information detailed above has been given to us by our suppliers, it is passed on in good faith and is to the best of our knowledge correct.

For and on behalf of Sampling Systems Ltd.

Für und im Namen von Sampling Systems Ltd.

Pour et au nom de Sampling Systems Ltd.

Da e per conto di Sampling Systems Ltd.

En nombre y por cuenta de Sampling Systems Ltd.

Por e em nome de Sampling Systems Ltd.

Kuldeep Flora
Quality Manager

This certificate is valid for 3 years from the date given above.
Diese Bescheinigung gilt für 3 Jahre ab dem obenstehenden Datum.
Ce certificat est valable 3 ans à compter de la date indiquée ci-dessus.
Questo certificato è valido per 3 anni dalla data indicata sopra.
Este certificado es válido durante 3 años a partir de la fecha especificada más arriba.
Este certificado é válido por 3 anos a partir da data indicada acima.

