

Specification Sheet

Specification Number: Revision: Date Issued: 9045H-125S 01 27<u>-Sep-21</u>

1. Part Number:

9045H-125S

2. Photograph:





3. Product Information:

Description: Use: Material of Construction: SteriWare® ULTRA Scoop 125ml – Sterile

Double bagged, single use, disposable powder scoop

100% High Density Polyethylene (HDPE)

- Conforms to FDA CFR 177.1520
- Conforms to EU Regulations 10/2011
- Conforms to EC Regulations 1935/2004
- Conforms to USP Class VI
- Conforms to USP 661
- Certified BSE/TSE free

Single piece, injection moulded

Class 100,000 Cleanroom

Class 10,000 Cleanroom

196mm

104mm

61mm

28g

- Method of Construction:
- Overall Length:
- Handle Length:
- Maximum Width:
- Nominal Weight:
- Moulding Environment:
- Packing Environment:



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com

SteriWare is a registered Trademark belonging to Sampling Systems Ltd.



Specification Number: 9045H-125S

Revision: Date Issued:

01 27-Sep-21

Bags:	Double Bagged, heat-sealed PE/PET bag.Inner bag printed with batch information
Outer Packaging:	Cardboard box
Box Liner:	LDPE bag
Quantity per Box:	50
Method of Sterilisation:	Gamma Irradiation (25-45kGy)
Recommended Storage Conditions:	Dry and ambient temperature
Shelf Life:	5 years from the date of manufacture
ATEX:	This device has not been ATEX tested
4. Change Control:	Sampling Systems operates a policy of continual improvement. To be kept informed of all major changes to any product in the
	SteriWare range please email your contact details to <u>change.control@sampling.com</u>
5. Country of Origin:	Made and packed in the UK.
6. Repeated Use:	This is a single-use, disposable product. Repeated use will lead to failure.
7. Approval:	Signed: Name: Kuldeep Flora

15Ky

Name:	Kuldeep Flora
Position:	Quality Manager
Date:	27-Sep-21



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com