

Specification Sheet

Specification Number: Revision: Date Issued: 9088H-03S 01 21-Sep-21

1. Part Number:

9088H-03S

2. Photograph:



3. Product Information:

Description: Use: Material of Construction:

Method of Construction: Nominal Bowl Volume: Length: Bowl Length: Bowl Width: Nominal Weight: Moulding Environment:

Packing Environment:

Bags:

Outer Packaging: Box Liner:

Quantity per Box:

Quantity per Бох

Method of Sterilisation: Recommended Storage Conditions:

Shelf Life:

ATEX:

SteriWare® ULTRA Sample Spoon 3ml - Sterile Double bagged, single use, disposable spoon 100% High Density Polyethylene (HDPE) • Conforms to FDA CFR 177.1520 • Conforms to EU Regulations 10/2011

- Conforms to EC Regulations 1935/2004
- Conforms to USP Class VI
- Conforms to USP 661
- Certified BSE/TSE free

Single piece, injection moulded

3ml

173mm

36mm

25mm

4g

Class 100,000 Cleanroom

Class 10,000 Cleanroom

- Double Bagged, heat-sealed PE/PET bag.
- Inner bag printed with batch information

Cardboard box

LDPE bag

50

Gamma Irradiation (25-45kGy)

Dry and ambient temperature

5 years from the date of manufacture

This device has not been ATEX tested



Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet Specification Number: 9088H-03S Revision: 21-Sep-21 Date Issued:

01

4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to change.control@sampling.com

This is a single-use, disposable device. Repeated use will lead to

5. Country of Origin:

Made and packed in the UK.

6. Repeated Use:

7. Approval:

Signed:

failure.

Name: Kuldeep Flora Position: **Quality Manager** Date: 21-Sep-21



Sampling Systems Ltd. 4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com