


## **Specification Sheet**

Specification Number: A566-30
Revision: 02
Date Issued: 21-Feb-22

**1. Part Number:** A566-30

## 2. Photograph:





## 3. Product Information:

Description TC Container 30L

Material of Construction:

Body: 316L stainless steel
Side Handles: 304 stainless steel

Lid: 304 stainless steel 316L stainless steel

Clamp: 316L stainless steel (may be 304 stainless steel)
Gasket: Silicone (conforms to FDA CFR 177.2600)

Method of Construction: Fully welded, the weld is ground and polished

Surface Finish: Better than 0.5 microns Ra

Dimensions: A = 119mm

B = 98mm C = 602mm D = 401mm E = 303mm

Nominal Volume: 30L

Handles: Two side handles welded on to patch plates

Graduation Marks: None
Pressure Rating: None
Nominal Weight (body, lid, clamp & gasket): 6.8 Kg

BSE/TSE statement: All polishing compounds used in the manufacture of this container

are of vegetable origin

Storage Conditions: Dry, ambient conditions

Note: The neck of the container is fitted with a 4" diameter ferrule which conforms to BS4825-3

Note. The neck of the container is litted with a 4 diameter lendle which comorns to B34623-3

4. Approval:

Signed:

Name: Kuldeep Flora

Position: Quality Manager

Date: 21-Feb-22



## Sampling Systems Ltd.

4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com