


Manual

# Alpha-1-Antitrypsin

ELISA

*For the determination of alpha-1-antitrypsin in stool*

Valid from 28.12.2022



**IC6200**

Distributed by:  
Ilex Life Sciences LLC  
1465 Sand Hill Rd Ste 2018  
Candler, NC 28715  
United States  
Tel: (828) 531-9949  
Email: [info@ilexlife.com](mailto:info@ilexlife.com)  
Web: <https://ilexlife.com/>

Manufactured by:  
ImmuChrom GmbH  
Lise-Meitner-Str. 13  
64646 Heppenheim  
Germany

Table of contents	Page
<b>1. Intended use</b>	<b>2</b>
<b>2. Introduction</b>	<b>2</b>
<b>3. General notes, warnings and precautions</b>	<b>2</b>
<b>4. Material delivered in the test package</b>	<b>3</b>
<b>5. Additional special equipment</b>	<b>4</b>
<b>6. Reagent preparation</b>	<b>4</b>
<b>7. Specimen</b>	<b>5</b>
Stool samples	5
<b>8. Procedure</b>	<b>5</b>
Principle of the method	5
Sample preparation	6
<b>9. Calculation of analytical results</b>	<b>7</b>
Standard curve	7
<b>10. Internal quality control</b>	<b>8</b>
Reference values	8
<b>11. Validation data</b>	<b>8</b>
Measuring range	8
Precision and reproducibility	8
Linearity	9
Detection limit	9
Limit of quantification	9
Recovery	9
Cross reactivity	9
<b>12. Limitations of the method</b>	<b>10</b>
<b>13. Disposal</b>	<b>10</b>
<b>14. Literature references</b>	<b>10</b>

## 1. Intended use

The *ImmuChrom* ELISA Kit is intended for the quantitative determination of alpha-1-antitrypsin stool. For research use only.

## 2. Introduction

Alpha-1-antitrypsin is a 52 kD glycoprotein produced in the liver, intestinal macrophages, monocytes and intestinal epithelial cells. It belongs to the group of acute-phase proteins and is one of the most important proteinase inhibitors in serum. Alpha-1-antitrypsin inhibits the proteinases trypsin and neutrophil elastase (1). A deficiency leads to increased proteolysis. Since alpha-1-antitrypsin is not appreciably broken down or resorbed in the intestine, it can be detected in the stool if the intestinal mucosa has increased permeability ("leaky gut" syndrome) (1,2). Patients with inflammatory bowel disease show elevated levels of alpha-1-antitrypsin in their stools (4).

Taking non-steroidal anti-inflammatory drugs (e.g. aspirin, ibuprofen, diclofenac) or COX-2 inhibitors (e.g. celecoxib) can lead to enteropathies, which result in an increase in the alpha-1-antitrypsin value in the stool. Before carrying out the determination, the corresponding medication should therefore not be taken for a period of 14 days, so as not to influence the measurement of the degree of intestinal inflammation.

In the case of diarrhea, due to the dilution effect, normal values can be measured in the stool despite the presence of inflammatory processes in the intestine.

### Applications

- Enteral loss of proteins- syndrome
- "Leaky gut" syndrome
- Enterocolitides of several origin

## 3. General notes, warnings and precautions

This assay was produced and put on the market for research use only.

Individual components from different batches and test kits should not be interchanged. The expiry dates stated on the relevant packaging must be observed.

The test kit reagents contain preservatives to protect against bacterial growth. Therefore contact with the skin and/or mucous membranes should be avoided.

The test kit contains components of human origin. The starting reagents were tested for antibodies against HIV1/2, hepatitis B and anti-HCV using immunoassay methods. All parameters tested were found negative. As a precautionary measure, all test kit reagents should always be treated as potentially infectious material in accordance with health care accident prevention regulations.

The substrate TMB (tetramethylbenzidine) is toxic by ingestion and skin contact. In the event of contact with the skin, the affected area must be washed immediately with plenty of water and soap.

Avoid contact of the stop solution, which consists of acid, with the skin. It causes burns on contact. You should therefore work with protective gloves and goggles. In the event of contact, the burned area must be immediately and thoroughly rinsed with plenty of water. If necessary, a doctor should be consulted.

Adherence to the prescribed protocol for performing the test is essential. ImmuChrom GmbH assumes no liability for any damage caused by unauthorized changes in the test procedure.

The guidelines for carrying out quality control in medical laboratories must be observed. Appropriate controls must be carried along.

The reagents must not be used after the expiration date.

Wear disposable gloves when handling specimens or kit reagents and wash hands thoroughly afterwards. Do not pipette by mouth. Do not eat, drink, smoke, or put on makeup in areas where specimens or kit reagents are being handled.

Samples may contain unknown interfering substances. This can lead to false high or false low results.

#### 4. Material delivered in the test package

Article no.	Compont	Description	Amount
IC6200mtp	MTP	Mikrotiter plate coated	12 x 8 wells
IC6200wp	WASHBUF	ELISA waschbuffer conc. 10 fold	100 ml
IC6200st	STD	Standards (1.0 ml) The concentrations are given in the specification	5 vials
IC6200ko	CTRL	Control 1 and 2 (1.0 ml) The concentrations are given in the specification	1 vial each
IC6200kg	CONJ	Conjugate, peroxidase labeled antibody	15 ml
IC6200su	SUB	TMB substrate (tetramethylbenzidine)	15 ml
IC6200sp	STOPP	Stop solution	7 ml
IC7400ex	EXT	Extraction buffer	150 ml

## 5. Additional special equipment

- Centrifuge, 3000xg
- Plastic reaction vials
- Stool sample extraction vials
- Vortex mixer
- Various pipettes
- Multichannel- or multipipette
- Foil to cover the microtiter plate
- Bidest. water
- Microtiterplate shaker
- ELISA reader with filter 450 nm (reference filter 620)

## 6. Reagent preparation

**Microtiter plate (MTP).** Take the needed number of stripes and assemble them on the holder. Please take care that the plate has reached 20-30° C before usage. Stripes which are not needed yet must be stored at 2-8°C. Please do not dispose of the holder until all stripes are used.

**Wash buffer (WASHBUF).** Dilute the wash buffer concentrate 1:10 with bidest. water (1 part buffer + 9 parts bidest. water). The dilution is stable for 14 days at 2-8°C.

Important: When storing the wash buffer concentrate at 2-8°C crystallization may occur. Before dilution, all crystals must be dissolved.

It is recommended to dilute only the amount of buffer which is used to process the given samples.

All other test reagents are stable at 2-8 °C up to the date of expiry stated on the label, unless otherwise specified.

## 7. Specimen

### Stool samples

Alpha-1-antitrypsin is extracted by the extraction buffer out of the stool sample in a ratio of 1:100 (e.g. 10 mg/ml).

#### Extraction in Stool extraction vials

For the extraction stick vials could be used.

We recommend to mix **15 mg** stool with **1.5 ml** extraction buffer (EXT), then vortex it until the mixture is homogenous. Transfer the resulting slurry to a plastic vial and centrifuge it for 10 min at 3000xg.

Dilute the supernatant **1:125** with diluted wash buffer (e.g. 8  $\mu$ l supernatant + 992  $\mu$ l diluted wash buffer).

## 8. Procedure

### Principle of the method

The  $\alpha$ -1-AT ELISA test determines human alpha-1-antitrypsin according to the "sandwich"-principle.  $\alpha$ -1-AT in sample, standard and controls binds to antibodies, which are coated to the microtiter plate. After a washing step a peroxidase labeled detection antibody is added. A second washing step is followed by the addition of the substrate which is converted to a colored product by the peroxidase. The reaction is terminated by the addition of an acidic stop solution. The optical densities are read at 450 nm (against the reference wavelength 620 nm) in a microtiter plate reader. The  $\alpha$ -1-AT concentration can be calculated from the standard curve.

## Sample preparation

All reagents and samples should be at 20-30°C and mixed well before use.

The position of standards, controls and samples are noted on a protocol sheet.

### 1. Washing step

Take out the needed stripes of the microtiter plate and wash 1x with 250 µl diluted WASHBUF per cavity. Remove residual buffer by tapping the plate on absorbent paper after the washing step.

### 2. Incubation samples

Pipette **100 µl STD, CTRL** and diluted **samples** in double values in the microtiter plate.

Cover the stripes with a cover film and incubate the microtiter plate by shaking for **60 min** (20-30 °C; 400 rpm, 2 mm orbit diameter).

### 3. Washing step

Discard the content of the microwells and wash 5x with 250 µl diluted WASHBUF per cavity. Remove residual buffer by tapping the plate on absorbent paper after the last washing step.

### 4. Incubation conjugate

Pipette **100 µl CONJ** in each microwell.

Cover the stripes with a cover film and incubate the microtiter plate by shaking for **60 min** (20-30 °C; 400 rpm, 2 mm orbit diameter).

### 5. Washing step

Discard the content of the microwells and wash 5x with 250 µl diluted WASHBUF per cavity. Remove residual buffer by tapping the plate on absorbent paper after the last washing step.

### 6. Incubation substrate

Pipette **100 µl SUB** in each microwell.

Incubate by shaking for **10-15 min** in the dark (20-30 °C; 400 rpm, 2 mm orbit diameter).

### 7. Stopping reaction

Pipette **50 µl STOPP** in each microwell, mix well.

### 8. Reading

Read the absorbance at 450 nm. If the microtiterplate reader allows to use a reference wavelength use 620 nm as reference wavelength.

Reading should be done within 5 min after stopping reaction.

## 9. Calculation of analytical results

For calculating the results we recommend to use the 4-parameter Marquart algorithm.

### Stool samples


The obtained  $\alpha$ -1-AT concentration is multiplied with **12.5**.

Dilution 1: 15 mg in 1.5 ml corresponds to a factor **100** (assumption: 1 g stool = 1 ml)

Dilution 2: Factor **125** (8  $\mu$ l supernatant + 992  $\mu$ l wash buffer)

Calculation: Conc. Patient [ $\mu$ g/ml] = obtained conc. [ng/ml] x 100 x 125 / 1000

### Standard curve



The curve given above is only for demonstration. It must not be used for calculation of your samples



## 10. Internal quality control

### Reference values

Stool: < 0.27 mg/g stool

Ref: G. Beckmann (Hrsg.). Mikroökologie des Darmes  
ISBN 3-87706-521-X; S.263

We recommend, that each laboratory should develop their own normal range. The values mentioned above are only for orientation and can deviate from other published data.

## 11. Validation data

### Measuring range

The measuring range of alpha-1-antitrypsin is between a sample concentration of 33.0 - 900 µg/ml.

### Precision and reproducibility

<b>Intra-Assay CV:</b>	9.6 % (77.0 ng/ml)	[n = 6]
	8.7 % (42.6 ng/ml)	[n = 6]
	11.2 % (13.0 ng/ml)	[n = 6]
<b>Inter-Assay CV:</b>	11.9 % (75.1 ng/ml)	[n = 6]
	9.7 % (44.3 ng/ml)	[n = 6]
	14.4 % (14.6 ng/ml)	[n = 6]

## Linearity

The dilution of the samples was done with diluted wash buffer.

Sample	Dilution factor	Expected [ng/ml]	Measured [ng/ml]	Recovery [%]
1	--	--	72.8	
	1:2	36.4	31.4	86.3
	1:4	18.2	15.1	83.0
	1:8	9.1	8.3	91.2
	1:16	4.8	3.5	76.1
2	--	--	46.4	
	1:2	23.2	19.9	85.8
	1:4	11.6	10.7	92.2
	1:8	5.8	4.0	77.6
3	--	--	15.7	
	1:2	7.9	7.2	91.1
	1:4	3.9	3.2	82.1

## Detection limit

0.3 ng/ml

For the determination of the detection limit 20 replicates of the standard 0 were measured. After addition of the 3 fold standard deviation to the mean value the concentration was read from the standard curve.

## Limit of quantification

0.6 ng/ml

For the determination of the detection limit 20 replicates of the standard 0 were measured. After addition of the 10 fold standard deviation to the mean value the concentration was read from the standard curve.

## Recovery

The recovery was found between 80.3 and 105.2 %

## Cross reactivity

Cross reactivity to other plasma proteins could not be detected in stool samples.

## 12. Limitations of the method

**Stool samples** with  $\alpha$ -1-AT concentrations above the standard curve should be diluted with diluted wash buffer (WASHBUF) and measured again.

In case of diarrhea it is possible that even patients with an inflammation in the gut show normal values.

## 13. Disposal

The substrate (SUB) must be disposed as non-halogenated solvent. The stop solution (STOPP) could be neutralized with NaOH and if the pH value is neutral it can be disposed as salt solution. (**Important:** Reaction will produce heat, be careful).

Please refer to the appropriate national guidelines.

## 14. Literature references

1. Braamskamp (2010). Protein-losing enteropathy in children. Eur J Pediatr (2010) 169:1179–1185
2. Beckmann, Ruffer (2000). Mikroökologie des Darmes. ISBN 3-87706-521-X D. 262-264
3. Martin et al (2000). Gastroenterologische Aspekte in der Naturheilkunde. ISBN 3-930620-29-4 S 213-214.
4. Karbach et al (1983). Alpha 1-antitrypsin, a reliable endogenous marker for intestinal protein loss and its application in patients with Crohn's disease. Gut. 1983 Aug;24(8):718-23
5. Martin (Hrsg.), Das Standardlabor in der naturheilkundlichen Praxis, 4. Auflage, 2014 ISBN 978-3-437-56303-4
6. Kirkamm, Martin (Hrsg.), Spezielle Labordiagnostik in der naturheilkundlichen Praxis 2014, ISBN 978-3-437-56323-2
7. Stein (Hrsg), Labordiagnostik in der Gastroenterologie und Hepatologie, 2018, ISBN 978-3-8374-1562-9
8. Tibble JA, et al. 1999. High prevalence of NSAID