

## Peak Needles

7410 Coca Cola Dr. unit 107  
Hanover MD 20176  
USA

### ***Guangdong Heshan Sentai PI***

### ***astic Factory Co., Ltd.***

Add:No.WenMin,Lane,DF93,HenSan,District,Jiangmen,City,Guangdo  
ng,Province.P.R.China

### STERILIZATION CERTIFICATE

TYPE OF PRODUCT SUBJECT OF STERILIZATION:TRITON CARTRIDGES  
NEEDLE

TYPE OF STERILIZATION :EO GAS

ITEM NO:TATTOO NEEDLES

DATE OF STERILIZATION: May 2023

LOT NO: 052023



We hereby certify that the above mentioned item which were  
sterilized by EO GAS on May,2023 With the lot NO.052023

\*The attachment is the detailed parameter report of EO GAS  
sterilization

**Peak Needles**

7410 Coca Cola Dr. unit 107  
 Hanover MD 20176  
 USA

Guangdong Heshan Sentai Plastic Factory Co., Ltd.						
Add:No. WenMin, Lane, DF93, HenSan, District, Jiangmen, City, Guangdong, Province. P. R. China						
EO-Sterilization Record Per Load						
Date	2023-05-30	Order:	1	Sterilization No:	052023	
Article Name	Production No	Specification	Quantity	Required Operation During Sterilization Cycle		
TRITON Cartridge needle	052023	51X36X21 cm	1-40		Required	Actual
				Tr. C	49-54	51
				Pressure (KPa)	_68- _71 P	68 P
				Humidity(%RH)	30-85%	70
				Exposure time(min)	12:00	12:00
				Speed of Air Removal (Kpa/min)	1	2.5
				Speed of Aeration(Kpa/min)	//	//
				Weight of EO used Before(kg)	//	90
				Weight of EO used Afer(kg9)	//	60
				Weight of EO used (kg)	8	8
				Pressure Changed(KPa)	s: // E: //	S-69.8 E:34.5
						2 FO 1



