

Peak Needles

7410 Coca Cola Dr. unit 107
Hanover MD 20176
USA

Guangdong Heshan Sentai Pl

astic Factory Co., Ltd.

Add:No.WenMin,Lane,DF93,HenSan,District,Jiangmen,City,Guangdo
ng,Province.P.R.China

STERILIZATION CERTIFICATE

TYPE OF PRODUCT SUBJECT OF STERILIZATION:TRITON CARTRIDGES
NEEDLE

TYPE OF STERILIZATION :EO GAS

ITEM NO:TATTOO NEEDLES

DATE OF STERILIZATION: Jul 2022

LOT NO: 072022

We hereby certify that the above mentioned item which were
sterilized by EO GAS on Jul,2022 With the lot NO.072022

*The attachment is the detailed parameter report of EO GAS
sterilization



Peak Needles

7410 Coca Cola Dr. unit 107
 Hanover MD 20176
 USA

Guangdong Heshan Sentai Plastic Factory Co., Ltd.
 Add: No. WenMin, Lane, DF93, HenSan, District, Jiangmen, City, Guangdong, Province. P. R. China

EO-Sterilization Record Per Load

Date: 2022-07-18		Order: 4		Sterilization No: 072022		
Article Name	Production No	Specification	Quantity	Required Operation During Sterilization Cycle		
					Required	Actual
TRITON Cartridge needle	072022-1	51X36X21 cm	1-150	Tr. C	49-54	52
TRITON Cartridge needle	072022-2	51X36X21 cm	151-300	Pressure (KPa)	_68- _71 P	70 P
TRITON Cartridge needle	072022-3	51X36X21 cm	301-450	Humidity (%RH)	30-85%	50
TRITON Cartridge needle	072022-4	51X36X21 cm	451-592	Exposure time (min)	12:00	12:00
				Speed of Air Removal (Kpa/min)	1	2.5
				Speed of Aeration (Kpa/min)	//	//
				Weight of EO used Before (kg)	//	90
				Weight of EO used Afer (kg9)	//	60
				Weight of EO used (kg)	8	8
				Pressure Changed (KPa)	s: // E: //	S-69.8 E:34.5



072022-1

Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release	
Conditioning	17:00	19:00	120	Nomal		
Air Removal	19:00	19:20	20	Nomal		
Pressure Time	19:20	20:00	40	Nomal		
Eo ijection	20:00	20:30	30	Nomal		
Exposure time	20:30	8:30	720	Nomal		
Eo removal	8:30	9:30	60			
Flushing	9:30	12:00	150			

072022-2

Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release	
Conditioning	17:00	19:00	120	Nomal		
Air Removal	19:00	19:20	20	Nomal		
Pressure Time	19:20	20:00	40	Nomal		
Eo ijection	20:00	20:30	30	Nomal		
Exposure time	20:30	8:30	720	Nomal		
Eo removal	8:30	9:30	60			
Flushing	9:30	12:00	150			



072022-3						
Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release	
Conditioning	17:00	19:00	120	Nomal		
Air Removal	19:00	19:20	20	Nomal		
Pressure Time	19:20	20:00	40	Nomal		
Eo ijection	20:00	20:30	30	Nomal		
Exposure time	20:30	8:30	720	Nomal		
Eo removal	8:30	9:30	60			
Flushing	9:30	12:00	150			

072022-4						
Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release	
Conditioning	17:00	19:00	120	Nomal		
Air Removal	19:00	19:20	20	Nomal		
Pressure Time	19:20	20:00	40	Nomal		
Eo ijection	20:00	20:30	30	Nomal		
Exposure time	20:30	8:30	720	Nomal		
Eo removal	8:30	9:30	60			
Flushing	9:30	12:00	150			

