

## Peak Needles

7410 Coca Cola Dr. unit 107  
Hanover MD 20176  
USA

## *Guangdong Heshan Sentai PI*

### *astic Factory Co., Ltd.*

Add:No.WenMin,Lane,DF93,HenSan,District,Jiangmen,City,Guangdo  
ng,Province.P.R.China

### STERILIZATION CERTIFICATE

TYPE OF PRODUCT SUBJECT OF STERILIZATION:TRITON CARTRIDGES  
NEEDLE

TYPE OF STERILIZATION :EO GAS

ITEM NO:TATTOO NEEDLES

DATE OF STERILIZATION: Aug 2023

LOT NO: 082023

We hereby certify that the above mentioned item which were  
sterilized by EO GAS on Aug,2023 With the lot NO.082023

\*The attachment is the detailed parameter report of EO GAS  
sterilization



**Peak Needles**

7410 Coca Cola Dr. unit 107  
 Hanover MD 20176  
 USA

**Guangdong Heshan Sentai Plastic Factory Co., Ltd.**

Add:No. WenMin, Lane, DF93, HenSan, District, Jiangmen, City, Guangdong, Province. P. R. China

**EO-Sterilization Record Per Load**

Date: 2023-09-14		Order: 1		Sterilization No: 082023	
Article Name	Production No	Specification	Quantity	Required Operation During Sterilization Cycle	
TRITON Cartridge needle	082023	51X36X21 cm	1-74	Required	Actual
				Tr. C	49-54 52
				Pressure (KPa)	68- 71 P 70 P
				Humidity (%RH)	30-85% 70
				Exposure time (min)	12:00 12:00
				Speed of Air Removal (Kpa/min)	1 2.5
				Speed of Aeration (Kpa/min)	// // // //
				Weight of EO used Before (kg)	// 90
				Weight of EO used Afer (kg9)	// 60
				Weight of EO used (kg)	8 8
				Pressure Changed (KPa)	s: // E: // S-69.8 E:34.3
					2 FO 1



052023

Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release
Conditioning	17:00	19:00	120	Nomal	
Air Removal	19:00	19:20	20	Nomal	
Pressure Time	19:20	20:00	40	Nomal	
Eo ijection	20:00	20:30	30	Nomal	
Exposure time	20:30	8:30	720	Nomal	
Eo removal	8:30	9:30	60		
Flushing	9:30	12:00	150		

