

Peak Needles

7410 Coca Cola Dr. unit 107
Hanover MD 20176
USA

Guangdong Heshan Sentai PI

astic Factory Co., Ltd.

Add:No.WenMin,Lane,DF93,HenSan,District,Jiangmen,City,Guangdo
ng,Province.P.R.China

STERILIZATION CERTIFICATE

TYPE OF PRODUCT SUBJECT OF STERILIZATION:TRITON CARTRIDGES
NEEDLE

TYPE OF STERILIZATION :EO GAS

ITEM NO:TATTOO NEEDLES

DATE OF STERILIZATION: Sep 2022

LOT NO: 092022

We hereby certify that the above mentioned item which were
sterilized by EO GAS on Sep,2022 With the lot NO.092022



*The attachment is the detailed parameter report of EO GAS
sterilization

Peak Needles

7410 Coca Cola Dr. unit 107
 Hanover MD 20176
 USA

Guangdong Heshan Sentai Plastic Factory Co., Ltd.

Add:No. WenMin, Lane, DF93, HenSan, District, Jiangmen, City, Guangdong, Province. P. R. China

EO-Sterilization Record Per Load

Date: 2022-09-26		Order: 1		Sterilization No: 092022	
Article Name	Production No	Specification	Quantity	Required Operation During Sterilization Cycle	
TRITON Cartridge needle	092022	51X36X21 cm	1-73	Required	Actual
				Tr. C	49-54 50
				Pressure (KPa)	_68- _71 P 68 P
				Humidity (%RH)	30-85% 75
				Exposure time (min)	12:00 12:00
				Speed of Air Removal (Kpa/min)	1 2.5
				Speed of Aeration (Kpa/min)	// // // //
				Weight of EO used Before (kg)	// 90
				Weight of EO used After (kg)	// 60
				Weight of EO used (kg)	8 8
				Pressure Changed (KPa)	s: // E: // S-69.8 E:3
					2 FO 1



092022

Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release
Conditioning	17:00	19:00	120	Nomal	
Air Removal	19:00	19:20	20	Nomal	
Pressure Time	19:20	20:00	40	Nomal	
Eo ijection	20:00	20:30	30	Nomal	
Exposure time	20:30	8:30	720	Nomal	
Eo removal	8:30	9:30	60		
Flushing	9:30	12:00	150		

泰
德
藥
業
有
限
公
司
2 F0-2