



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Customer: Power BioPharms
Product identity: Strawberry Squares 1872022PWB0000455 12mg CBD 10mg D9 10mg Rosin
Client/Metric ID: .
Laboratory ID: 22-008164-0001

Summary

Potency:

Analyte per 3.75g	Result	Limits	Units	Status	
CBD per 3.75g	15.5		mg/3.75g		CBD-Total per Serving Size 16.7 mg/3.75g
CBD-A per 3.75g	1.42		mg/3.75g		
Δ8-THC per 3.75g	0.178		mg/3.75g		THC-Total per Serving Size 8.25 mg/3.75g
Δ9-THC per 3.75g	8.25		mg/3.75g		(Reported in milligrams per serving)

Residual Solvents:

All analytes passing and less than LOQ.

Pesticides:

All analytes passing and less than LOQ.

Metals:

Less than LOQ for all analytes.

Microbiology:

Less than LOQ for all analytes.



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Customer: Power BioPharms
 United States of America (USA)
Product identity: Strawberry Squares 1872022PWB0000455 12mg CBD 10mg D9 10mg Rosin
Client/Metric ID: .
Sample Date:
Laboratory ID: 22-008164-0001
Evidence of Cooling: No
Temp: 25.2 °C
Relinquished by: UPS
Serving Size #1: 3.75 g

Sample Results

Potency per 3.75g	Method: J AOAC 2015 V98-6 (mod) ^b	Units mg/se	Batch: 2206017	Analyze: 7/15/22 7:46:00 PM	
Analyte	Result	Limits	Units	LOQ	Notes
CBD per 3.75g	15.5		mg/3.75g	0.123	
CBD-A per 3.75g	1.42		mg/3.75g	0.123	
CBD-Total per 3.75g	16.7		mg/3.75g	0.231	
CBG per 3.75g	< LOQ		mg/3.75g	0.123	
CBG-A per 3.75g	< LOQ		mg/3.75g	0.123	
CBG-Total per 3.75g	< LOQ		mg/3.75g	0.230	
CBN per 3.75g	< LOQ		mg/3.75g	0.123	
Δ8-THC per 3.75g	0.178		mg/3.75g	0.123	
Δ9-THC per 3.75g	8.25		mg/3.75g	0.123	
THC-A per 3.75g	< LOQ		mg/3.75g	0.123	
THC-Total per 3.75g	8.25		mg/3.75g	0.231	

Microbiology

Analyte	Result	Limits	Units	LOQ	Batch	Analyzed Method	Status	Notes
Aerobic Plate Count	< LOQ		cfu/g	10	2205886	07/15/22 AOAC 990.12 (Petrifilm) ^P		
E.coli	< LOQ		cfu/g	10	2205884	07/15/22 AOAC 991.14 (Petrifilm) ^P		
Total Coliforms	< LOQ		cfu/g	10	2205884	07/15/22 AOAC 991.14 (Petrifilm) ^P		
Mold (RAPID Petrifilm)	< LOQ		cfu/g	10	2205885	07/16/22 AOAC 2014.05 (RAPID) ^P		
Yeast (RAPID Petrifilm)	< LOQ		cfu/g	10	2205885	07/16/22 AOAC 2014.05 (RAPID) ^P		



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Solvents											
Method: Residual Solvents by GC/MS ^b											
Units µg/g											
Batch 2206025											
Analyze 07/18/22 11:09 AM											
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
1,4-Dioxane	< LOQ	380	100	pass		2-Butanol	< LOQ	5000	200	pass	
2-Ethoxyethanol	< LOQ	160	30.0	pass		2-Methylbutane (Isopentane)	< LOQ		200		
2-Methylpentane	< LOQ		30.0			2-Propanol (IPA)	< LOQ	5000	200	pass	
2,2-Dimethylbutane	< LOQ		30.0			2,2-Dimethylpropane (neo-pentane)	< LOQ		200		
2,3-Dimethylbutane	< LOQ		30.0			3-Methylpentane	< LOQ		30.0		
Acetone	< LOQ	5000	200	pass		Acetonitrile	< LOQ	410	100	pass	
Benzene	< LOQ	2.00	1.00	pass		Butanes (sum)	< LOQ	5000	400	pass	
Cyclohexane	< LOQ	3880	200	pass		Ethyl acetate	< LOQ	5000	200	pass	
Ethyl benzene	< LOQ		200			Ethyl ether	< LOQ	5000	200	pass	
Ethylene glycol	< LOQ	620	200	pass		Ethylene oxide	< LOQ	50.0	20.0	pass	
Hexanes (sum)	< LOQ	290	150	pass		Isopropyl acetate	< LOQ	5000	200	pass	
Isopropylbenzene (Cumene)	< LOQ	70.0	30.0	pass		m,p-Xylene	< LOQ		200		
Methanol	< LOQ	3000	200	pass		Methylene chloride	< LOQ	600	60.0	pass	
Methylpropane (Isobutane)	< LOQ		200			n-Butane	< LOQ		200		
n-Heptane	< LOQ	5000	200	pass		n-Hexane	< LOQ		30.0		
n-Pentane	< LOQ		200			o-Xylene	< LOQ		200		
Pentanes (sum)	< LOQ	5000	600	pass		Propane	< LOQ	5000	200	pass	
Tetrahydrofuran	< LOQ	720	100	pass		Toluene	< LOQ	890	100	pass	
Total Xylenes	< LOQ		400			Total Xylenes and Ethyl benzene	< LOQ	2170	600	pass	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Pesticides											
Method: AOAC 2007.01 & EN 15662 (mod) ^b						Units mg/kg		Batch 2206003		Analyze 07/15/22 03:32 PM	
Analyte	Result	Limits	LOQ	Status	Notes	Analyte	Result	Limits	LOQ	Status	Notes
Abamectin ^y	< LOQ	0.50	0.250	pass		Acephate ^y	< LOQ	0.40	0.250	pass	
Acequinocyl ^y	< LOQ	2.0	1.00	pass		Acetamidprid ^y	< LOQ	0.20	0.100	pass	
Aldicarb ^y	< LOQ	0.40	0.200	pass		Azoxystrobin ^y	< LOQ	0.20	0.100	pass	
Bifenazate ^y	< LOQ	0.20	0.100	pass		Bifenthrin ^y	< LOQ	0.20	0.100	pass	
Boscalid ^y	< LOQ	0.40	0.200	pass		Carbaryl ^y	< LOQ	0.20	0.100	pass	
Carbofuran ^y	< LOQ	0.20	0.100	pass		Chlorantraniliprole ^y	< LOQ	0.20	0.100	pass	
Chlorfenapyr ^y	< LOQ	1.0	0.500	pass		Chlorpyrifos ^y	< LOQ	0.20	0.100	pass	
Clofentezine ^y	< LOQ	0.20	0.100	pass		Cyfluthrin ^y	< LOQ	1.0	0.500	pass	
Cypermethrin ^y	< LOQ	1.0	0.500	pass		Daminozide ^y	< LOQ	1.0	0.500	pass	
Diazinon ^y	< LOQ	0.20	0.100	pass		Dichlorvos ^y	< LOQ	1.0	0.500	pass	
Dimethoate ^y	< LOQ	0.20	0.100	pass		Ethoprophos ^y	< LOQ	0.20	0.100	pass	
Etofenprox ^y	< LOQ	0.40	0.200	pass		Etoxazole ^y	< LOQ	0.20	0.100	pass	
Fenoxycarb ^y	< LOQ	0.20	0.100	pass		Fenpyroximate ^y	< LOQ	0.40	0.200	pass	
Fipronil ^y	< LOQ	0.40	0.200	pass		Fonicamid ^y	< LOQ	1.0	0.400	pass	
Fludioxonil ^y	< LOQ	0.40	0.200	pass		Hexythiazox ^y	< LOQ	1.0	0.400	pass	
Imazalil ^y	< LOQ	0.20	0.100	pass		Imidacloprid ^y	< LOQ	0.40	0.200	pass	
Kresoxim-methyl ^y	< LOQ	0.40	0.200	pass		Malathion ^y	< LOQ	0.20	0.100	pass	
Metalaxyl ^y	< LOQ	0.20	0.100	pass		Methiocarb ^y	< LOQ	0.20	0.100	pass	
Methomyl ^y	< LOQ	0.40	0.200	pass		MGK-264 ^y	< LOQ	0.20	0.100	pass	
Myclobutanil ^y	< LOQ	0.20	0.100	pass		Naled ^y	< LOQ	0.50	0.250	pass	
Oxamyl ^y	< LOQ	1.0	0.500	pass		Paclotbutrazole ^y	< LOQ	0.40	0.200	pass	
Parathion-Methyl ^y	< LOQ	0.20	0.200	pass		Permethrin ^y	< LOQ	0.20	0.100	pass	
Phosmet ^y	< LOQ	0.20	0.100	pass		Piperonyl butoxide ^y	< LOQ	2.0	1.00	pass	
Prallethrin ^y	< LOQ	0.20	0.200	pass		Propiconazole ^y	< LOQ	0.40	0.200	pass	
Propoxur ^y	< LOQ	0.20	0.100	pass		Pyrethrin I (total) ^y	< LOQ	1.0	0.500	pass	
Pyridaben ^y	< LOQ	0.20	0.100	pass		Spinosad ^y	< LOQ	0.20	0.100	pass	
Spiromesifen ^y	< LOQ	0.20	0.100	pass		Spirotetramat ^y	< LOQ	0.20	0.100	pass	
Spiroxamine ^y	< LOQ	0.40	0.200	pass		Tebuconazole ^y	< LOQ	0.40	0.200	pass	
Thiacloprid ^y	< LOQ	0.20	0.100	pass		Thiamethoxam ^y	< LOQ	0.20	0.100	pass	
Trifloxystrobin ^y	< LOQ	0.20	0.100	pass							

Metals										
Analyte	Result	Limits	Units	LOQ	Batch	Analyzed	Method	Status	Notes	
Arsenic	< LOQ	0.200	mg/kg	0.0173	2206121	07/19/22	AOAC 2013.06 (mod.) ^p	pass		
Cadmium	< LOQ	0.200	mg/kg	0.0173	2206121	07/19/22	AOAC 2013.06 (mod.) ^p	pass		
Lead	< LOQ	0.500	mg/kg	0.0173	2206121	07/19/22	AOAC 2013.06 (mod.) ^p	pass		
Mercury	< LOQ	0.100	mg/kg	0.00867	2206121	07/19/22	AOAC 2013.06 (mod.) ^p	pass		



These test results are representative of the individual sample selected and submitted by the client.

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

^p = ISO/IEC 17025:2017 accredited method.

^v = TNI accredited analyte.

Units of Measure

cfu/g = Colony forming units per gram

g = g

µg/g = Microgram per gram

mg/kg = Milligram per kilogram = parts per million (ppm)

mg/3.75g = Milligram per 3.75g

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory



Derrick Tanner
General Manager



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56



**Hemp & Cannabis: Usable / Extract / Finished Product
 Chain of Custody Record**

Document Control ID: 2832 Revision: 5
 Effective: 01/04/2022

ORELAP ID: OR100028 ANAB ISO 17025 ID: AT-1508

Company: Power BioPharms Contact: Deanna Petrin Address: 8101 Royal Ridge Parkway City: Irving State: TX Zip Code: 75063 <input checked="" type="checkbox"/> Email Results: DROPBOX deanna@devmgf.com <input checked="" type="checkbox"/> Ph: (469) - 373 - 3200 <i>Billing Contact (if different)</i> Name: CC AUTH ON FILE Email: Address: City: State: Zip: Ph: () -			Analysis Requested Pesticides Oregon (P2120) <input type="checkbox"/> Residual Solvents Oregon (H0008) <input type="checkbox"/> Heavy Metals (H0013) <input type="checkbox"/> Mycotoxins (H0042) <input type="checkbox"/> Micro Profile D (M1010) <input type="checkbox"/> Terpenes (H0030) <input type="checkbox"/> Potency-Basic (H0014) <input type="checkbox"/> Potency Basic + Expanded (H0010) <input type="checkbox"/> Potency Basic + ADCs (H0015) <input type="checkbox"/> Other:							PO Number: Project ID: Batch ID: Sampled by: Custom Reporting: Source Material: <input checked="" type="checkbox"/> - Ind. Hemp product <input type="checkbox"/> - Rec. Cannabis Reporting Type: <input type="checkbox"/> - Compliance <input checked="" type="checkbox"/> - R&D Report to: <input type="checkbox"/> - METRC <input type="checkbox"/> - ODA <input type="checkbox"/> - USDA <input type="checkbox"/> - Other: Turnaround time (TAT - Business Days): <input type="checkbox"/> - 5BD <input checked="" type="checkbox"/> - 3BD* <input type="checkbox"/> - 2BD* <i>*Check for availability</i>					
Lab ID	Client Sample Identification	Sample date	Pesticides Oregon (P2120)	Residual Solvents Oregon (H0008)	Heavy Metals (H0013)	Mycotoxins (H0042)	Micro Profile D (M1010)	Terpenes (H0030)	Potency-Basic (H0014)	Potency Basic + Expanded (H0010)	Potency Basic + ADCs (H0015)	Other:	Material Type †	Weight (Units)	Comments/Metric ID
	Strawberry Squares 1872022PWB0000455 12mg CBD 10mg D9 10mg Rosin	07/07/22	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	E	45g	Contains Active as Noted
	Blueberry Squares 1872022PWB0000454 12mg CBD 10mg D9 10mg Rosin	07/07/22	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	E	45g	Please report in mg/serving
Standard Serving Sizes: D8: 3.3g D9: 3.75g HHC: 3.3g D10: 3.3g THCO: 3.3g CBD: 3.3g Mother Liquor: 3.3g Diamonds: 5g Hearts: 6g															
Signature - Relinquished By:			Date	Time	Signature - Received By:			Date	Time	Lab Use Only:					
Deanna Petrin			07/07/22		CH			7/12	1056	Shipped Via: UPS or Client drop off Evidence of cooling: <input type="checkbox"/> Yes <input checked="" type="checkbox"/> No - Temp (°C): 25.2 Sample in good condition: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No Payment: <input type="checkbox"/> Cash <input type="checkbox"/> Check <input type="checkbox"/> CC <input type="checkbox"/> Net: Prelog storage:					

† - Material Type Codes: Plant Material (P) ; Isolate (I) ; Concentrate/Extract (C) ; Tincture/Topical (T) ; Edible (E) ; Beverage (B) ; Vapor Product (V)

Samples submitted to Columbia Laboratories with testing requirements constitute an agreement for services in accordance with the current terms of service associated with this COC. By signing "Relinquished by" you are agreeing to these terms

12423 NE Whitaker Way
 Portland, OR 97230

P: (503) 254-1794 | Fax: (503) 254-1452
info@columbialaboratories.com

Page _____ of _____
www.columbialaboratories.com



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Revision: 3 Document ID: 3120

Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC 2007.1 & EN 15662		Units: mg/Kg		Batch ID: 2206003				
Method Blank		Laboratory Control Sample						
Analyte	Blank Result	Blank Limits	Notes	LCS Result	LCS Spike	LCS % Rec	Limits	Notes
Abamectin	0.000	< 0.250		0.851	1.000	85.1	50.0	150
Acephate	0.000	< 0.250		0.810	1.000	81.0	60.0	120
Acequinocyl	0.000	< 1.000		2.967	4.000	74.2	40.0	160
Acetamiprid	0.000	< 0.100		0.331	0.400	82.7	60.0	120
Aldicarb	0.000	< 0.200		0.679	0.800	84.9	60.0	120
Azoxystrobin	0.000	< 0.100		0.346	0.400	86.5	60.0	120
Bifenazate	0.000	< 0.100		0.326	0.400	81.5	60.0	120
Bifenthrin	0.000	< 0.100		0.303	0.400	75.7	50.0	150
Boscalid	0.000	< 0.200		0.618	0.800	77.2	60.0	120
Carbaryl	0.000	< 0.100		0.326	0.400	81.5	60.0	120
Carbofuran	0.000	< 0.100		0.341	0.400	85.2	60.0	120
Chlorantraniliprole	0.000	< 0.100		0.321	0.400	80.1	60.0	120
Chlorfenapyr	0.382	< 0.500		1.549	2.000	77.4	60.0	120
Chlorpyrifos	0.000	< 0.100		0.294	0.400	73.4	60.0	120
Clofentazine	0.000	< 0.100		0.253	0.400	63.2	60.0	120
Cyfluthrin	0.000	< 0.500		1.568	2.000	78.4	50.0	150
Cypermethrin	0.000	< 0.500		1.547	2.000	77.4	50.0	150
Daminozide	0.000	< 0.500		0.621	2.000	31.1	60.0	120
Diazinon	0.000	< 0.100		0.332	0.400	83.0	60.0	120
Dichlorvos	0.000	< 0.500		1.756	2.000	87.8	60.0	120
Dimethoate	0.000	< 0.100		0.334	0.400	83.5	60.0	120
Ethiofophos	0.000	< 0.100		0.306	0.400	76.4	60.0	120
Etofenprox	0.000	< 0.200		0.674	0.800	84.3	50.0	150
Etoxazole	0.000	< 0.100		0.328	0.400	82.0	60.0	120
Fenoxycarb	0.000	< 0.100		0.307	0.400	76.7	60.0	120
Fenpyroximate	0.000	< 0.200		0.606	0.800	75.8	60.0	120
Fipronil	0.002	< 0.200		0.602	0.800	75.2	60.0	120
Fonicamid	0.000	< 0.250		0.884	1.000	88.4	60.0	120
Fludioxonil	0.000	< 0.200		0.644	0.800	80.6	50.0	150
Hexythiazox	0.000	< 0.250		0.725	1.000	72.5	60.0	120
Imazalil	0.000	< 0.100		0.311	0.400	77.9	60.0	120
Imidacloprid	0.000	< 0.200		0.661	0.800	82.6	60.0	120
Kresoxim-methyl	0.000	< 0.200		0.631	0.800	78.8	60.0	120
Malathion	0.000	< 0.100		0.334	0.400	83.4	60.0	120
Metaxalyl	0.000	< 0.100		0.322	0.400	80.5	60.0	120
Methiocarb	0.000	< 0.100		0.320	0.400	80.0	60.0	120
Methomyl	0.000	< 0.200		0.730	0.800	91.3	60.0	120
MGK 264	0.000	< 0.100		0.283	0.400	70.7	50.0	150
Myclobutanil	0.000	< 0.100		0.334	0.400	83.6	60.0	120
Naled	0.000	< 0.250		0.606	1.000	60.6	50.0	150
Oxamyl	0.000	< 0.500		1.858	2.000	92.9	60.0	120
Paclobutrazole	0.000	< 0.200		0.668	0.800	83.5	60.0	120
Parathion-Methyl	0.000	< 0.200		0.722	0.800	90.2	50.0	150
Permethrin	0.000	< 0.100		0.281	0.400	70.2	50.0	150
Phosmet	0.000	< 0.100		0.308	0.400	76.9	50.0	150
Piperonyl butoxide	0.000	< 0.500		1.687	2.000	84.4	60.0	120
Prallethrin	0.000	< 0.100		0.298	0.400	74.5	60.0	120
Propiconazole	0.000	< 0.200		0.597	0.800	74.6	60.0	120
Propoxur	0.000	< 0.100		0.333	0.400	83.3	60.0	120
Pyrethrin (Summe)	0.000	< 0.100		0.341	0.413	82.6	60.0	120
Pyridaben	0.000	< 0.100		0.315	0.400	78.9	50.0	150
Spinosad	0.000	< 0.100		0.310	0.388	79.8	50.0	150
Spiromesifen	0.000	< 0.100		0.304	0.400	75.9	60.0	120
Spirotetramat	0.000	< 0.100		0.320	0.400	80.0	60.0	120
Spiroxamine	0.000	< 0.200		0.654	0.800	81.8	60.0	120
Tebuconazole	0.000	< 0.200		0.622	0.800	77.7	60.0	120
Thiacloprid	0.000	< 0.100		0.337	0.400	84.3	60.0	120
Thiamethoxam	0.000	< 0.100		0.357	0.400	89.1	60.0	120
Trifloxystrobin	0.000	< 0.100		0.323	0.400	80.8	60.0	120

Q6



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Revision: 3 Document ID: 3120
Legacy ID: CFL-C21 Worksheet Validated 10/30/2020

Laboratory Pesticide Quality Control Results

AOAC 2007.1 & EN 15662		Units: mg/Kg				Batch ID: 2206003				
Matrix Spike/Matrix Spike Duplicate Recoveries					Sample ID: 22-008103-0001					
Analyte	Result	MS Res	MSD Res	Spike	RPD%	Limit	MS % Rec	MSD % Rec	Limits	Notes
Abamectin	0.000	0.730	0.687	1.000	6.0%	< 30	73.0%	68.7%	50 - 150	
Acephate	0.000	0.988	0.978	1.000	1.0%	< 30	98.8%	97.8%	50 - 150	
Acequinocyl	0.000	4.457	4.382	4.000	1.7%	< 30	111.4%	109.5%	50 - 150	
Acetamiprid	0.000	0.415	0.398	0.400	4.2%	< 30	103.6%	99.4%	50 - 150	
Aldicarb	0.000	0.812	0.807	0.800	0.6%	< 30	101.6%	100.9%	50 - 150	
Azoxystrobin	0.000	0.424	0.390	0.400	8.4%	< 30	106.1%	97.5%	50 - 150	
Bifenazate	0.000	0.378	0.358	0.400	5.5%	< 30	94.6%	89.5%	50 - 150	
Bifenthrin	0.000	0.360	0.351	0.400	2.5%	< 30	90.1%	87.8%	50 - 150	
Boscalid	0.000	0.679	0.628	0.800	7.9%	< 30	84.9%	78.5%	50 - 150	
Carbaryl	0.000	0.428	0.393	0.400	8.4%	< 30	107.0%	98.3%	50 - 150	
Carbofuran	0.000	0.444	0.430	0.400	3.2%	< 30	111.1%	107.6%	50 - 150	
Chlorantraniliprole	0.000	0.300	0.291	0.400	3.0%	< 30	74.9%	72.7%	50 - 150	
Chlorfenapyr	0.000	1.715	1.718	2.000	0.2%	< 30	85.7%	85.9%	50 - 150	
Chlorpyrifos	0.000	0.324	0.305	0.400	6.0%	< 30	81.0%	76.2%	50 - 150	
Clofentezine	0.000	0.292	0.288	0.400	1.7%	< 30	73.1%	71.9%	50 - 150	
Cyfluthrin	0.000	1.995	1.829	2.000	8.7%	< 30	99.8%	91.5%	30 - 150	
Cypermethrin	0.000	1.927	1.851	2.000	4.0%	< 30	96.4%	92.5%	50 - 150	
Daminozide	0.000	0.834	0.830	2.000	0.5%	< 30	41.7%	41.5%	30 - 150	
Diazinon	0.000	0.365	0.364	0.400	0.2%	< 30	91.3%	91.1%	50 - 150	
Dichlorvos	0.000	1.896	1.767	2.000	7.1%	< 30	94.8%	88.3%	50 - 150	
Dimethoate	0.000	0.400	0.386	0.400	3.8%	< 30	100.1%	96.4%	50 - 150	
Ethoprophos	0.000	0.364	0.369	0.400	1.2%	< 30	91.0%	92.1%	50 - 150	
Etofenprox	0.000	0.897	0.876	0.800	2.4%	< 30	112.2%	109.5%	50 - 150	
Etoxazole	0.000	0.384	0.370	0.400	3.7%	< 30	95.9%	92.4%	50 - 150	
Fenoxycarb	0.000	0.374	0.382	0.400	2.1%	< 30	93.4%	95.4%	50 - 150	
Fenpyroximate	0.000	0.599	0.577	0.800	3.7%	< 30	74.8%	72.1%	50 - 150	
Fipronil	0.000	0.903	0.859	0.800	5.0%	< 30	112.8%	107.4%	50 - 150	
Fonicamid	0.000	0.893	0.839	1.000	6.2%	< 30	89.3%	83.9%	50 - 150	
Fludioxonil	0.000	0.696	0.638	0.800	8.7%	< 30	87.1%	79.8%	50 - 150	
Hexythiazox	0.000	1.230	1.343	1.000	8.8%	< 30	123.0%	134.3%	50 - 150	
Imazalil	0.000	0.400	0.400	0.400	0.1%	< 30	100.1%	100.0%	50 - 150	
Imidacloprid	0.000	0.521	0.496	0.800	4.8%	< 30	65.1%	62.0%	50 - 150	
Kresoxim-methyl	0.000	0.711	0.682	0.800	4.2%	< 30	88.9%	85.2%	50 - 150	
Malathion	0.000	0.381	0.349	0.400	8.7%	< 30	95.2%	87.2%	50 - 150	
Metaxalyl	0.000	0.381	0.387	0.400	1.5%	< 30	95.3%	96.8%	50 - 150	
Methiocarb	0.000	0.367	0.359	0.400	2.2%	< 30	91.6%	89.6%	50 - 150	
Methomyl	0.000	0.714	0.744	0.800	4.2%	< 30	89.2%	93.0%	50 - 150	
MGK 264	0.000	0.395	0.369	0.400	6.8%	< 30	98.7%	92.2%	50 - 150	
Myclobutanil	0.000	0.354	0.347	0.400	2.0%	< 30	88.5%	86.7%	50 - 150	
Naled	0.000	1.021	0.970	1.000	5.1%	< 30	102.1%	97.0%	50 - 150	
Oxamyl	0.000	1.784	1.912	2.000	6.9%	< 30	89.2%	95.6%	50 - 150	
Paclobutrazole	0.000	0.865	0.791	0.800	8.9%	< 30	108.1%	98.9%	50 - 150	
Parathion-Methyl	0.000	0.882	0.788	0.800	11.3%	< 30	110.2%	98.4%	30 - 150	
Permethrin	0.000	0.280	0.263	0.400	6.3%	< 30	70.0%	65.7%	50 - 150	
Phosmet	0.000	0.349	0.327	0.400	6.5%	< 30	87.3%	81.8%	50 - 150	
Piperonyl butoxide	0.000	1.951	1.887	2.000	3.3%	< 30	97.6%	94.4%	50 - 150	
Prallethrin	0.000	0.403	0.370	0.400	8.6%	< 30	100.8%	92.4%	50 - 150	
Propiconazole	0.000	0.728	0.712	0.800	2.2%	< 30	91.0%	89.0%	50 - 150	
Propoxur	0.000	0.398	0.391	0.400	1.8%	< 30	99.6%	97.8%	50 - 150	
Pyrethrin (Summe)	0.000	0.343	0.324	0.413	5.8%	< 30	83.1%	78.4%	50 - 150	
Pyridaben	0.000	0.456	0.434	0.400	5.0%	< 30	113.9%	108.4%	50 - 150	
Spinosad	0.000	0.389	0.358	0.388	8.1%	< 30	100.1%	92.4%	50 - 150	
Spiromesifen	0.000	0.613	0.557	0.400	9.7%	< 30	153.4%	139.2%	50 - 150	Q
Spirotetramat	0.000	0.251	0.243	0.400	3.4%	< 30	62.8%	60.7%	50 - 150	
Spiroxamine	0.000	0.774	0.749	0.800	3.3%	< 30	96.8%	93.6%	50 - 150	
Tebuconazole	0.000	0.716	0.702	0.800	2.0%	< 30	89.5%	87.8%	50 - 150	
Thiacloprid	0.000	0.389	0.381	0.400	2.0%	< 30	97.2%	95.3%	50 - 150	
Thiamethoxam	0.000	0.322	0.333	0.400	3.4%	< 30	80.5%	83.3%	50 - 150	
Trifloxystrobin	0.000	0.349	0.329	0.400	5.8%	< 30	87.2%	82.3%	50 - 150	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Revision: 1 Document ID: 7148
Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

Laboratory Control Sample									
Batch ID: Z206017									
Analyte	LCS	Result	Spike	Units	% Rec	Limits		Evaluation	Notes
CBDVA	1	0.0341	0.033	%	102	80.0	- 120	Acceptable	
CBDV	1	0.0379	0.033	%	114	80.0	- 120	Acceptable	
CBE	1	0.0344	0.033	%	103	80.0	- 120	Acceptable	
CBDA	1	0.0311	0.032	%	96.0	90.0	- 110	Acceptable	
CBGA	1	0.0311	0.032	%	96.8	80.0	- 120	Acceptable	
CBG	1	0.0318	0.034	%	94.3	80.0	- 120	Acceptable	
CBD	1	0.0355	0.036	%	97.5	90.0	- 110	Acceptable	
THCV	1	0.0349	0.033	%	105	80.0	- 120	Acceptable	
d8THCV	1	0.0344	0.033	%	103	80.0	- 120	Acceptable	
THCVa	1	0.0319	0.033	%	95.6	80.0	- 120	Acceptable	
CBN	1	0.0338	0.034	%	99.1	90.0	- 110	Acceptable	
exo-THC	1	0.0344	0.033	%	103	80.0	- 120	Acceptable	
d9THC	1	0.0327	0.034	%	97.2	90.0	- 110	Acceptable	
d8THC	1	0.0315	0.032	%	97.2	90.0	- 110	Acceptable	
CBL	1	0.0315	0.033	%	94.4	80.0	- 120	Acceptable	
9S-HHC	3	0.0309	0.033	%	92.6	80.0	- 120	Acceptable	
CBC	1	0.0370	0.033	%	111	80.0	- 120	Acceptable	
9R-HHC	3	0.0293	0.033	%	87.8	80.0	- 120	Acceptable	
THCA	1	0.0302	0.031	%	97.3	90.0	- 110	Acceptable	
CBGA	1	0.0322	0.033	%	96.5	80.0	- 120	Acceptable	
CBLA	1	0.0325	0.033	%	97.5	80.0	- 120	Acceptable	
d8THCO	3	0.0334	0.033	%	100	80.0	- 120	Acceptable	
CBT	1	0.0329	0.033	%	98.6	80.0	- 120	Acceptable	
d9THCO	3	0.0339	0.033	%	102	80.0	- 120	Acceptable	

Method Blank

Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.003	%	< 0.003	Acceptable	
CBDV	<LOQ	0.003	%	< 0.003	Acceptable	
CBE	<LOQ	0.003	%	< 0.003	Acceptable	
CBDa	<LOQ	0.003	%	< 0.003	Acceptable	
CBGA	<LOQ	0.003	%	< 0.003	Acceptable	
CBG	<LOQ	0.003	%	< 0.003	Acceptable	
CBD	<LOQ	0.003	%	< 0.003	Acceptable	
THCV	<LOQ	0.003	%	< 0.003	Acceptable	
d8THCV	<LOQ	0.003	%	< 0.003	Acceptable	
THCVa	<LOQ	0.003	%	< 0.003	Acceptable	
CBN	<LOQ	0.003	%	< 0.003	Acceptable	
exo-THC	<LOQ	0.003	%	< 0.003	Acceptable	
d9THC	<LOQ	0.003	%	< 0.003	Acceptable	
d8THC	<LOQ	0.003	%	< 0.003	Acceptable	
CBL	<LOQ	0.003	%	< 0.003	Acceptable	
9S-HHC	<LOQ	0.003	%	< 0.003	Acceptable	
CBC	<LOQ	0.003	%	< 0.003	Acceptable	
9R-HHC	<LOQ	0.003	%	< 0.003	Acceptable	
THCA	<LOQ	0.003	%	< 0.003	Acceptable	
CBGA	<LOQ	0.003	%	< 0.003	Acceptable	
CBLA	<LOQ	0.003	%	< 0.003	Acceptable	
d8THCO	<LOQ	0.003	%	< 0.003	Acceptable	
CBT	<LOQ	0.003	%	< 0.003	Acceptable	
d9THCO	<LOQ	0.003	%	< 0.003	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation

Units of Measure:

% - Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794

Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56



Revision: 1 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

J AOAC 2015 V98-6		Batch ID: 2206017						
Sample Duplicate		Sample ID: 22-008163-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBDV	0.00371	0.00377	0.003	%	1.59	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBD A	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBGA	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBG	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBD	1.50	1.53	0.003	%	2.50	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
d8THCV	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
THCVA	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBN	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
d9THC	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
d8THC	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
9S-HHC	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBC	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
9R-HHC	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBCA	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
d8THCO	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
CBT	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	
d9THCO	<LOQ	<LOQ	0.003	%	NA	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Revision: Document ID:
Legacy ID: Effective:

Laboratory Quality Control Results

Residual Solvents				Batch ID: 2206025					
Method Blank				Laboratory Control Sample					
Analyte	Result	LOQ	Notes	Result	Spike	Units	% Rec	Limits	Notes
Propane	ND	< 200		456	572	µg/g	79.7	60 - 120	
Isobutane	ND	< 200		515	731	µg/g	70.5	60 - 120	
Butane	ND	< 200		503	731	µg/g	68.8	60 - 120	
2,2 Dimethylpropane	ND	< 200		835	936	µg/g	89.2	60 - 120	
Methanol	ND	< 200		1630	1650	µg/g	98.8	60 - 120	
Ethylene Oxide	ND	< 30		42.7	56.2	µg/g	76.0	60 - 120	
2-Methylbutane	ND	< 200		1570	1620	µg/g	96.9	60 - 120	
Pentane	ND	< 200		1580	1610	µg/g	98.1	60 - 120	
Ethanol	ND	< 200		1570	1620	µg/g	96.9	70 - 130	
Ethyl Ether	ND	< 200		1580	1600	µg/g	98.8	60 - 120	
2,2-Dimethylbutane	ND	< 30		166	167	µg/g	99.4	60 - 120	
Acetone	ND	< 200		1570	1620	µg/g	96.9	60 - 120	
2-Propanol	ND	< 200		1570	1610	µg/g	97.5	60 - 120	
Ethyl Formate	ND	< 500		1280	1620	µg/g	79.0	70 - 130	
Acetonitrile	ND	< 100		609	635	µg/g	95.9	60 - 120	
Methyl Acetate	ND	< 500		1420	1630	µg/g	87.1	70 - 130	
2,3-Dimethylbutane	ND	< 30		176	177	µg/g	99.4	60 - 120	
Dichloromethane	ND	< 60		482	498	µg/g	96.8	60 - 120	
2-Methylpentane	ND	< 30		157	166	µg/g	94.6	60 - 120	
MTBE	ND	< 500		1390	1600	µg/g	86.9	70 - 130	
3-Methylpentane	ND	< 30		171	175	µg/g	97.7	60 - 120	
Hexane	ND	< 30		173	174	µg/g	99.4	60 - 120	
1-Propanol	ND	< 500		1460	1620	µg/g	90.1	70 - 130	
Methylethylketone	ND	< 500		1390	1600	µg/g	86.9	70 - 130	
Ethyl acetate	ND	< 200		1530	1610	µg/g	95.0	60 - 120	
2-Butanol	ND	< 200		1540	1620	µg/g	95.1	60 - 120	
Tetrahydrofuran	ND	< 100		463	507	µg/g	91.3	60 - 120	
Cyclohexane	ND	< 200		1520	1610	µg/g	94.4	60 - 120	
2 methyl 1 propanol	ND	< 500		1390	1640	µg/g	84.8	70 - 130	
Benzene	ND	< 1		4.73	5.22	µg/g	90.6	60 - 120	
Isopropyl Acetate	ND	< 200		1530	1610	µg/g	95.0	60 - 120	
Heptane	ND	< 200		1510	1610	µg/g	93.8	60 - 120	
1-Butanol	ND	< 500		1330	1610	µg/g	82.6	70 - 130	
Propyl Acetate	ND	< 500		1370	1610	µg/g	85.1	70 - 130	
1,4-Dioxane	ND	< 100		470	508	µg/g	92.5	60 - 120	
2-Ethoxyethanol	ND	< 30		158	165	µg/g	95.8	60 - 120	
Methylisobutylketone	ND	< 500		1350	1610	µg/g	83.9	70 - 130	
3-Methyl-1-butanol	ND	< 500		1320	1600	µg/g	82.5	70 - 130	
Ethylene Glycol	ND	< 200		405	492	µg/g	82.3	60 - 120	
Toluene	ND	< 100		451	497	µg/g	90.7	60 - 120	
Isobutyl Acetate	ND	< 500		1320	1610	µg/g	82.0	70 - 130	
1-Pentanol	ND	< 500		1300	1600	µg/g	81.3	70 - 130	
Butyl Acetate	ND	< 500		1320	1610	µg/g	82.0	70 - 130	
Ethylbenzene	ND	< 200		855	980	µg/g	87.2	60 - 120	
m,p-Xylene	ND	< 200		835	985	µg/g	84.8	60 - 120	
o-Xylene	ND	< 200		823	965	µg/g	85.3	60 - 120	
Cumene	ND	< 30		142	168	µg/g	84.5	60 - 120	
Anisole	ND	< 500		1150	1600	µg/g	71.9	70 - 130	
DMSO	ND	< 500		1140	1610	µg/g	70.8	70 - 130	
1,2-dimethoxyethane	ND	< 50		143	165	µg/g	86.7	70 - 130	
Triethylamine	ND	< 500		1050	1620	µg/g	64.8	70 - 130	Q6
N,N-dimethylformamide	ND	< 150		370	481	µg/g	76.9	70 - 130	
N,N-dimethylacetamide	ND	< 150		357	480	µg/g	74.4	70 - 130	
Pyridine	ND	< 50		137	171	µg/g	80.1	70 - 130	
1,2-Dichloroethane	ND	< 1		0.825	1	µg/g	82.5	70 - 130	
Chloroform	ND	< 1		0.81	1	µg/g	81.0	70 - 130	
Trichloroethylene	ND	< 1		0.749	1	µg/g	74.9	70 - 130	



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56

Revision: Document ID:
Legacy ID: Effective:

QC - Sample Duplicate			Sample ID: 22-007934-0002					
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Accept/Fail	Notes
Propane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Isobutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Butane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylpropane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Methanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethylene Oxide	ND	ND	30	µg/g	0.0	< 20	Acceptable	
2-Methylbutane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Pentane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Ether	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2,2-Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Acetone	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Propanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Ethyl Formate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Acetonitrile	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Methyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
2,3 Dimethylbutane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Dichloromethane	ND	ND	60	µg/g	0.0	< 20	Acceptable	
2-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
MTBE	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methylpentane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Hexane	ND	ND	30	µg/g	0.0	< 20	Acceptable	
1-Propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Methylethylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethyl acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-Butanol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Tetrahydrofuran	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Cyclohexane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
2-methyl-1-propanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Benzene	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Isopropyl Acetate	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Heptane	ND	ND	200	µg/g	0.0	< 20	Acceptable	
1 Butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Propyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,4-Dioxane	ND	ND	100	µg/g	0.0	< 20	Acceptable	
2 Ethoxyethanol	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Methylisobutylketone	ND	ND	500	µg/g	0.0	< 20	Acceptable	
3-Methyl-1-butanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylene Glycol	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Toluene	ND	ND	100	µg/g	0.0	< 20	Acceptable	
Isobutyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1-Pentanol	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Butyl Acetate	ND	ND	500	µg/g	0.0	< 20	Acceptable	
Ethylbenzene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
m,p-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
o-Xylene	ND	ND	200	µg/g	0.0	< 20	Acceptable	
Cumene	ND	ND	30	µg/g	0.0	< 20	Acceptable	
Anisole	ND	ND	500	µg/g	0.0	< 20	Acceptable	
DMSO	ND	ND	500	µg/g	0.0	< 20	Acceptable	
1,2-dimethoxyethane	ND	ND	50	µg/g	0.0	< 20	Acceptable	
Triethylamine	ND	ND	500	µg/g	0.0	< 20	Acceptable	
N,N-dimethylformamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
N,N-dimethylacetamide	ND	ND	150	µg/g	0.0	< 20	Acceptable	
Pyridine	ND	ND	50	µg/g	0.0	< 20	Acceptable	
1,2 Dichloroethane	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Chloroform	ND	ND	1	µg/g	0.0	< 20	Acceptable	
Trichloroethylene	ND	ND	1	µg/g	0.0	< 20	Acceptable	

Abbreviations

ND - None Detected at or above MRL
RPD - Relative Percent Difference
LOQ - Limit of Quantitation
Q6 - Quality control outside QC limits. Data acceptable based on remaining QC.

Units of Measure:

µg/g Microgram per gram or ppm



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 22-008164/D003.R000
Report Date: 07/20/2022
ORELAP#: OR100028
Purchase Order:
Received: 07/12/22 10:56





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.