

Appendix B Technical Specification

B.1 Specifications

B.1.1 Main Unit

Lead	Standard 12-lead, Nehb
Acquisition Mode	Simultaneous 12-lead
Record Format	Standard leads: 3×4, 3×4+1R, 3×4+3R, 6×2, 6×2+1R, 6×2+3R, 12×1
	Nehb lead: 6×1, 3×2
	VCG: 6×1+3, 3×2+3, 3×2+3+1R, 3×2+3+3R, Frank
Record Mode	Economic, Auto, Manual, Upload, Cycle, Trigger
Lead Format	Standard leads: 3×4, 3×4+1R, 6×2, 6×2+1R, 12×1
	Nehb lead: 6×1, 3×2
	VCG: 3×2+3, 6×1+3, Frank
Long-term Recording	Record for a long term (30 s~300 s) and rhythm analysis
Measurement Parameters	Standard leads: HR, PR interval, QRS duration, QT/QTc interval, P/QRS/T axis, RV5/SV1 voltage and RV5+SV1 voltage
	Nehb lead: HR, PR interval, P duration, T duration, QRS duration, QT/QTc interval, P/QRS/T axis, P amplitude
Filters	AC, low-pass and high-pass filters
CMRR	>89 dB
	>100 dB (with AC interference filter)
Input CIR current	≤0.1 μA
Patient Leak Current	<10 μA
Time Constant	≥3.2 s
Frequency Response	0.05 Hz~250 Hz

Noise Level	$\leq 15 \mu V_{p-v}$
Sensitivity Threshold	$20 \mu V_{p-v}$
Signal Gain	1.25 mm/mV, 2.5 mm/mV, 5 mm/mV, 10 mm/mV, 20 mm/mV, 40 mm/mV, 10/5 mm/mV, 20/10 mm/mV, Auto Gain (Auto Gain is just for the Automatic mode)
Calibration Voltage	1 mV $\pm 5\%$
Accuracy of Input Signal Reproduction	<p>Using the method described in 4.2.7.1 of AAMI EC11 to test the overall system error, which is within $\pm 5\%$;</p> <p>Using method A and D described in 4.2.7.1 of AAMI EC11 to test frequency response.</p> <p>Because of sampling characteristics and the asynchronism between sample rate and signal rate of the ECG machine, digital systems may produce a noticeable modulating effect from one cycle to the next, particularly in pediatric recordings. This phenomenon, which is not physiologic, shall be clearly described in the operator's and service manuals.</p>
Input Circuit:	Floating circuit input
Input Impedance	$\geq 2.5 M\Omega$ (full-band)
Sampling Rate of Signals	8000 Hz

B.1.2 Recorder Specification

Recorder	<p>Thermal Dot Matrix Word Printing System</p> <p>8 points/mm (perpendicular)</p> <p>40 points/mm (horizontal, 25 mm/s)</p>
Recording Paper	210mm \times 140mm-140P (recommended) or 210mm \times 150mm-140P Z-fold paper
Paper Speed	(5, 6.25, 10, 12.5, 25, 50) mm/s, $\pm 3\%$

B.1.3 Wireless Network (Optional)

Applicable Standard	IEEE 802.11b/g/n (2.4G)	IEEE 802.11a/n (5G)
Frequency Range	2.412 GHz~2.472 GHz	4.9 GHz~5.975 GHz
Band Width	20~40MHz	20~40MHz
Radiated Power	+18dBm	+13.5dBm
Signal Path	1-13 (China)	
Type and Frequency Characteristics of the Modulation	CCK/DSSS/OFDM/MCS7/MCS0	

B.1.4 Other Specification

Patient Cable	Standard 12-lead cable with defibrillation-proof
Display on LCD	1280×800, 9-inch LCD touch screen, the whole instrument work status, time, heart rate, and with the backlight
Safety Classification	IEC60601-1 Class I Type CF
AC Power Supply	100 V~240 V, 50 Hz /60 Hz, 110 VA
DC Power Supply	Rechargeable lithium battery, 14.8 V/ 4400mAh. In environment temperature ranging from 20 °C to 30 °C and with the machine turning off, the charging time is not more than 4 hours to charge the battery to 90%.
	In environment temperature ranging from 20 °C to 30 °C, the continuous working time is not less than 3 hours while the ECG device is continuously printing.

B.2 Environment Requirements

1	Transportation	
	Environment Temperature	-20 °C ~ +55 °C
	Relative Humidity	≤95 % (No condensation)
	Air Pressure	70 kPa ~ 106 kPa
	Transportation: avoid direct sunshine and rain.	
2	Storage	
	Environment Temperature	-20 °C ~ +55 °C
	Relative Humidity	≤95 % (No condensation)
	Air Pressure	70 kPa ~ 106 kPa
	The packed ECG should be stored in the well-ventilated room without corrosive gases.	
3	Using	
	Environment temperature	+5 °C ~ +40 °C
	Relative humidity	≤95 % (No condensation)
	Air pressure	70 kPa ~ 106 kPa