

WUJIANG PAILIXIN METAL PRODUCT CO, LTD

吴江派利欣金属制品有限公司

Add: No.597 SanXing Rd Economic Development Zone WuJiang JiangSu China

Tel:0512-63009978

Fax:0512-63009378

## EO-Sterilization Batch Report(环氧乙烷灭菌批记录)

Date: 2020/01/21

QR NO. QR19040


Client: Painful Pleasure

Add: 7410 Coca Dr. Unit 107, Hanover MD 21076, USA

Contact: Marc

Mobilc: 410-948-7783

E-mail: marc@painfulpleasures.com

General Information			
Sterilization No.		190930-1	
Customer name		PAINFUL PLEASURE	
Brand name		Peak	
Date of reception		2019.09.30	
Temperature of Environment		50°C	
Product name	Product type	Product Lot#	Cases
STAINLESS STEEL PIN WITH TUBE	1RL 0.20MM*B、 5RS 0.25MM etc	0901	142CTNS
PLASTIC RUBBER SLEEVE			2CTNS
Sterilization process data			
Preconditioning time		1.5-2.5hrs	
Date of sterilization		2019.09.30	
Station number		7	
Batch number		190901	
Aeration time		7 Days	
Results of Physical process control			
The physical Acceptance criteria of the preconditioning the cycl and the aeration were in compliance with the validated specifications of cycle no.: <u>GBS</u>			
Microbiological process data			
The microbiological process control with biological indicators has been performed according to the BI sterility test son(032005)			
No of biological indicators indicators 19 per chamber load.		BeiJin Four Ring Medical, Lot No.20170203	
Date of incubation		2019.09.30	
Incubation conditions		7 Days 37°C	
Result of microbiological process control			
[ ] No growth from Bacillus subtulis has been detected			
Remarks			
QA Approval			
Date of release		Signature of QA	
2019.09.30			

We hereby certify that the above mentioned goods(as described by the client)have been sterilized by "Ethylene Oxide".