

QC

Certificate of Analysis

embryotools

Parc Científic de Barcelona // Avda. Doctor Marañón, 8
08028 Barcelona, Spain
NIF B66034612 // info@embryotools.com
Phone: + 34 934 497 198

REQUESTED BY: Hunter Scientific Limited (Unit 1, Priors Hall Widdington, Saffron Walden CB11 3SB Essex United Kingdom of Great Britain and Northern Ireland)

ASSAY REQUESTED BY CUSTOMER: SMA - Sperm motility assay

OPERATION PROCEDURE: SOP-SMA-00/WI-SMA-05

TYPE OF ASSAY: Indirect

INTERNAL NUMBER: SMA.020.1632.2023

DATE: 28/06/2023 - 29/06/2023

Product information provided by the customer (Embryotools cannot be held responsible for the veracity of this information)

DESCRIPTION OF TEST PRODUCT: Borosilicate Pasteur Pipette

REF: PPB

LOT NUMBER: 5061

EXP. DATE: 2027-04

PROTOCOL:

A human sperm sample was processed in duplicate using silica-based density gradients (95%, 70% and 45% concentration) and test product incubated with previously tested medium for 60 min at 37°C. Afterwards, the resulting sperm preparation was incubated with the extracted medium (test group) or control tested medium (control group), to achieve a final concentration of approximately 10 million progressive sperms per milliliter. Test and control groups were incubated in parallel in the same incubator chamber at 37.3°C and optimal %CO₂ and %O₂, and sperm motility assessed at 0, 4, 8 and 24 hours after processing. Sperms were classified according to their mobility patterns in progressive (PS), non-progressive (NPS) and immobile (IS).

CONTROL AND TEST ASSAY RESULTS:

Motility evaluation of control and tested group.

	Motility evaluation											
	0 hours			4 hours			8 hours			24 hours		
	PS%	NPS%	IS%	PS%	NPS%	IS%	PS%	NPS%	IS%	PS%	NPS%	IS%
Control	93.05	1.90	5.05	90.95	0.90	8.15	90.05	2.90	7.05	65.80	20.45	13.75
Borosilicate Pasteur Pipette (lot.:5061)	93.15	2.25	4.60	91.75	2.25	6.00	89.65	1.35	9.00	72.80	11.55	15.65

* progressive (PS), non-progressive (NPS) and immobile (IS)

Page 1 of 2. N°. SMA.020.1632.2023



Sperm motility index (SMI) values:

	Sperm motility index			
	0 hours	4 hours	8 hours	24 hours
Borosilicate Pasteur Pipette (lot.:5061)	1.00	1.01	1.00	1.11

SUMMARY OF OBSERVATIONS: Test and control samples were incubated in the same incubator at 37.3°C with a tri-gas atmosphere with optimal %CO2 and %O2. The results of this assay refer to the items tested.

Sperm motility index (SMI) corresponds to the value obtained by dividing the percentage of motile sperms in the test group by the percentage of motile sperms in the control group.
Values <0.75 by 8 h post incubation are considered as indicators of poor assisted reproduction outcomes. This cut-off value has been used as Embryotools reference to indicate sperm toxicity.

These results are representative of the test samples submitted by the customer.

Nuno Costa-Borges, PhD

Scientific Director

Gloria Calderón, PhD

Quality Assurance