



Test Report No.T32120240597SN Date: Apr 26, 2021 Page 1 of 6

MASKON LIMITED

RM 18, BLK E, 10/F, LEE SUM FACTORY BLDG, 21-25 NG FONG ST., SAN PO KONG, KOWLOON, HONG KONG

The following samples were submitted and identified by/on behalf of the client as:

MEDICAL FACE MASK

Case No. : CA321202436929

Lot No. / Batch Code : NOT PROVIDED

Sample Description : BEIGE MASK

Country of Origin : HONG KONG

Sample Receiving Date : MAR 24, 2021

Testing Period : MAR 24, 2021 - APR 26, 2021

	~ 10
Test Requested	Conclusion
EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods (Excluded Clause 5.2.6 and Clause 6)	PASS (Type IIR)

**** FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S) ********

Signed for and on behalf of SGS Hong Kong Ltd.

Au Kam Chi, Gigi Technical Manager Signed for and on behalf of SGS Hong Kong Ltd.

30

Wong Kin Man, Gilman Technical Development Manager Signed for and on behalf of SGS Hong Kong Ltd.

Tsang Chuk Hai Senior Microbiologist

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.aga.com/wnTerms-and-Conditions for electronic focuments, subject to Terms and Conditions for Electronic Documents at http://www.aga.com/wnTerms-and-Conditions for electronic Documents agas, Affairston is drawn to the limitation of liability indemnification and jurisdiction is used self-and statement. Any holder of this document is advised that information contained hereon reflects the Company's Information or its intervention only and within the limits of Citient's Instructions, if any, The Company's ade responsibility is to its Citient and this document does not account the company and the company and the representation of the company and the compan

inless officewise stated the results shown in this test report refer only to the sample(s) tested and auch sample(s) are retained for 30 days only





Test Report No.T32120240597SN Date: Apr 26, 2021 Page 2 of 6

EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

Scope : This document specifies construction, design, performance requirements and

test methods for medical face masks intended to limit the transmission of infective agents from staff to patients during surgical procedures and other medical settings with similar requirements. A medical face mask with an appropriate microbial barrier can also be effective in reducing the emission of infective agents from the nose and mouth of an asymptomatic carrier or a patient with clinical symptoms. This European Standard is not applicable to

masks intended exclusively for the personal protection of staff.

Number of Specimen : 200 pcs of complete product

Clause	Test Items/requirement Requirements	Test Result Summary
5 5.1	General	
5.1.1	Materials and construction	PASS The mask is composed of a filter layer that is bonded between layers of fabric. The mask was not disintegrated, split or tear during intended use, and no objectionable matter was observed
.ask <u>(</u>)n	Design IVLASKOn	by visual assessment. ask on. of PASS
M°	$MaskOn.$ by M°	Length: 17.3 cm Width: 9.6 cm (Folded); 15.9 cm (Expanded)
5.2	Performance requirements	222
5.2.2^ a.al.25.2.3^	Bacterial filtration efficiency (BFE) Breathability (Differential Pressure)	> 98% () /< 60 Pa/cm²
5.2.4	Splash resistance	Penetration not seen at 16.0 kPa
5.2.5^	Microbial cleanliness (Bioburden)	≤ 30 cfu/g
5.2.6	Biocompatibility	Not Conducted as per client requested
5.2.7	Summary of performance requirements	See Table 1
<u>6</u>	Marking, labelling and packaging	Not Conducted as per client requested

This document is leaved by the Company adulyct to its General Conditions of Service printed overleaf, validable on request or accessible at http://www.ags.com/wh/Terms-and-Conditions argue and, for electron format documents, audiget to Terms and Conditions for Electronic Documents at http://www.ags.com/wh/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions/Terms-and-Conditions-and-Conditi

Unless otherwise stated the results shown in this test report refer only to the eample(s) tested and such sample(s) are retained for 30 days only





Test Report No.T32120240597SN Date: Apr 26, 2021 Page 3 of 6

EN 14683:2019+AC:2019 Medical face masks - Requirements and test methods

Table 1 Performance requirements for medical face masks

Characteristics	Type I a	Type II	Type IIR
Bacterial filtration efficiency (BFE), %	≥ 95	≥ 98	≥ 98
Differential pressure, Pa/cm ²	< 40	< 40	< 60
Splash resistance (kPa) #	Not Required	Not Required	≥ 16.0
Microbial cleanliness (cfu/g)	≤ 30	≤ 30	≤ 30

- Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.
- * An acceptable quality limit of 4,0 % is met for a single sampling plan when 29 or more of the 32 tested specimens show "pass" results.

Note:

- ^ Results of compliance for tests requested is justified according to decision rule based on the non-binary statement with guard band (is equal to the expanded measurement uncertainty with a 95% coverage probability, w = U₈₀) as stated in ILAC-G8:09/2019 Clause 4.2.3.

 - "Fail One or more measured values were observed out of tolerance at the points tested". The specific false reject risk is up to 2.5%.

This document is issued by the Company subject to its General Conditions of Service printed overleaf, shallable on request or accessible at http://www.aga.com/en/Terms-and-Conditions appr and, for electronic format documents, subject to Terms and Conditions for Electronic Cocuments at http://www.aga.com/en/Terms-and-Conditions/Terms-and-Condit

Unises otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

SSS Horg Kang Limited | Office: Units 303 & 305, 3/F Building 22E, Phase 3, HK Science Park, New Tentories, Hong Kong 1 (852) 2334 4481 1 (852) 2754 3125 • mixtg.hkd3sgs.com Laboratory: 1/F, Units 16-29 3/F, 4/F & 5/F, Dn Wui Centre, 25 Lok Yip Boad, Fanling, New Tentories, Hong Kong. www.sgsgroup.com.hk





Test Report No.T32120240597SN Date: Apr 26, 2021 Page 4 of 6

Result 1 Bacterial filtration efficiency (BFE) (EN14683:2019+AC:2019 Appendix B)

Test Side : Beige colour without earloop (Inside)

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Dimensions of test specimen : 173 mm x 94 mm

BFE Test Area : 49 cm²
BFE Flow Rate : 28.3 l/min

Test bacteria : Staphylococcus aureus ATCC 6538

Positive Control Average : 2.4 x 103 CFU
Negative Monitor Count : <1 CFU

Test Specimen	Percent BFE (%)
1	99.9
2	99.8
3	99.9
4	99.8
5	99.8

Result 2 Determination of Breathability (EN14683:2019+AC:2019 Appendix C Differential pressure)

Test Side : Beige colour without earloop (Inside)

Pre-Conditioning : Minimum of 4 hours at 21±5% and 85±5% R.H.

Test Area VI 2.S.K. UM; 4.9 cm²/ Flow Rate : 8 l/min

ri in hui		ΔP (Pa/cm²)			
Test Location	Specimen 1	Specimen 2	Specimen 3	Specimen 4	Specimen 5
Top Centre	26.5	34.3	33.4	26.5	30.4
Centre	23.5	27.5	26.5	22.6	26.5
Bottom Centre	23.5	25.5	27.5	22.6	28.4
Centre Left	23.5	21.6	26.5	21.6	20.6
Centre Right	20.6	24.5	21.6	22.6	23.5
Average	23.5	26.7	27.1	23.2	25.9

Result 3 Splash resistance (ISO 22609:2004)

Test Side : Beige colour with earloop (Outside)

Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.

Test Condition : 21±5 °C and (85±10)% R.H.
Test Pressure : 16.0 kPa (120 mmHg)

No of Test Specimen Tested : 32 No of Test Specimen Passed : 32

Test Specimen #	Synthetic Blood Penetration	
1-32	None Seen	

Note: Targeting-plate method was used.

Sample was tested in the stage of spreading the pleats out when mounted on the test fixture.

This document is leased by the Company subject to its General Conditions of Service printed overless, establish on request or accessible at http://www.aga.com/articless-accessible and Conditions of Electronic Documents at http://www.aga.com/articless-acc

Unless otherwise stated the results shown in this text report refer only to the sample(s) tested and auch sample(s) are retained for 30 days only:

SSS Horg Kang Limited | Office: Units: 303 & 305, 3/F Building: 22E, Phase 3, HK Science Park, New Tenitories, Hong Kong # (852) 2334 4481 # (852) 2364 3125 # mktg hk@sgs.com Laboratory: 1/F, Units: 16-29 3/F, 4/F & 5/F, On Wui Centre, 25 Lok Yip Read, Fanling, New Temitories, Hong Kong. www.sgsgroup.com.hk





Test Report

No.T32120240597SN

Date: Apr 26, 2021

Page 5 of 6

Result 4 Microbial cleanliness (Bioburden)(EN 14683:2019+AC:2019 Annex D)

Test Methods

Bioburden

The analyses were performed according to EN 14683:2019+AC:2019 and ISO 11737-1:2018

Test Results

	SGS Sample No.:HKHC	S Sample No.:HKHC210300000958-101		
Article Number	Mask Weight	Total Bioburden, cfu/mask	Total Bioburden, cfu/g	
1	3.65g	9	2.47	
2	3.65g	108	29.59	
3	3.74g	51	13.64	
4	3.64g	30	8.24	
5	3.67g	99	26.98	
6	3.64g	42	11.54	
7	3.60g	21	5.83	
8	3.640	24	6.59	
n	13.65g SK Un.	<i>0</i>) ²⁴ / ₂₄	IVIAS 6.58 M.	
10	3.65g	75	20.55	
MaskC	η by Mean:	48.3 z On	by 13.2	

Recovery Efficiency Correction Factor 2.0

Microbial Cleanliness (Bioburden): 26.0 cfu/g

Standard requirement#: ≤30 cfu/g

Note:

- Results reported on the submitted sample on an as received basis.
- cfu = Colony Forming Units
- Extraction method: by shaking at 250rpm for 5 minutes
- # EN 14683:2019+AC:2019 Medical face masks Requirements and test methods Performance requirements for medical face masks - Microbial cleanliness

This document is issued by the Company subject to its General Conditions of Service primarile or excessible in http://www.ags.com/w/Terms-and-Conditions.agar and, for electronic format documents, subject to Terms and Conditions for Electronic Documents is http://www.ags.com/w/Terms-and-Conditions/Terms-Electronic Conditions/Terms-Electronic Conditions/Terms-Electr

Unless otherwise stated the results shown in this test report refer only to the eample(s) tested and such sample(s) are retained for 30 days only





Test Report No.T32120240597SN Date: Apr 26, 2021 Page 6 of 6

Sample Photo:

Sample Picture (As received)





Taskom. Wive was

Mask On. SGS authenticate the photo on original report only

*** End of Report ***

Mo Mack On by Mo Mack On

This document is leased by the Company subject to its General Conditions of Service printed overlead, establish on request or accessible at http://www.ags.com/en/Terms-and-Conditions are reported from the company subject to the General Conditions for Residual and a highly-way agg, convive/Terms-and-Conditions/Terms-and-Con

Unless otherwise stated the results shown in this text report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

SOS Horg Kang Limited Diffice: Units 303 & 305, 3/F Building 22E, Phase 3, HK Science Park, New Tentories, Hong Kong 1 (852) 2334 4481 1 (852) 2754 3126 • mktg.hkdBsgs.com Laboratory: 1/F, Units 16-29 3/F, 4/F & 5/F, Dn Wui Centre, 25 Lok Yip Read, Faming, New Tentories, Hong Kong. www.sgsgroup.com.bk