

Guangdong Heshan Sentai Plastic Factory Co., Ltd.

Add:No.WenMin,Lane,DF93,HenSan,District,Jiangmen,City,Guangdong,Province.P.R.China

STERILIZATION CERTIFICATE

TYPE OF PRODUCT SUBJECT OF STERILIZATION:TRITON CARTRIDGES
NEEDLE

TYPE OF STERILIZATION :EO GAS

ITEM NO:TATTOO NEEDLES

DATE OF STERILIZATION: Jan 2022

LOT NO: 012022

We hereby certify that the above mentioned item which were
sterilized by EO GAS on Jan,2022 With the lot NO.012022



*The attachment is the detailed parameter report of EO GAS
sterilization



Sentai

EO-Sterilization Record Per Load

Date: 2022-01-23		Order: 2		Sterilization No: 012022		
Article Name	Production No	Specification	Quantity	Required Operation During Sterilization Cycle		
					Required	Actual
TRITON Cartridge needle	012022-1	51X36X21 cm	1-100	Tr. C	49-54	51
TRITON Cartridge needle	012022-2	51X36X21 cm	101-167	Pressure (KPa)	_68- _71 P	70 P
				Humidity (%RH)	30-85%	60
				Exposure time (min)	12:00	12:00
				Speed of Air Removal (Kpa/min)	1	2.5
				Speed of Aeration (Kpa/min)	//	//
				Weight of EO used Before (kg)	//	80
				Weight of EO used After (kg)	//	71
				Weight of EO used (kg)	8	9
				Pressure Changed (KPa)	s: // E: //	S-69.8 E:35.5
						2 FO. J

012022-1					
Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release
Conditioning	17:00	19:00	120	Nomal	
Air Removal	19:00	19:20	20	Nomal	
Pressure Time	19:20	20:00	40	Nomal	
EO injection	20:00	20:30	30	Nomal	
Exposure time	20:30	8:30	720	Nomal	
EO removal	8:30	9:30	60		
Flushing	9:30	12:00	150		
012022-2					
Required Operation	Start Time	End Time	Actual Time	Sterilization Information	Release
Conditioning	13:00	15:00	120	Nomal	
Air Removal	15:00	15:20	20	Nomal	
Pressure Time	15:20	16:00	40	Nomal	
EO injection	16:00	16:30	30	Nomal	
Exposure time	16:30	4:30	720	Nomal	
EO removal	7:00	7:30	60		
Flushing	7:30	10:00	150		