CERTIFICATE OF

CALIBRATION

Date Of Issue 8th Feb 2023 Certificate Number 004415



0601



Electronic Temperature Instruments Ltd.
Unit 1, Teville Industrials, Worthing
West Sussex BN14 8NW, UK
tel: 01903 202151
email: sales@etiltd.co.uk

www.etiltd.com

Page 1 of 1	
Approved Signatory	Signature
Sean Wells	alelle.

Customer Name: Signatrol Ltd

Address: Unit E2, Green Lane Business Park

Tewkesbury GL20 8SJ

Order Number: 46815

Date Received: 24th Jan 2023 Date Calibrated: 07th Feb 2023

Procedure: The Instrument was stabilised in ambient conditions, then calibrated by comparison against

traceable references in stirred liquid baths. The reported 'Instrument reading' is based on repeated measurements once stability criteria has been met in line with ETI procedure P5

Ambient Temperature: 23 ± 3 °C

Ambient Humidity: < 60 %rh

Temperature Scale: International Temperature Scale of 1990

Inst. Description: ETI Reference Thermometer

Integral Ø3.3 x 130mm 1/10th DIN Pt100 Liquid Probe

Inst. Serial Number: D12360142

Peripheral Device: N/A
Peripheral Device N°: N/A

Calibrated Range/Scale: -50.0 °C to +134.0 °C

Resolution: 0.01 °C

Comments: Immersion depth 110mm

Results:	Test Temperature °C	Instrument Reading °C	Error °C	Uncertainty °C
	-50.003	-50.02	-0.017	0.045 + Resolution
	-18.001	-18.02	-0.019	0.040 + Resolution
	-0.004	-0.01	-0.006	0.040 + Resolution
	4.002	4.01	0.008	0.040 + Resolution
	24.996	25.01	0.014	0.040 + Resolution
	100.008	100.02	0.012	0.040 + Resolution
	133.987	134.00	0.013	0.040 + Resolution

End Of Report

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability of long term stability.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.