

CERTIFICATE OF CALIBRATION

Date Of Issue
19th October 2012

Certificate Number
191012/N1



Certificate Issued By:



Electronic Temperature Instruments Ltd
Easting Close, Worthing
West Sussex BN14 8HQ England
Telephone +44 01903 202151
Facsimile +44 01903 202445
email: sales@etiltd.co.uk
Website: www.etiltd.co.uk

Page 1 of 2 pages

Approved Signatory

Signature

J. Carswell
D. Carter

Customer Name: SIGNATROL LIMITED
Address: 1st & 2nd FLOOR OFFICES
105 CHURCH STREET
TEWKESBURY
GLOUCESTERSHIRE
GL20 5AB

Order Number: 44129

Ref Number: 332/33048

Date Received: 15th October 2012

Date Calibrated: 17/18th October 2012

Ambient Temperature: 22 °C ± 2 °C

Ambient Humidity: <60 % rh

Temperature Scale: International Temperature Scale of 1990

Instrument Type : EUROTRON MICROCAL 1 SIMULATOR

Instrument Serial Number: 0049272

Procedure: The instrument was stabilised at ambient temperature, then calibrated by measuring its output on traceable reference equipment.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Results indicate performance of instrument at time of measurement, with no warranty as to specification, repeatability or long term stability.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurements to recognised national standards, and to units of measurement realised at the National Physical Laboratory or other recognised national standards laboratories. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No 0601

Certificate Number
191012/N1

Page 2 of 2 pages

Instrument Serial No 0049272

Results

<u>Type "J" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
-100.0	-4.633	-4.636 1	-100.1	± 0.25 °C
700.0	39.132	39.130 6	700.0	± 0.15 °C

<u>Type "K" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
-100.0	-3.554	-3.555 5	-100.0	± 0.25 °C
900.0	37.326	37.324 1	900.0	± 0.15 °C

<u>Type "T" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
-100.0	-3.379	-3.374 5	-99.9	± 0.25 °C
300.0	14.862	14.864 8	300.1	± 0.15 °C

<u>Type "S" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
800.0	7.345	7.341 2	799.8	± 1 °C
1400.0	14.373	14.371 4	1399.9	± 1 °C

<u>Type "R" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
800.0	7.950	7.947 5	799.8	± 1 °C
1400.0	16.040	16.034 8	1399.6	± 1 °C

<u>Type "E" Output Temperature °C</u>	<u>Nominal mV</u>	<u>Measured mV</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
-100.0	-5.237	-5.238 4	-100.0	± 0.25 °C
650.0	49.116	49.108 9	649.8	± 0.15 °C

<u>PT100 Output Temperature °C</u>	<u>Nominal Ohms</u>	<u>Measured Ohms</u>	<u>Equivalent Temperature °C</u>	<u>Uncertainty of Measurement</u>
0.0	0.00	100.040 8	0.10	± 0.025 °C
600.0	313.71	313.647 1	599.81	± 0.18 °C

End of Report