

CERTIFICATE OF CALIBRATION

ISSUED BY ROTRONIC INSTRUMENTS (UK) LTD

DATE OF ISSUE: 19th January 2021

CERTIFICATE NUMBER: 37981



rotronic
MEASUREMENT SOLUTIONS

Calibrated by: G Thompson

Approved Signatory: M Smith 

Unit 1a Crompton Fields, Crompton Way,
Crawley, West Sussex, RH10 9EE.

Telephone: 01293 571000

Fax: 01293 571008

Email: service@rotronic.co.uk

www.rotronic.co.uk

Page 1 of 1

Dates Measurements Performed:

17th to 19th January 2021

Calibration Procedure Used: RUKP2

Customer Details : Signatrol Limited, Unit E2, Green Lane Business Park, Tewkesbury,
: Gloucestershire, GL20 8SJ

Customer's Order Number : 46152

Rotronic Ref Number : 37981

Instrument Description : Humidity and temperature generator with chilled mirror hygrometer control

Manufacturer : Michell Instruments

Model Type (s) : OptiCal

Serial Number (s) : 071476/154675/153905

The Chamber was calibrated by comparison against a chilled mirror hygrometer certified as traceable to National Standards. The hygrometer was also calibrated in terms of temperature by comparison with platinum resistance thermometers, inserted into the chamber, which are traceable to national standards. The applied relative humidity was calculated using the measured dew point and the measured temperature. The indicated values were taken from the instruments display and are given in the table below. The calibration was conducted in controlled laboratory conditions of 23 °C ± 2 °C. The temperature scale used is ITS-90.

Applied Dew Point (°C)	Calibration Uncertainty Dew Point *(°C)	Calculated Relative Humidity (%rh)	Calibration Uncertainty * (%rh)	Applied Temperature setpoint (°C)	Calibration Uncertainty * (°C)	Observed Dew/Frost Point (°C)	Instrument Error Dew Point (°C)	Applied Relative Humidity setpoint (%rh)	Indicated Relative Humidity (%rh)	Instrument Error (%rh)	Measured Temperature of Optical (°C)
-9.82	±0.17	11.6	±0.5	21.2	±0.17	-9.6	-0.2	10.0	10.7	-0.9	21.1
0.88	±0.17	25.8	±0.5	21.2	±0.17	1.0	+0.1	25.0	26.0	+0.2	21.1
10.79	±0.17	51.5	±0.8	21.2	±0.17	10.9	+0.1	50.0	52.0	+0.5	21.1
16.99	±0.17	77.2	±1.2	21.0	±0.17	17.2	+0.2	75.0	78.4	+1.2	21.1
19.90	±0.17	92.0	±1.4	21.0	±0.17	20.1	+0.2	90.0	93.4	+1.4	21.2

* The uncertainties quoted apply only to values obtained during the calibration and are not indicative of long-term stability of the instrument under calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. (TSDC26 Issue 10)