

CERTIFICATE OF CALIBRATION

ISSUED BY ROTRONIC INSTRUMENTS (UK) LTD

DATE OF ISSUE: 20th March 2024

CERTIFICATE NUMBER: 43767



0766



Calibrated by: G Thompson

Approved Signatory: M Smith

Page 1 of 1

Dates Measurements Performed:

18th to 20th March 2024

Calibration Procedure

RUKP2

Used:

Unit 1a Crompton Fields, Crompton Way,
Crawley, West Sussex, RH10 9EE.

Telephone: 01293 571000

uk.service.crawley@processsensing.com

www.processsensing.com

Customer Details : Signatrol Ltd, Unit E2, Green Lane Business Park, Tewkesbury
: Gloucestershire, GL20 8SJ
Customer's Order Number : 47154
Rotronic Ref Number : 43767
Instrument Description : Humidity and temperature generator with chilled mirror hygrometer control
Manufacturer : Michell Instruments
Model Type (s) : Optical
Serial Number (s) : 071476/154675/153905

The hygrometer was calibrated by comparison against a chilled mirror hygrometer certified as traceable to National Standards. The hygrometer was also calibrated in terms of temperature by comparison with platinum resistance thermometers, which are traceable to national standards. The applied relative humidity was calculated using the measured dew point and the measured temperature. The indicated values were taken from the instruments display and are given in the table below. The calibration was conducted in an environmental chamber. The calibration was conducted in controlled laboratory conditions of 23 °C ± 2 °C. The probe under calibration was fully immersed. The temperature scale used is ITS-90. The measurements shown below are exclusive to the instrument on the certificate.

Applied Dew Point (°C)	Calibration Uncertainty Dew Point ** (°C)	Calculated Relative Humidity (%rh)	Calibration Uncertainty ** (%rh)	Applied Temperature Setpoint Optical (°C)	Calibration Uncertainty ** (°C)	Applied Relative Humidity Setpoint Optical (%rh)	Indicated Relative Humidity Optical (%rh)	Instrument Error (%rh)	Indicated Dewpoint Optical (°Cdp)***	Instrument Error (°Cdp)	Measured Temp. Optical (°C)	Instrument Error (°C)
-8.19*	±0.17	12.2	±0.5	21.0	±0.17	10	11.4	-0.8	-8.9	-0.7	21.1	+0.1
1.22	±0.17	26.6	±0.5	21.0	±0.17	25	27.0	+0.4	1.4	+0.2	21.1	+0.1
11.08	±0.17	52.5	±0.7	21.0	±0.17	50	53.4	+0.9	11.3	+0.2	21.1	+0.1
17.42	±0.17	79.0	±0.9	21.0	±0.17	75	80.4	+1.4	17.7	+0.3	21.2	+0.2
19.99	±0.17	92.7	±1.1	21.0	±0.17	90	94.0	+1.3	20.2	+0.2	21.2	+0.2

*Ice was on the reference hygrometer

** The uncertainties quoted apply only to values obtained during the calibration and are not indicative of long-term stability of the instrument under calibration.

***calculated Psychrometric conversions from HW4 V3.9

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory. (TSDC26 Issue 12)