

# CERTIFICATE OF CALIBRATION

Certificate Number **RH01024**  
Date of Issue 10<sup>th</sup> January 2020



0179



48 Lancaster Way Business Park  
Ely, Cambridgeshire  
CB6 3NW  
United Kingdom

Tel: +44 (0)1353 658000  
Fax: +44 (0)1353 658199  
e-mail: uk.info@michell.com  
Website: www.michell.com

Approved Signatory  
Mr S. Daines

Accredited to  
ISO/IEC 17025:2017

*S. Daines*

Page 1 of 2

<b>Customer</b>	Signatrol Limited
<b>Customer address</b>	Unit E2 Green Lane Business Park, Tewkesbury, Gloucestershire GL20 8SJ
<b>Customer order number</b>	45909
<b>Received</b>	06 <sup>th</sup> January 2020
<b>Instrument</b>	Michell Instruments Opti-Cal. Integrated humidity calibrator.
<b>Opti-Cal serial number</b>	071476
<b>Serial number Instrument</b>	154675
<b>Sensor</b>	153905 <b>Temperature probe</b> 155972
<b>Michell reference number</b>	T41713
<b>Measurements performed</b>	09 <sup>th</sup> & 10 <sup>th</sup> January 2020
<b>Laboratory temperature</b>	21 °C ±2 °C
<b>Laboratory humidity</b>	32 %rh ±10 %rh

The Opti-Cal instrument was calibrated by comparison using reference instruments with UKAS calibration certificates.

The dew point sensor was removed from the Opti-Cal chamber and placed in a climatic chamber with the reference hygrometer sensor. Another, non-functional sensor was used to plug the sensor hole. The two reference temperature probes were inserted 170 mm into the Opti-Cal chamber, through the left and right hand ports of the front access plate. The Opti-Cal chamber humidity was set to match the calibration conditions for each test point of this certificate.

The sample air from the Michell Instruments dew point generator was divided, with one path to the reference hygrometer and the other to the hygrometer being calibrated.

The optical surfaces of the hygrometers were cleaned using de-ionised water, prior to the calibration.

The output used from the hygrometer was the digital display with a resolution to 0.1, for all parameters.

At each relative humidity, time was allowed to ensure that the calibration conditions had stabilised. This was confirmed while recording the 10 readings (at 2 minute intervals) that are averaged to give the figures recorded in this certificate.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

# CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory 0179

Certificate Number **RH01024**

Page 2 of 2

No adjustment was made to the hygrometer before the calibration was performed.

Applied temperature °C	Applied RH %rh	Test instrument					
		Indicated temperature °C	Correction temperature °C	Expanded uncertainty temperature °C	Indicated RH %rh	Correction RH %rh	Expanded uncertainty RH %rh
21.05	9.94	21.0	+0.1	±0.24	9.9	0.0	±0.24
20.99	25.14	21.0	0.0	±0.24	25.3	-0.2	±0.59
20.96	49.88	20.9	+0.1	±0.24	50.3	-0.4	±1.2
21.00	74.89	21.0	0.0	±0.24	75.8	-0.9	±1.8
20.99	90.07	21.0	0.0	±0.24	91.1	-1.0	±2.1

Applied temperature °C	Applied Dew point °C	Test instrument		
		Indicated Dew point °C	Correction Dew point °C	Expanded uncertainty Dew point °C
21.05	-10.52	-10.5	0.0	±0.25
20.99	0.31	0.4	-0.1	±0.25
20.96	10.12	10.2	-0.1	±0.25
21.00	16.38	16.5	-0.1	±0.25
20.99	19.30	19.5	-0.2	±0.25

Dew points below 0 °C are measured over ice, unless otherwise stated.

The indicated sample flow rate through the dew point sensor was 1.0 l/min.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k=2$ , providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The uncertainties quoted in the Certificate of Calibration only apply to the measured value obtained during the period of calibration and are not indicative of the long-term stability of the instrument under test.