

# 1. Certificate of Analysis

<b>Scope</b>	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
<b>Product Lot</b>	100793
<b>Date Issued</b>	Jun2023

Population	MFI specification	Results (Pass/Fail)
SpectraComp eGFP Negative	<3000	Pass
SpectraComp eGFP Low	10k - 30k	Pass
SpectraComp eGFP Medium	100k - 250k	Pass
SpectraComp eGFP High	1M - 2.5M	Pass

**\*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings**

Approved by: Jonathan Yang 06 / 13 / 2023  
Jonathan Yang / Director, Quality Assurance and Regulatory Affairs

Slingshot Biosciences Internal Part Number: 11-0630v1