

November 8, 2021

Angela Drysdale
VP, Regulatory Affairs
Abbott Diagnostics Scarborough, Inc.
10 Southgate Road
Scarborough, ME 04074

Device: BinaxNOW COVID-19 Antigen Self Test

EUA Number: EUA210264

Company: Abbott Diagnostics Scarborough, Inc.

Indication: Non-prescription home use for the qualitative detection of nucleocapsid protein antigen from SARS-CoV-2 with:

Self-collected anterior nasal (nares) swab samples from individuals aged 15 years or older with symptoms of COVID-19 within the first seven days of symptom onset.

Adult collected anterior nasal (nares) swab samples from individuals aged two years or older with symptoms of COVID-19 within the first seven days of symptom onset.

Self-collected anterior nasal (nares) swab samples from individuals aged 15 years or older, or adult collected anterior nasal (nares) swab samples from individuals aged two years or older, with or without symptoms or other epidemiological reasons to suspect COVID-19 when tested twice over three days with at least 24 hours (and no more than 48 hours) between tests.

Dear Ms. Drysdale:

On March 31, 2021, based on your¹ request the Food and Drug Administration (FDA) issued an Emergency Use Authorization (EUA) for the BinaxNOW COVID-19 Antigen Self Test pursuant to Section 564 of the Federal Food, Drug, and Cosmetic Act (the Act) (21 U.S.C. §360bbb-3) for the indications stated in the letter.² Based on your request, FDA granted updates to the

¹ For ease of reference, this letter will use the term “you” and related terms to refer to Abbott Diagnostics Scarborough, Inc.

² The March 31, 2021, letter authorized your product for the qualitative detection of nucleocapsid protein antigen from SARS-CoV-2 from individuals with or without symptoms or other epidemiological reasons to suspect COVID-19 infection when tested twice over three days with at least 36 hours between tests. This test was authorized for non-prescription home use with self-collected direct anterior nasal (nares) swab samples from individuals aged 15 years or older or adult collected anterior nasal swab samples from individuals aged two years or older.

authorized labeling on April 1, 2021.³ In addition, based on your request, the March 31, 2021, letter was revised and reissued by FDA on August 10, 2021.⁴

On July 6, 2021, you requested to amend this EUA. Based on that request, and having concluded that revising the August 10, 2021, EUA is appropriate to protect the public health or safety under section 564(g)(2)(C) of the Act (21 U.S.C. § 360bbb-3(g)(2)(C)), FDA is reissuing the August 10, 2021, letter in its entirety with the revisions incorporated.⁵ Accordingly, your product⁶ is hereby authorized pursuant to section 564 of the Act when used pursuant to the Scope of Authorization (Section II) and Conditions of Authorization (Section IV) of this reissued letter.

On February 4, 2020, pursuant to Section 564(b)(1)(C) of the Act, the Secretary of the Department of Health and Human Services (HHS) determined that there is a public health emergency that has a significant potential to affect national security or the health and security of United States citizens living abroad, and that involves the virus that causes COVID-19. Pursuant to Section 564 of the Act, and on the basis of such determination, the Secretary of HHS then declared that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of the virus that causes COVID-19 subject to the terms of any authorization issued under Section 564(a) of the Act.⁷

³ On April 1, 2021, your request was granted via email acknowledgement to include minor updates to the BinaxNOW COVID-19 Antigen Self Test Instructions.

⁴ On August 10, 2021, the revisions to the March 31, 2020, letter and authorized labeling included: (1) addition of the end user option to use digital electronic-based step-by-step “BinaxNOW COVID-19 Antigen Self Test Instructions” either via a website or by downloading the “NAVICA App” mobile application onto a compatible smartphone, (2) addition of the option to facilitate automatic reporting of test results via the “NAVICA App” mobile application, (3) updates to the performance data to include results of usability studies evaluating the digital electronic-based step-by-step “BinaxNOW COVID-19 Antigen Self Test Instructions,” (4) updates to the Conditions of Authorization to add new Conditions related to circulating variants (Conditions R. and S.) and update timeframe in Condition Q; (5) removal of Conditions of Authorization R. and S. from the March 31, 2021, letter that were fulfilled, and (6) updates to Fact Sheet for Healthcare Providers and Fact Sheet for Individuals to reflect language used in more recent authorizations.

⁵ The revisions to the August 10, 2021, letter and authorized labeling include: (1) updates to the intended use to include use of your product with “*self-collected anterior nasal (nares) swab samples from individuals aged 15 years or older with symptoms of COVID-19 within the first seven days of symptom onset,*” “*adult collected anterior nasal (nares) swab samples from individuals aged two years or older with symptoms of COVID-19 within the first seven days of symptom onset,*” and “*self-collected anterior nasal (nares) swab samples from individuals aged 15 years or older, or adult collected anterior nasal (nares) swab samples from individuals aged two years or older, with or without symptoms or other epidemiological reasons to suspect COVID-19 when tested twice over three days with at least 24 hours (and no more than 48 hours) between tests*”, (2) update the outer box labels and Instructions for Use (IFU) limitations section to reflect the updates to the intended use, (3) offer the test kit with 1 test and 10 tests options, in addition to the 2 kit option currently offered, (4) Update Condition of Authorization Q. below to give a 4 month extension, and (5) updates to the letter and fact sheets to reflect the updated intended use and for consistency with language used in more recent authorizations.

⁶ For ease of reference, this EUA will use the term “your product” to refer to the BinaxNOW COVID-19 Antigen Self Test used for the indication identified above.

⁷ U.S. Department of Health and Human Services, *Determination of a Public Health Emergency and Declaration that Circumstances Exist Justifying Authorizations Pursuant to Section 564(b) of the Federal Food, Drug, and Cosmetic Act*, 21 U.S.C. § 360bbb-3. 85 FR 7316 (February 7, 2020).

FDA considered the totality of scientific information available in authorizing the emergency use of your product for the indication above. A summary of the performance information FDA relied upon is included in the “BinaxNOW COVID-19 Antigen Self Test Healthcare Provider Instructions for Use” (identified below).

Having concluded that the criteria for issuance of this authorization under Section 564(c) of the Act are met, I am authorizing the emergency use of your product, described in the Scope of Authorization of this letter (Section II), subject to the terms of this authorization.

I. Criteria for Issuance of Authorization

I have concluded that the emergency use of your product meets the criteria for issuance of an authorization under Section 564(c) of the Act, because I have concluded that:

1. The SARS-CoV-2 can cause a serious or life-threatening disease or condition, including severe respiratory illness, to humans infected by this virus;
2. Based on the totality of scientific evidence available to FDA, it is reasonable to believe that your product may be effective in diagnosing COVID-19, and that the known and potential benefits of your product when used for diagnosing COVID-19, outweigh the known and potential risks of your product; and
3. There is no adequate, approved, and available alternative to the emergency use of your product.⁸

II. Scope of Authorization

I have concluded, pursuant to Section 564(d)(1) of the Act, that the scope of this authorization is limited to the indication above.

Authorized Product Details

Your product is a lateral flow immunoassay intended for the qualitative detection of nucleocapsid protein antigen from SARS-CoV-2. This test is authorized for non-prescription home use with self-collected anterior nasal (nares) swab samples from individuals aged 15 years or older with symptoms of COVID-19 within the first seven days of symptom onset. This test is also authorized for non-prescription home use with adult collected anterior nasal (nares) swab samples from individuals aged two years or older with symptoms of COVID-19 within the first seven days of symptom onset.

This test is also authorized for non-prescription home use with self-collected anterior nasal (nares) swab samples from individuals aged 15 years or older, or adult collected anterior nasal (nares) swab samples from individuals aged two years or older, with or without symptoms or other epidemiological reasons to suspect COVID-19 when tested twice over three days with at least 24 hours (and no more than 48 hours) between tests. Your product does not differentiate

⁸ No other criteria of issuance have been prescribed by regulation under Section 564(c)(4) of the Act.

between SARS-CoV and SARS-CoV-2.

The SARS-CoV-2 nucleocapsid protein antigen is generally detectable in anterior nasal (nares) swabs during the acute phase of infection. Positive results indicate the presence of viral antigens, but clinical correlation with past medical history and other diagnostic information is necessary to determine infection status. Positive results do not rule out bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease. Individuals who test positive with your product should self-isolate and seek follow-up care with their physician or healthcare provider as additional testing may be necessary.

Negative results should be treated as presumptive and confirmation with a molecular assay, if necessary, for patient management, may be performed. Negative results do not rule out SARS-CoV-2 infection and should not be used as the sole basis for treatment or patient management decisions, including infection control decisions. Negative results should be considered in the context of an individual’s recent exposures, history and the presence of clinical signs and symptoms consistent with COVID-19.

For serial testing programs, additional confirmatory testing with a molecular test for negative results may be necessary, if there is a high likelihood of COVID-19, such as, an individual with a close contact with COVID-19 or with suspected exposure to COVID-19 or in communities with high prevalence of infection. Additional confirmatory testing with a molecular test for positive results may also be necessary, if there is a low likelihood of COVID-19, such as in individuals without known exposures to COVID-19 or residing in communities with low prevalence of infection.

Individuals who test negative and continue to experience COVID-like symptoms of fever, cough and/or shortness of breath may still have SARS-CoV-2 infection and should seek follow up care from their healthcare provider.

Individuals should report their test result through the NAVICA App and provide all results obtained with this product to their healthcare provider in order to receive appropriate medical care. All healthcare providers will report all test results they receive from individuals who use the authorized product to relevant public health authorities in accordance with local, state, and federal requirements using appropriate LOINC and SNOMED codes, as defined by the Laboratory In Vitro Diagnostics (LIVD) Test Code Mapping for SARS-CoV-2 Tests provided by the Centers for Disease Control and Prevention.

The individual using your product has the option to follow the “BinaxNOW COVID-19 Antigen Self Test Instructions” provided as a hard copy with the kit or follow digital electronic-based step-by-step “BinaxNOW COVID-19 Antigen Self Test Instructions” either via a website or by downloading the “NAVICA App” mobile application onto a compatible smartphone.⁹ To use your

⁹ The digital electronic “BinaxNOW COVID-19 Antigen Self Test Healthcare Provider Instructions for Use” can be accessed via <https://www.globalpointofcare.abbott/en/product-details/binaxnow-covid-19-antigen-self-test-us.html>. Compatible smart phone for the NAVICA App includes Apple iPhone running Operation System (iOS): latest major version and two prior major versions (iPhone running iOS v12 or later), and Android Phones: latest major version and two prior major versions (Android phone running Android OS v9 or later). Additional smart phone models as may be requested, and for which you receive appropriate authorization, in accordance with Condition L. below.

product the individual performing the test first opens the test kit before 6 drops of extraction reagent from a dropper bottle are added to the top hole of the swab well. A nasal swab specimen is then either self-collected or adult-collected following the instructions provided before being inserted into the test card through the bottom hole of the swab well, and firmly pushed upwards until the swab tip is visible through the top hole. The swab is rotated 3 times clockwise and the card is closed, bringing the extracted sample into contact with the test strip. Test results are interpreted visually after 15 minutes based on the presence or absence of visually detectable pink/purple colored lines. Test results should not be read after 30 minutes. Upon completion of the test and result interpretation the user should share their results with their healthcare provider and report their test result through the NAVICA App.

The BinaxNOW COVID-19 Antigen Self Test includes the following materials or other authorized materials: Test Cards, Extraction Reagent bottles, and Nasal Swabs.

Your product includes an internal control test line that must generate the expected result for a test to be considered valid, as outlined in the “BinaxNOW COVID-19 Antigen Self Test Healthcare Provider Instructions for Use” and the “BinaxNOW COVID-19 Antigen Self Test Instructions.”

The labeling entitled “BinaxNOW COVID-19 Antigen Self Test Instructions,” the “BinaxNOW COVID-19 Antigen Self Test Healthcare Provider Instructions for Use,” the website accessible digital electronic “BinaxNOW COVID-19 Antigen Self Test Instructions” the “BinaxNOW COVID-19 Antigen Self Test” kit box labels (1-test, 2-tests and 10-tests) (available at <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas>), the “NAVICA App” software application with “BinaxNOW COVID-19 Antigen Self Test Instructions” and the following fact sheets pertaining to the emergency use, is required to be made available as set forth in the Conditions of Authorization (Section IV), and are collectively referred to as “authorized labeling”:

- Fact Sheet for Healthcare Professionals: Abbott Diagnostics Scarborough, Inc. - The BinaxNOW COVID-19 Antigen Self Test
- Fact Sheet for Individuals: Abbott Diagnostics Scarborough, Inc. - The BinaxNOW COVID-19 Antigen Self Test

The above described product, when accompanied by the authorized labeling as set forth in the Conditions of Authorization (Section IV) is authorized to be distributed and used under this EUA, despite the fact that it does not meet certain requirements otherwise required by applicable federal law.

I have concluded, pursuant to Section 564(d)(2) of the Act, that it is reasonable to believe that the known and potential benefits of your product, when used consistent with the Scope of Authorization of this letter (Section II), outweigh the known and potential risks of your product.

I have concluded, pursuant to Section 564(d)(3) of the Act, based on the totality of scientific evidence available to FDA, that it is reasonable to believe that your product may be effective in diagnosing COVID-19, when used consistent with the Scope of Authorization of this letter

(Section II), pursuant to Section 564(c)(2)(A) of the Act.

FDA has reviewed the scientific information available to FDA, including the information supporting the conclusions described in Section I above, and concludes that your product (as described in the Scope of Authorization of this letter (Section II)) meets the criteria set forth in Section 564(c) of the Act concerning safety and potential effectiveness.

The emergency use of your product under this EUA must be consistent with, and may not exceed, the terms of this letter, including the Scope of Authorization (Section II) and the Conditions of Authorization (Section IV). Subject to the terms of this EUA and under the circumstances set forth in the Secretary of HHS’s determination under Section 564(b)(1)(C) of the Act described above and the Secretary of HHS’s corresponding declaration under Section 564(b)(1) of the Act, your product is authorized for the indication above.

III. Waiver of Certain Requirements

I am waiving the following requirements for your product during the duration of this EUA:

- Current good manufacturing practice requirements, including the quality system requirements under 21 CFR Part 820 with respect to the design, manufacture, packaging, labeling, storage, and distribution of your product, but excluding Subpart H (Acceptance Activities, 21 CFR 820.80 and 21 CFR 820.86), Subpart I (Nonconforming Product, 21 CFR 820.90), and Subpart O (Statistical Techniques, 21 CFR 820.250).

IV. Conditions of Authorization

Pursuant to Section 564(e) of the Act, I am establishing the following conditions on this authorization:

Abbott Diagnostics Scarborough, Inc. (You) and Authorized Distributor(s)¹⁰

- A. Your product must comply with the following labeling requirements: the intended use statement in 21 CFR 809.10(a)(2), (b)(2); adequate directions for use in 21 U.S.C. 352(f) and 21 CFR 809.10(b)(5), (7), and (8); appropriate limitations on the use of the device including information required under 21 CFR 809.10(a)(4); and any available information regarding performance of the device, including requirements under 21 CFR 809.10(b)(12).
- B. You and authorized distributor(s) must make available the “BinaxNOW COVID-19 Antigen Self Test Instructions” and the “Fact Sheet for Individuals” for your product in the shipped kit using the “BinaxNOW COVID-19 Antigen Self Test” kit box labels (1-test, 2-tests and 10-tests) and make these two documents electronically available on your website.

¹⁰ “Authorized Distributor(s)” are identified by you, Abbott Diagnostics Scarborough, Inc., in your EUA submission as an entity allowed to distribute your product.

- C. You and authorized distributor(s) must maintain records of customer complaint files and report to FDA any significant complaints about usability or deviations from the established performance characteristics of which you and authorized distributor(s) become aware.
- D. You and authorized distributor(s) must inform relevant public health authorities of this EUA, including the terms and conditions herein, and any updates made to your product and/or the authorized labeling.
- E. Through a process of inventory control, you and authorized distributor(s) must maintain records of the locations (e.g., pharmacies, doctor’s offices, etc.) to which your product is distributed and the number of tests distributed to each location.
- F. You and authorized distributor(s) must collect information on the performance of your product and have a process in place to track adverse events, including any occurrence of false positive or false negative results and significant deviations from the established performance characteristics of the product of which you become aware and report any such events to FDA in accordance with 21 CFR Part 803. Serious adverse events, especially unexpected biosafety concerns, should immediately be reported to the Division of Microbiology (DMD)/Office of Health Technology 7 (OHT7)-Office of In Vitro Diagnostics and Radiological Health (OIR)/Office of Product Evaluation and Quality (OPEQ)/Center for Devices and Radiological Health (CDRH) (via email: CDRH-EUARreporting@fda.hhs.gov).
- G. You and authorized distributor(s) are authorized to make available additional information relating to the emergency use of your product that is consistent with, and does not exceed, the terms of this letter of authorization.
- H. You and authorized distributor(s) using your product must ensure that any records associated with this EUA are maintained until otherwise notified by FDA. Such records will be made available to FDA for inspection upon request.

Abbott Diagnostics Scarborough, Inc. (You)

- I. You must notify FDA of any authorized distributor(s) of your product, including the name, address, and phone number of any authorized distributor(s).
- J. You must provide authorized distributor(s) with a copy of this EUA and communicate to authorized distributor(s) any subsequent revisions that might be made to this EUA and its authorized accompanying materials, including the authorized labeling.
- K. You must make the authorized “BinaxNOW COVID-19 Antigen Self Test Healthcare Provider Instructions for Use” and the “Fact Sheet for Healthcare Professionals” electronically available on your website. Additionally, you must provide the opportunity to request a copy of the “BinaxNOW COVID-19 Antigen Self Test Healthcare Provider

Instructions for Use” and “Fact Sheet for Healthcare Professionals” in paper form, and after such request, promptly provide the requested labeling at no additional cost.

- L. You may request changes to this EUA for your product, including to the Scope of Authorization (Section II in this letter) or to the authorized labeling, including requests to make available additional authorized labeling specific to an authorized distributor. Such additional labeling may use another name for the product but otherwise must be consistent with the authorized labeling, and shall not exceed the terms of authorization of this letter. Any request for changes to this EUA should be submitted to the DMD/OHT7-OIR/OPEQ/CDRH and require appropriate authorization from FDA prior to implementation.
- M. You must comply with the following requirements pursuant to FDA regulations: 21 CFR 820 Subpart H (Acceptance Activities, 21 CFR 820.80 and 21 CFR 820.86), Subpart I (Nonconforming Product, 21 CFR 820.90), and Subpart O (Statistical Techniques, 21 CFR 820.250).
- N. You must have lot release procedures and the lot release procedures, including the study design and statistical power, must ensure that the product released for distribution meet the clinical and analytical performance claimed in the authorized labeling.
- O. If requested by FDA, you must submit your lot release procedures to FDA, including sampling protocols, testing protocols, and acceptance criteria, that you use to release lots of your product for distribution in the U.S. If such lot release procedures are requested by FDA, you must provide them within 48 hours of the request.
- P. You must evaluate the analytical limit of detection and assess traceability¹¹ of your product with any FDA-recommended reference material(s). After submission to and concurrence with the data by FDA, you will update your labeling to reflect the additional testing. Such labeling updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- Q. You must evaluate the clinical performance of your product to support the serial screening claim in an FDA agreed upon post authorization clinical evaluation study within 6 months of the date of this letter (unless otherwise agreed to with DMD/OHT7-OIR/OPEQ/CDRH). After submission to and concurrence with the data by FDA, you must update the authorized labeling to reflect the additional testing. Such labeling updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.
- R. You must evaluate the impact of SARS-CoV-2 viral mutations on your product’s performance. Such evaluations must occur on an ongoing basis and must include any additional data analysis that is requested by FDA in response to any performance concerns you or FDA identify during routine evaluation. Additionally, if requested by FDA, you must submit records of these evaluations for FDA review within 48 hours of

¹¹ Traceability refers to tracing analytical sensitivity/reactivity back to an FDA-recommended reference material.

the request. If your evaluation identifies viral mutations that affect the stated expected performance of your device, you must notify FDA immediately (via email: CDRH-EUA-Reporting@fda.hhs.gov).

- S. If requested by FDA, you must update your labeling within 7 calendar days to include any additional labeling risk mitigations identified by FDA, such as those related to the impact of viral mutations on test performance. Such updates will be made in consultation with, and require concurrence of, DMD/OHT7-OIR/OPEQ/CDRH.

Conditions Related to Printed Materials, Advertising and Promotion

- T. All descriptive printed matter, advertising, and promotional materials relating to the use of your product shall be consistent with the authorized labeling, as well as the terms set forth in this EUA and meet the requirements set forth in section 502(a), (q)(1), and (r) of the Act, as applicable, and FDA implementing regulations.
- U. No descriptive printed matter, advertising, or promotional materials relating to the use of your product may represent or suggest that this test is safe or effective for the detection of SARS-CoV-2.
- V. All descriptive printed matter, advertising, and promotional materials relating to the use of your product shall clearly and conspicuously state that:
- This product has not been FDA cleared or approved; but has been authorized by FDA under an EUA;
 - This product has been authorized only for the detection of proteins from SARS- CoV-2, not for any other viruses or pathogens; and,
 - The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

The emergency use of your product as described in this letter of authorization must comply with the conditions and all other terms of this authorization.

V. Duration of Authorization

This EUA will be effective until the declaration that circumstances exist justifying the authorization of the emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 is terminated under Section 564(b)(2) of the Act or the EUA is revoked under Section 564(g) of the Act.

Sincerely,

Jacqueline A. O'Shaughnessy, Ph.D.
Acting Chief Scientist
Food and Drug Administration

Enclosure