

T-Square[®]


Test Procedure



Allow the kit and specimen to come to room temperature prior to testing.
AVOID PLACING ANYTHING IN THE MOUTH 10 MINUTES PRIOR TO TESTING.


Step 1

Remove the test cube and the sponge collector from the foil pouch by tearing at the notch. Place the test cube upright on a level surface.



Step 2


Put the sponge end of the collector on your tongue to collect oral fluid for about 7 minutes or 3 minutes if using the saturation indicator, until the color on saturation indicator strip appears RED in the indicator window.



Step 3

Open the test cube and place the fully saturated sponge collector inside the test cube. Press the sponge collector down firmly until it reaches the bottom of the test cube, then close the cube lid tightly while compressing the collector. Keep test cube upright on a flat surface and follow Step 4.

Note: Make sure the sponge collector is inserted vertically and the handle of collector is put into the clamp.



Step 4

1. Interpreting Alcohol Test Result:

Read result at 2 minutes.

NEGATIVE: No color changes on reaction pad.


POSITIVE: Compare the color of reaction pad with provided color chart to estimate the approximate blood alcohol level.

2. Interpreting Drug Test Results:


Read results at 5 minutes. A negative result will be shown by two color lines on the strip.

The intensity of the line does not matter. A positive result will be shown by a single color line in the C section of the test strip. If there are no color lines present or a single color line in the T section of the test strip, the results are invalid, and the test should be repeated.

Drug Test Results



Blood Alcohol Concentration



NOTE: Results after more than 2 minutes may be not accurate