


Reference Material Product Information Sheet

Epichem's Quality System conforms to ISO9001:2015 as certified by ECAAS Pty Ltd - Certification number 616061.



Name	Triprolidine- <i>N</i> -oxide oxalate salt (1:1)
Synonym(s)	Triprolidine- <i>N</i> -oxide oxalate; 2-((<i>E</i>)-1-(4-methylphenyl)-3-(1-oxidopyrrolidin-1-yl)prop-1-enyl)pyridine ethanedioate
Epichem Item #	EPL-AA204 Batch 1
CAS #	1310575-86-0 (freebase)
Molecular Formula	C ₁₉ H ₂₂ N ₂ O.C ₂ H ₂ O ₄
Molecular Weight	384.43 g/mol
Appearance	Blue crystalline powder
Melting Point	134.6-139.3°C
Combustion Analysis	Required (%): C:65.6; H:6.3; N:7.3. Found (%): C:65.6; H:6.4; N:7.3.
Purity*	97.2%
Date of Manufacture	8 September 2016
Storage Requirements	Store in tightly closed vessel. Protect from heat, light and moisture.
Special Precautions	This compound is for laboratory use only. Its toxicological properties may not have been fully established. It should be handled only by suitably qualified personnel.
Intended Use	This compound is suitable for the identification of impurities and degradants in pharmaceutical materials. The purity assay is considered as relative contribution.
Date of Shipment	TBA This certificate is valid for one year from the date of shipment provided the substance is unopened and stored under the recommended conditions.
Retest Date	TBA (Proper Storage and Handling Required)

* NATA accreditation does not cover the performance of this service

EPL-AA204 Batch 1

Revision 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902


I. Identity

The identity of this product was established using the following analyses:

Ia. ¹HNMR Spectrum

Conditions: 400 MHz, DMSO-d₆

¹HNMR spectrum consistent with chemical structure.



EPL-AA204 Batch 1


Revision 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

Ib. Mass Spectrum

The mass spectrum of this material was analysed by Liquid Chromatography Mass Spectroscopy (LCMS) using in-house EM005.WI08.

Method: 5% to 100% ACN in water gradient (+0.1% formic acid)
Poroshell 120 EC-C18, 4.6 x 50 mm, 2.7 micron



Theoretical value: 295.2 [M+H]⁺.

The signal of the Mass Spectrum is consistent with the theoretical value and its interpretation is consistent with the structural formula.


EPL-AA204 Batch 1

Revision 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

Ic. IR Spectrum

The infra-red spectrum of this material was analysed by Fourier-Transform Infrared Spectroscopy (FTIR) using in-house EM005.WI09.



The interpretation of the signals of the Fourier-Transform Infrared Spectrum is consistent with the structural formula.

EPL-AA204 Batch 1

Revision 1


Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

II. Purity

The purity of this material was analysed by high performance liquid chromatography (HPLC) using in-house EM005.WI07.

HPLC Conditions:

Column	Conditions				Detector	Injector
Agilent Poroshell 120 EC-C18 4.6 x 50mm 2.7 micron	25°C				DAD 288nm	Auto 1.0 µL 0.4 mg/mL in 100% acetonitrile (NO MODIFIERS)
	Time (min)	% Line A (Water + 0.1% (v/v) TFA)	% Line B (Acetonitrile + 0.1% (v/v) TFA)	Flow rate (mL/min)		
	0.00	95	5	1.0		
	6.00	71	29	1.0		
	12.60	5	95	1.0		
	17.60	5	95	1.0		
	18.60	95	5	1.0		
	21.60	95	5	1.0		



EPL-AA204 Batch 1

Revision 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

Area Percent Report – Sorted by Signal

Peak Number	Retention Time (rounded)	Area	Area % (rounded)
1	4.18	0.06	0.01
2	4.62	0.02	0.00
3	4.84	0.03	0.01
4	5.00	0.11	0.02
5	5.50	426.36	99.27
6	5.83	0.01	0.00
7	6.05	0.62	0.14
8	6.75	0.04	0.01
9	6.99	0.38	0.09
10	8.11	0.18	0.04
11	8.22	1.66	0.39
12	10.83	0.03	0.01
13	17.14	0.01	0.00
Totals			100 (rounded)

For the calculation the system peaks were ignored. The content of the analyte was determined as a ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100%.

Results:

Average 99.3% (average of 10 duplicate analyses)

III. Water Content

Method: Karl-Fischer titration using in-house EM005.WI04.

Results:

Average 2.0%

IV. Ash Content

Method: BP2016 Ash (Appendix XI J) as per WS001/C29372

Result:

Contains <0.1% ash.

V. Residual Solvents

Method: ¹H NMR

Result:

Contains 0.1% isopropanol by ¹H NMR analysis.

VI. Final Result

Chromatographic purity (HPLC)	99.3%
Water content	2.0%
Ash content	<0.1%
Residual solvents	0.1%
Purity*	97.2%

This purity is assessed to be 97.2%.

Product Reviewed By:

Product Released By:

James Rixson, PhD
Head of Production

Boon Tan
Quality Manager

Release Date: 16 November 2020

*NATA accreditation does not cover the performance of this service.

The calculation of the purity follows the formula:

$$\text{Purity(\%)} = \frac{((\text{Chromatographicpurity[HPLC]}) \times (100 - (\text{watercontent} + \text{ashcontent} + \text{volatilecontents})))}{100}$$

EPL-AA204 Batch 1

Revision 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902