



Accredited for compliance with ISO17034 and ISO/IEC17025. This document shall not be reproduced except in full. Accreditation Number: 20126

NATA is a signatory to the ILAC Mutual Recognition Arrangement for the mutual recognition of the equivalence of reference materials certificates.



Our Formula. Your Success.

Reference Material Product Information Sheet

Epichem's Quality System conforms to ISO9001:2015 as certified by ECAAS Pty Ltd - Certification number 616061.

	F			
Name	4-(dimethylamino)-1-((1RS)-1-(3-(dimethylamino)propyl)-1-(4-fluorophenyl)-1,3-dihydroisobenzofuran-5-yl)butan-1-one dihydrobromide			
BP/EP Name	Citalopram Impurity G			
Epichem Item #	EPL-AA214 Batch 3			
CAS#	2575516-42-4			
Molecular Formula	C ₂₅ H ₃₃ FN ₂ O ₂ .HBr.HBr			
Molecular Weight	574.38 g/mol			
Appearance	Off-white powder			
Melting Point	210.6-217.5°C (decomposition)			
Combustion Analysis	Required (%): C:52.3; H:6.1; N:4.9. Found (%): C:52.2; H:6.3; N:4.6.			
Purity*	96.6%			
Date of Manufacture	16 September 2019			
Storage Requirements	Protect from heat, light and moisture.			
Special Precautions	This compound is for laboratory use only. Its toxicological properties may not have been fully established. It should be handled only by suitably qualified personnel.			
Intended Use	This compound is suitable for the identification of impurities and degradants in pharmaceutical materials. The purity assay is considered as relative contribution.			
Date of Shipment	TBA			
	This certificate is valid for one year from the date of shipment provided the substance is unopened and stored under the recommended conditions.			
Retest Date	TBA (Proper Storage and Handling Required)			

^{*} NATA accreditation does not cover the performance of this service

EPL-AA214 Batch 3

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia ABN 80 106 769 902 Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au

Form PC008.F07 **Product Information Sheet** Page 1 of 7


I. Identity

The identity of this product was established using the following analyses:

Ia. ¹HNMR Spectrum

Conditions: 400 MHz, DO

¹H NMR spectrum consistent with chemical structure.




Ib. Mass Spectrum

The mass spectrum of this material was analysed by Liquid Chromatography Mass Spectroscopy (LCMS) using inhouse EM005.WI08.

Method: ACN/water gradient (+ 0.1% formic acid).

ZORBAX SB-C8, 4.6 x 30 mm, 3.5 micron.

Retention		Mol. Weight
Time (MS)	MS Area	or Ion
2.082	22308098	413.25 I
		207.15 I




Theoretical value: 207.1 [M+H]+.

The signal of the Mass Spectrum is consistent with the theoretical value and its interpretation is consistent with the structural formula.

Form PC008.F07 Product Information Sheet Page 3 of 7

Ic. IR Spectrum

The infra-red spectrum of this material was analysed by Fourier-Transform Infrared Spectroscopy (FTIR) using inhouse EM005.WI09.




The interpretation of the signals of the Fourier-Transform Infrared Spectrum is consistent with the structural formula.

II. Purity

The purity of this material was analysed by high performance liquid chromatography (HPLC) using in-house EM005.WI07.

HPLC Conditions:

Column	Conditi	Conditions			Detector	Injector
Agilent Poroshell						
120 EC-C18	Time	% Line A (Water +	% Line B (Acetonitrile	Flow rate		1.0 μL
4.6 x 50mm	(min)	0.1% (v/v) TFA)	+ 0.1% (v/v) TFA)	(mL/min)		0.3 mg/mL in
no a somm	0.00	90	10	1.0		100% acetonitrile
2.7 micron	7.00	69	31	1.0		(NO MODIFIERS)
	13.40	5	95	1.0		
	18.40	5	95	1.0		
	19.40	90	10	1.0		
	22.40	90	10	1.0		



EPL-AA214 Batch 3

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902


Area Percent Report - Sorted by Signal

Peak Number	Retention Time (rounded)	Area	Area % (rounded)
1	5.72	0.04	0.01
2	5.82	0.14	0.04
3	6.06	0.20	0.06
4	6.17	4.17	1.28
5	6.27	320.92	98.39
6	6.60	0.08	0.03
7	6.73	0.03	0.01
8	7.85	0.25	0.08
9	7.90	0.34	0.11
Totals			100 (rounded)

For the calculation the system peaks were ignored. The content of the analyte was determined as a ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100%.

Results:

Average 98.4% (average of 10 duplicate analyses)



Form PC008.F07 Product Information Sheet Page 6 of 7

III. Water Content

Method: Karl-Fischer titration using in-house EM005.WI04.

Results:

Average 0.4%

IV. Ash Content

Method: BP 2019 Appendix XI J Method II

Result:

Contains 0.2% ash.

V. Residual Solvents

Method: ¹HNMR

Result:

Contains 0.6% acetone and 0.6% isopropanol by ¹H NMR analysis.

VI. Final Result

Chromatographic purity (HPLC)	98.4%
Water content	0.4%
Ash content	0.2%
Residual solvents	1.2%
Purity*	96.6%

This purity is assessed to be 96.6%.

Product Reviewed By: Product Released By:

James Rixson, PhD Carol Worth, PhD **Head of Production** Quality Manager

Release Date: 10 October 2022

The calculation of the purity follows the formula:

((Chromatographic purity [HPLC])x(100-(water content+a shcontent+volatile contents)))

EPL-AA214 Batch 3

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia www.epichem.com.au Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 ABN 80 106 769 902

^{*}NATA accreditation does not cover the performance of this service.