


Reference Material Product Information Sheet

Epichem's Quality System conforms to ISO9001:2015 as certified by ECAAS Pty Ltd - Certification number 616061.



Name	8-chloro-6,11-dihydro-11-(1-methyl-4-piperidinylidene)-5H-Benzo[5,6]cyclohepta[1,2-b]pyridine
BP/EP Name	Loratadine Impurity G
USP Name	Loratadine Related Compound B
Synonym(s)	8-Chloroazatadine; N-Methyl-desloratadine; 8-Chloro-11-(1-methylpiperidin-4-ylidene)-6,11-dihydro-5H-benzo[5,6]cyclohepta[1,2-b]pyridine
Epichem Item #	EPL-AA279 Batch 1
CAS #	38092-89-6
Molecular Formula	C ₂₀ H ₂₁ ClN ₂
Molecular Weight	324.86 g/mol
Appearance	White powder
Melting Point	117.8-120.3°C
Combustion Analysis	Required (%): C:74.0; H:6.5; N:8.6. Found (%): C:73.8; H:6.6; N:8.6.
Purity*	99.6%
Date of Manufacture	15 March 2021
Storage Requirements	Protect from heat, light and moisture.
Special Precautions	This compound is for laboratory use only. Its toxicological properties may not have been fully established. It should be handled only by suitably qualified personnel.
Intended Use	This compound is suitable for the identification of impurities and degradants in pharmaceutical materials. The purity assay is considered as relative contribution.
Date of Shipment	TBA This certificate is valid for one year from the date of shipment provided the substance is unopened and stored under the recommended conditions.
Retest Date	TBA (Proper Storage and Handling Required)

* NATA accreditation does not cover the performance of this service

EPL-AA279 Batch 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia

Tel + 61 (0)8 6167 5200

Fax + 61 (0)8 6167 5201

www.epichem.com.au

ABN 80 106 769 902

I. Identity

The identity of this product was established using the following analyses:


Ia. ¹H NMR Spectrum

Conditions: 400 MHz, CD₃CO

¹H NMR spectrum consistent with chemical structure.

Acquisition Time (sec)	3.7547	Comment	LBC398-173@9H 1H MeOD (E:\dataexternal\epichem) cygoh 8	Date	11 Feb 2021 17:18:56		
Date Stamp	11 Feb 2021 17:18:56	File Name	\naphthalene\company\NMR files\LBC398-173@9H_CD3OD\1\data\111r				
Frequency (MHz)	400.13	Nucleus	1H	Number of Transients	8	Origin	spect
Original Points Count	24038	Owner	nmr	Points Count	32768	Pulse Sequence	zg
Receiver Gain	114.00	SW(cyclical) (Hz)	6402.05	Solvent	METHANOL-d4	Spectrum Offset (Hz)	2792.9836
Spectrum Type	STANDARD	Sweep Width (Hz)	6401.85	Temperature (degree C)	24.237		

LBC398-173@9H_CD3OD.001.001.1r.esp



EPL-AA279 Batch 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902


Ib. Mass Spectrum

The mass spectrum of this material was analysed by Liquid Chromatography Mass Spectroscopy (LCMS) using in-house EM005.WI08.

Method: ACN/water gradient (+ 0.1% formic acid).

ZORBAX SB-C8, 4.6 x 30 mm, 3.5 micron.

Retention Time (MS)	MS Area	Mol. Weight or Ion
2.175	14926618	327.10 I
		326.15 I
		325.10 I
		163.90 I
		163.10 I



Theoretical value: 325.1 [M+H]⁺.


The signal of the Mass Spectrum is consistent with the theoretical value and its interpretation is consistent with the structural formula.

EPL-AA279 Batch 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

Ic. IR Spectrum

The infra-red spectrum of this material was analysed by Fourier-Transform Infrared Spectroscopy (FTIR) using in-house EM005.WI09.



The interpretation of the signals of the Fourier-Transform Infrared Spectrum is consistent with the structural formula.

EPL-AA279 Batch 1


Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

II. Purity

The purity of this material was analysed by high performance liquid chromatography (HPLC) using in-house EM005.WI07.

HPLC Conditions:

Column	Conditions				Detector	Injector
Agilent ZORBAX Eclipse XDB-C18 4.6 x 100mm 3.5 micron	45°C				DAD 282nm	Auto 1.0 µL 0.3 mg/mL in 100% acetonitrile (NO MODIFIERS)
	Time (min)	% Line A (Water + 0.1% (v/v) TFA)	% Line B (Acetonitrile + 0.1% (v/v) TFA)	Flow rate (mL/min)		
	0.00	95	5	1.0		
	8.00	65	35	1.0		
	14.00	5	95	1.0		
	19.00	5	95	1.0		
	20.00	95	5	1.0		
	24.00	95	5	1.0		



EPL-AA279 Batch 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

Area Percent Report – Sorted by Signal

Peak Number	Retention Time (rounded)	Area	Area % (rounded)
1	7.00	0.33	0.08
2	7.08	412.67	99.85
3	7.54	0.14	0.03
4	7.66	0.03	0.01
5	7.73	0.08	0.02
6	9.36	0.04	0.01
Totals			100 (rounded)

For the calculation the system peaks were ignored. The content of the analyte was determined as a ratio of the peak area of the analyte and the cumulative areas of the purities, added up to 100%.

Results:

Average 99.8% (average of 10 duplicate analyses)

EPL-AA279 Batch 1

Epichem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epichem.com.au ABN 80 106 769 902

III. Water Content

Method: Karl-Fischer titration using in-house EM005.WI04.

Results:

Average 0.1%

IV. Ash Content

Method: BP2020 Ash Appendix XI J Method II

Result:

Contains <0.1% ash.

V. Residual Solvents

Method: ¹HNMR

Result:

No significant impurities detected by ¹H NMR analysis.

VI. Final Result

Chromatographic purity (HPLC)	99.7%
Water content	0.1%
Ash content	<0.1%
Residual solvents	<0.1%
Purity*	99.6%

This purity is assessed to be 99.6%.

Product Reviewed By:

Product Released By:

James Rixson, PhD
Head of Production

Carol Worth, PhD
Quality Manager

Release Date: 22 July 2022

*NATA accreditation does not cover the performance of this service.
The calculation of the purity follows the formula:

$$\text{Purity(\%)} = \frac{((\text{Chromatographicpurity[HPLC]}) \times (100 - (\text{watercontent} + \text{ashcontent} + \text{volatilecontents})))}{100}$$

EPL-AA279 Batch 1

Epicchem Pty Ltd, Suite 5, 3 Brodie-Hall Drive, Bentley WA 6102, Australia
Tel + 61 (0)8 6167 5200 Fax + 61 (0)8 6167 5201 www.epicchem.com.au ABN 80 106 769 902