

NuMega Resonance Labs

Fast Reliable Services

11526 Sorrento Valley Road, Suite B-2 ♦ San Diego, CA 92121 ♦ USA
Phone 858-793-6057 ♦ Fax 858-793-2607

December 11, 2023

NuMega Reference #

2310NERO.587B

Carrie Treat
Neurogan Inc
8821 Production Ave
San Diego, CA 92121
clt@neurogan.com

Quantitative Analysis Report

Bruker 500 MHz, AVII-500 and AV-500

Sample ID	Spermidine
Solvent	D2O
Standard	Ethylene Carbonate
Weight of Standard (wt _{st})	1.891
Weight of Sample (wt _{sa})	15.145
Sample Molecular Weight (MW _{sa})	254.63
Standard Molecular Weight (MW _{st})	88.06
Integration of Sample (I _{sa})	138.34
Integration of Standard (I _{st})	100
# of Equivalent Nuclei in Standard (n _{st})	4
# of Equivalent Nuclei in Sample (n _{sa})	2
Assigned Purity of Standard % (P)	99.92
Purity % *	99.8%

Formula Used for Calculation:
$$WT\% = \frac{wt_{st}}{wt_{sa}} \times \frac{MW_{sa}}{MW_{st}} \times \frac{I_{sa}}{I_{st}} \times \frac{n_{st}}{n_{sa}} \times P$$