

ISO9001 & ISO13485 & ISO11137

Product Name:	Examination Gloves			
Product Application:	NEST gloves can be used in household, industrial, medical, beauty and other			
	industries, and it is an essential hand protection product.			
Product Materials:	Latex / Nitrile + colloidal oatmeal			
product Specifications:	Temperature range: stored at room temperature			
	Shelf life: 3 Years after date of production (ensure package is in good)			
Sterilization:	No			
BSE/TSE:	These products are deemed animal free.			
Performance Testing:	Each manufacturing lot is sampled and tested in accordance with standard operating procedures.			
	Appearance inspection: qualified Tensile Strength Test: qualified			
	Leak test: qualified Packaging inspection: qualified			
Features:	 Single use only Made in Malaysia 			
	 Ambidextrous: Can be worn left and right, regardless of left or right hand 			
	 Glove mouth rolled edge design, easy to wear, prevent rollback 			
	 Textured outer surface of the glove increases friction and facilitates gripping 			
	6. Powder-Free			
	7. Non-Sterile			
	8. Boxed for easy extraction.			
	9. Available in three materials			
	(1) Nitrile Examination Glove:			
	Tensile Strength: Minimum 18.0 MPa (Before Aging)			
	• Palm thickness: 0.07 mm			
	(2) Latex Examination Glove:			
	 Natural rubber latex and compounding materials 			
	Tensile Strength: Minimum 18.0 MPa (Before Aging)			
	Palm thickness: 0.11 mm			
	(3) Colloidal Oatmeal Coated Nitrile Glove			
	Colloidal Oatmeal Coated Latex Glove			
	Oats Extractions			
	Oats extractions is a patented coating recognized by the FDA as a skin protectant, whic			
	have been proven by clinical studies could reduce and control rash and dry skin			
	symptoms. Clinical studies have shown that colloidal oatmeal could reduce and control			

DATA SHEET





ISO9001 & ISO13485 & ISO11137

rash, dry skin.

Product Range

Nitrile Examination Glove

Cat. No.	Size	Palm Width (mm) ±4	Length	/Inner Box	/Case
902001	XS	76	Min 230mm (ASTM)	100	1000
902011	S	86		100	1000
902021	М	98		100	1000
902031	L	107		100	1000
902041	XL	115		90	900

Latex Examination Glove

Cat. No.	Size	Palm Width (mm) ±4	Length	/Inner Box	/Case
901001	XS	76	Min 230mm	100	1000
901011	S	86	(ASTM)	100	1000
901021	м	98		100	1000
901031	L	107	-	100	1000
901041	XL	115		90	900

Colloidal Oatmeal Coated Nitrile Glove

Cat. No.	Size	Palm Width (mm) ±4	Length	/Inner Box	/Case
903011	S	84	Min 230mm (ASTM)	100	1000
903021	М	94		100	1000
903031	L	105		100	1000

Colloidal Oatmeal Coated Latex Glove

Cat. No.	Size	Palm Width (mm) ±4	Length	/Inner Box	/Case
904011	S	84	Min 230mm	100	1000
904021	М	94	(ASTM)	100	1000
904031	L	105		100	1000

Head Office No. 530, Xida Road, Meicun Industrial Park, Xinwu District, Wuxi, Jiangsu, China Tel: +86+ 510-6800 6788 Email: info@nest-wuxi.com Online: www.cell-nest.com Oversea

NEST USA (New Jersey/ Phoenix) NEST scientific 株式会社 (Yokohama,Japan) NEST Scientific Europe B.V (Netherlands) Nest Scientific (MENA) FZE (Sharjah, United Arab Emirates)





NEST USA (New Jersey/ Phoenix)

Emirates)

NEST scientific 株式会社 (Yokohama, Japan)

Nest Scientific (MENA) FZE (Sharjah, United Arab

NEST Scientific Europe B.V (Netherlands)

Oversea

Product Picture

Nitrile Examination Glove



Due to chromatic aberration of the display screen, there may be chromatic aberration between the picture and the real object.

Latex Examination Glove



The picture is only for reference, please make the object as the standard.