

DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-24957 Mfr. Reference: BNX005 Centers for Disease Control and Prevention (CDC) National Institute for Occupational Safety and Health (NIOSH) National Personal Protective Technology Laboratory (NPPTL) 626 Cochrans Mill Road Pittsburgh, PA 15236-0070 Phone: 412-386-4000 Fax: 412-386-4051 July 21, 2021

Simon Ma Chief Technology Officer BNX Converting, LLC 16727 Park Row Houston, TX 77084

Dear Simon Ma:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted June 3, 2021. This request was for approval of the model C95W (white) N95 air-purifying filtering facepiece respirator. The complete respirator configuration is detailed on assembly matrix file name *1115AMb.xls*, revision B, dated: 06/21/2021.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired to be produced in languages other than English. Approval number TC-84A-9363 has been assigned. This respirator is approved for protection against particulates at N95 filter efficiency level.

The approval label is included as an attachment to this letter. The abbreviated label has been accepted as submitted. The cautions and limitations are listed on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely,

Jeffrey Peterson Chief, Conformity Verification and Standards Development Branch

Enclosures



National Institute for Occupational Safety and Health National Personal Protective Technology Laboratory Evaluation and Testing Branch

TEST REPORT

Task Number: TN-24957

Manufacturer: BNX Converting LLC

Prepared by: Nathan Smichnick

Date: 07/14/2021

Tests Conducted by: Nathan Smichnick

Respirator Tested: C95W

Tests Completed

Test Description	STP Number
A. Exhalation Resistance TestB. Inhalation Resistance TestC. Sodium Chloride (NaCl) N95 Test	TEB-APR-STP-0003 TEB-APR-STP-0007 TEB-APR-STP-0059

Overall Results

The respirator system tested met the requirements of all the above procedures.