



Certificate of Analysis

Product Item : Nylon050G045I-A

Membrane : Nylon membrane with Quartz pre-filter

Diameter : 65mm

Pore Size : 0.45um

Housing material : PP

Lot.: 0317GG16W

Page : 1/1

Date : 2016/03/15

Item	Specification
Bio-Safety	USP Class VI Biological Test for plastics
Connection	Inlet: Stepped Hose Barb with female Luer slip interior Outlet: Stepped Hose Barb with female Luer slip interior
Work condition	<350kPa
Water Entry Pressure	5 Bar
Air flow rate	p=1bar : 18-20L/min
Hepa Test	Filtration Efficiency: >99.95% (PSL 0.3um)

The information is based on our present state of knowledge and is intended to provide general notes on our products on their use. It should therefore not be constructed as guaranteeing specific properties of the products described or their suitability for a particular application. Any existing industrial property rights must be observed. The quality of our products guaranteed under our General Condition of Sale.