



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

NIOSH Reference: TN-17685  
Mfr. Reference: MAK-1105

Centers for Disease Control  
and Prevention (CDC)  
National Institute for Occupational  
Safety and Health (NIOSH)  
National Personal Protective  
Technology Laboratory (NPPTL)  
P.O. Box 18070  
Pittsburgh, PA 15236-0070  
Phone: 412-386-4000  
Fax: 412-386-4051

April 19, 2011

Mr. Alexander Freedman, President  
Makrite Industries Inc.  
236 Upland Avenue  
Newton Highlands, Massachusetts 02461-2003

Dear Mr. Freedman:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted March 8, 2011. This request was for an approval of the model 9500-N95 filtering facepiece air purifying respirator for protections against particulates at a N95 filter efficiency level, reference the assembly matrix MAK1105AM1.xls.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. The approval number TC-84A-5411 has been assigned. This respirator is approved for protections against particulates at a N95 filter efficiency level (N95).

The CD enclosed with this letter contains the final respirator label. The abbreviated label has been accepted as submitted. The cautions and limitations which apply to this approval are on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.


The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that the respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

Page 2 – Mr. Alexander Freedman – TN-17685

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely yours,



Heinz W. Ahlers  
Chief, Technology Evaluation Branch  
National Personal Protective Technology Laboratory

Enclosures