



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

NIOSH Reference: TN-17766
Mfr. Reference: MAK-1109

Centers for Disease Control
and Prevention (CDC)
National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
P.O. Box 18070
Pittsburgh, PA 15236-0070
Phone: 412-386-4000
Fax: 412-386-4051
June 2, 2011

Mr. Alexander Freedman, President
Makrite Industries Inc.
236 Upland Avenue
Newton Highlands, Massachusetts 02461-2003

Dear Mr. Freedman:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted April 20, 2011. This request was for an approval of the model 9500V-N95 filtering facepiece air purifying respirator with exhalation valve for protections against particulates at a N95 filter efficiency level, reference the assembly matrix MAK1109AM1.xls.

This request is granted. Approvals are granted only for documentation written in the English language. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. The approval number TC-84A-5460 has been assigned. This respirator is approved for protections against particulates at a N95 filter efficiency level (N95).

The CD enclosed with this letter contains the final respirator label. The abbreviated label has been accepted as submitted. The cautions and limitations which apply to this approval are on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that the respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

Page 2 – Mr. Alexander Freedman – TN-17766

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Heinz W. Ahlers", written over a horizontal line.

Heinz W. Ahlers
Chief, Technology Evaluation Branch
National Personal Protective Technology Laboratory

Enclosures