

# Lumify iOS Device Compatibility List

Memo 261705 G



Devices new to this list revision are in **RED**

Device Name	Year Released	Screen Size (in.)	Connectivity Options	Internal Storage (in GB)	Philips Case Available
iPhone 13 mini	2021	5.4 in	Wi-Fi / LTE	128/256/512	No
iPhone 13	2021	6.1 in	Wi-Fi / LTE	128/256/512	No
iPhone 13 Pro	2021	6.1 in	Wi-Fi / LTE	128/256/512 and 1TB	No
iPhone 13 Pro Max	2021	6.7 in	Wi-Fi / LTE	128/256/512 and 1TB	No
iPad (9 <sup>th</sup> gen)	2021	10.2	Wi-Fi / LTE	64/256	Yes
iPhone 12 Pro Max	2020	6.7	Wi-Fi / LTE	128/256/512	No
iPhone 12 Mini	2020	5.4	Wi-Fi / LTE	64/128/256	No
iPad (8 <sup>th</sup> gen)	2020	10.2	Wi-Fi / LTE	32/128	Yes
iPhone 12 Pro	2020	6.1	Wi-Fi / LTE	128/256/512	No
iPhone 12	2020	6.1	Wi-Fi / LTE	64/128/256	No
iPhone 11 Pro Max	2019	6.5	Wi-Fi / LTE	64/256/512	No
iPhone 11 Pro	2019	5.8	Wi-Fi / LTE	64/256/512	No
iPhone 11	2019	6.1	Wi-Fi / LTE	64/128/256	Yes
iPhone XS Max	2018	6.5	Wi-Fi / LTE	64/256/512	No
iPhone XS	2018	5.8	Wi-Fi / LTE	64/256/512	Yes
iPhone XR	2018	6.1	Wi-Fi / LTE	64/256/512	Yes
iPhone X	2017	5.8	Wi-Fi / LTE	64/256	Yes
iPhone 8 Plus	2017	5.5	Wi-Fi / LTE	64/256	No
iPhone 8	2017	4.7	Wi-Fi / LTE	64/256	Yes
iPhone 7 Plus	2016	5.5	Wi-Fi / LTE	32/128/256	No
iPhone 7	2016	4.7	Wi-Fi / LTE	32/128/256	Yes
iPad (7 <sup>th</sup> gen)	2019	10.2	Wi-Fi / LTE	32/128	Yes
iPad Air (3 <sup>rd</sup> gen)	2019	10.5	Wi-Fi / LTE	64 / 256	No



**Philips Ultrasound, Inc.**

22100 Bothell Everett Hwy

Bothell, WA 98021-8431, USA

[www.healthcare.philips.com/ultrasound](http://www.healthcare.philips.com/ultrasound)

© Koninklijke Philips N.V. 2017. All rights reserved. Printed in USA.

Reproduction or transmission in whole or in part, in any form or by any means, electronic, mechanical, or otherwise, is prohibited without the prior written consent of the copyright owner.  
261705 Rev G -- May 2022

# Lumify iOS Device Compatibility List

Memo 261705 G



Device Name	Year Released	Screen Size (in.)	Connectivity Options	Internal Storage (in GB)	Philips Case Available
iPad Mini 5	2019	7.9	Wi-Fi / LTE	64/256	Yes
iPad (6 <sup>th</sup> gen)	2018	9.7	Wi-Fi / LTE	32/128	Yes
iPad (5 <sup>th</sup> gen)	2017	9.7	Wi-Fi / LTE	32/128	Yes
iPad Pro (2 <sup>nd</sup> gen)	2017	12.9	Wi-Fi / LTE	64/256/512	No
iPad Pro (2 <sup>nd</sup> gen)	2017	10.5	Wi-Fi / LTE	64/256/512	No

## Supported Devices\*

Please note that Reacts tele-ultrasound solution is not available on Lumify when using the above iOS devices. All iOS devices currently compatible require use of Lumify Power Module and utilize USB-C transducer cables and will all continuously scan for 2 to 2.5 hrs.

Devices on the supported list satisfy all of the following criteria: 1. Passed software testing on the latest Lumify software, 2. Passed Philips hardware testing, 3. Is readily available from the device manufacturer.

Philips cannot test every device available on the market, so this list represents a range of devices to satisfy a variety of user needs with respect to form factor, price, battery life and connectivity options.



**Philips Ultrasound, Inc.**  
22100 Bothell Everett Hwy  
Bothell, WA 98021-8431, USA  
[www.healthcare.philips.com/ultrasound](http://www.healthcare.philips.com/ultrasound)

© Koninklijke Philips N.V. 2017. All rights reserved. Printed in USA.  
Reproduction or transmission in whole or in part, in any form or by any means, electronic, mechanical, or otherwise, is prohibited without the prior written consent of the copyright owner.  
261705 Rev G -- May 2022

# Lumify iOS Device Compatibility List

Memo 261705 G



## No Longer Supported\*\*

Device Name	Year Released	Screen Size (in.)	Connectivity Options	Internal Storage (in GB)	Philips Case Available

\*\*The devices on this list represent devices that at one point supported the Lumify system but no longer do for one or more of the following reasons: 1. Did not pass software testing on the latest Lumify software, 2. Did not pass Philips hardware testing, or 3. Is no longer available from the manufacturer. Lumify may still run on these devices, but due to the above criteria, Philips can no longer verify Lumify compatibility

Tablet Compatibility is subject to change, and Philips reserves the right to make changes without notice. Devices listed under “No Longer Supported” are not tested for use with the current version of Lumify by Philips.



**Philips Ultrasound, Inc.**  
22100 Bothell Everett Hwy  
Bothell, WA 98021-8431, USA  
[www.healthcare.philips.com/ultrasound](http://www.healthcare.philips.com/ultrasound)

© Koninklijke Philips N.V. 2017. All rights reserved. Printed in USA.  
Reproduction or transmission in whole or in part, in any form or by any means, electronic, mechanical, or otherwise, is prohibited without the prior written consent of the copyright owner.  
261705 Rev G -- May 2022