

(This Report is issued subject to the terms & conditions set out below)



Setsco Services Pte Ltd 18 Teban Gardens Crescent Singapore 608925

Tel: (65) 6566 7777 Fax: (65) 6566 7718 www.setsco.com Business Reg. No, 196900269D

Your Ref: -

Our Ref: FB-8500163443/SS/4

Date: 09/06/2021

Page 1 of 3

Subject

: Differential pressure test for Granzilla Aircare 3 ply disposable face mask

(Black) submitted by Grantech Pte Ltd on 29/05/2021

Purpose

: The purpose of the test is to measure the differential pressure of a medical

face mask

Tested For

: GRANTECH PTE LTD

629 Aljunied Road

#04-17 Cititech Industrial Building

Singapore 389838 Attn: Mr SY Ong

Method of Test

: BS EN 14683 : 2019 Medical Face Masks - Requirements and Test Methods

Clause 5.2.3 Annex C Method for determination of breathability (differential

pressure)

Description of

Samples

50 pieces of Granzilla Aircare 3 ply disposable face mask (Black) were

received

(Refer to Photograph 1)

Lotcode

20210104

Model

SML-A21

Meltblown Used

HSL200826

F. 1/4

Khairuddin Testing Officer

Salim Suwignjo

Assistant Manager (CMTD)
Construction Technology Division

Terms & Conditions

(1) The Report is prepared for the sole use of the Client and is prepared based upon the Item submitted, the services required by the Client and the conditions under which the Services are performed by SETSCO. The Report is not intended to be representative of similar or equivalent Services on similar or equivalent items. The Report does not constitute an endorsement by SETSCO of the item.

(2) SETSCO agrees to use reasonable diligence in the performance of the Services but no warranties are given and done may be implied directly or indirectly relating to the Services, the Report or the facilities of SETSCO.

(3) SETSCO shall under no circumstances be liable to the Client or its agents, servants or representatives, in contract, tort (including negligence or breach of statutory duty) or otherwise for any direct or indirect loss or damage suffered by the Client, its agents, servants or representative howsoever arising or whether connected with the Services provided by SETSCO herein.

(4) The Report may not be used in any publicity material without the written consent of SETSCO.

(5) The Report shall not be reproduced in part or in full unless approval in writing has been given by SETSCO.



FB-8500163443/SS/4

Page 2 of 3


Test Result : Differential Pressure

Sample Reference		Granzilla Aircare 3 ply disposable face mask (Black)				
Conditioning		From 03/06/2021 to 04/06/2021 @ 21°C, 80%RH				
Date of Test		04/06/2021				
Test Condition		23°C, 52%RH				
Air Flow (L/min)		8				
Tested Area (cm ²)		4.9				
Type of Mask		Multi Material Mask				
Specimen Reference		S1	S2	S3	S4	S5
Differential Pressure, ΔP (Pa/cm²)	Top 1	30.00	22.86	26.73	26.53	35.10
	Top 2	33.27	24.49	26.53	27.14	35.92
	Тор 3	35.51	24.08	29.18	29.39	31.84
	Bottom 1	30.20	24.90	32.24	29.80	26.12
	Bottom 2	34.69	23.27	32.20	28.98	36.73
	Bottom 3	36.88	24.29	31.43	27.76	37.96
	Average	33.76	23.98	29.39	28.27	33.95
Requirement BS EN 14683 Clause 5.2.7 Summary of Performance Requirements		Differential Pressure, (Pa/cm²)				
		Type I < 40				
		Type II < 40				
	Type IIR < 60					





Photograph 1 Granzilla Aircare 3 ply disposable face mask (Black) as received



Photograph 2 Granzilla Aircare 3 ply disposable face mask (Black) Outer (Top) and Inner (Bottom)



