

TEST REPORT

(This Report is issued subject to the terms & conditions set out below)

SetSCO Services Pte Ltd
18 Teban Gardens Crescent
Singapore 608925
Tel : (65) 6566 7777
Fax: (65) 6566 7718
www.setsco.com
Business Reg. No. 196900269D

Your Ref: -

Our Ref: FB-8500163443/SS/5

Date: 09/06/2021

Page 1 of 3

Subject : Flammability test for Granzilla Aircare 3 ply disposable face mask (Black) submitted by Grantech Pte Ltd on 29/05/2021

Purpose : The purpose of the test is to determine the flammability of a medical face mask

Tested For : **GRANTECH PTE LTD**
629 Aljunied Road
#04-17 Cititech Industrial Building
Singapore 389838
Attn: Mr SY Ong

Method of Test : ASTM F 2100-19 Standard Specification for Performance of Materials Used in Medical Face Masks – Clause 6.2 / 9.5 Flammability ; Determine flammability as specified in 16 CFR 1610

Description of Samples : 50 pieces of Granzilla Aircare 3 ply disposable face mask (Black) were received
(Refer to Photograph 1)
Lotcode : 20210104
Model : SML-A21
Meltblown Used : HSL200826


Khairuddin
Testing Officer


Salim Suwignjo
Assistant Manager (CMTD)
Construction Technology Division

Terms & Conditions:

- (1) The Report is prepared for the sole use of the Client and is prepared based upon the Item submitted, the services required by the Client and the conditions under which the Services are performed by SETSCO. The Report is not intended to be representative of similar or equivalent Services on similar or equivalent items. The Report does not constitute an endorsement by SETSCO of the item.
- (2) SETSCO agrees to use reasonable diligence in the performance of the Services but no warranties are given and done may be implied directly or indirectly relating to the Services, the Report or the facilities of SETSCO.
- (3) SETSCO shall under no circumstances be liable to the Client or its agents, servants or representatives, in contract, tort (including negligence or breach of statutory duty) or otherwise for any direct or indirect loss or damage suffered by the Client, its agents, servants or representative howsoever arising or whether connected with the Services provided by SETSCO herein.
- (4) The Report may not be used in any publicity material without the written consent of SETSCO.
- (5) The Report shall not be reproduced in part or in full unless approval in writing has been given by SETSCO.



LA-1993-0051-C LA-1998-0144-D
LA-1993-0067-G LA-2000-0181-F
LA-1994-0066-A LA-2012-0519-E
LA-1987-0001-B LA-1987-0001-B-1

Test Result : Flammability

Sample Reference	Granzilla Aircare 3 ply disposable face mask (Black)				
Drying of Specimen	09/06/2021 From 1030hrs to 1045hrs @ 105°C				
Date of Test	09/06/2021				
Test Condition	23°C, 52%RH				
Specimen Reference	S1	S2	S3	S4	S5
Flame Application Time	1 second				
Burn Time (sec)	No burn time	No burn time	No burn time	No burn time	No burn time
Visual Observation	DNI	DNI	DNI	DNI	DNI
Classification	Class 1	Class 1	Class 1	Class 1	Class 1
Requirement Accordance to ASTM 2100-19	Flame Spread – Class 1				


REMARKS:

1. Specimens were tested in their original state
2. DNI = Did not ignite






Photograph 1 Granzilla Aircare 3 ply disposable face mask (Black) as received



Photograph 2 Granzilla Aircare 3 ply disposable face mask (Black) Inner (Top) and Outer (Bottom)

