Test Report No. 7191244814-EEC20-04-WBH dated 02 Oct 2020

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.



Add value. Inspire trust.

SUBJECT

Flammability Test

CLIENT

Grantech Pte Ltd 629 Aljunied Road #04-17 Cititech Industrial Building Singapore 389838

Attn: Peter Chong It Kong

SAMPLE SUBMISSION DATE/ TEST DATE

22 Sep 2020 / 01 Oct 2020

DESCRIPTION OF SAMPLES

One sample described as below was received:

Product Name / Brand Name	Manufacturer	Lot Number
GRANZILLA AIRCARE 3 Ply Disposable Face Mask	Grantech Pte Ltd	Test Sample



Laboratory: TÜV SÜD PSB Pte. Ltd. No.1 Science Park Drive Singapore 118221 Phone: +65-6885 1333 Fax: +65-6776 8670 E-mail: enquiries@tuvsud.com https://www.tuvsud.com/en-sg Co. Reg: 199002667R

Regional Head Office: TÜV SÜD Asia Pacific Pte. Ltd. 1 Science Park Drive, #02-01 Singapore 118221

Test Report No. 7191244814-EEC20-04-WBH dated 02 Oct 2020



DESCRIPTION OF SAMPLE (cont'd)



Figure 1: "GRANZILLA AIRCARE 3 Ply Disposable Face Mask" sample as received

METHOD OF TEST:

ASTM F2100-19 Standard Specification for Performance of Materials Used in Medical Face Masks -Clause 6.2/ 9.5 Flammability (test as specified in 16 CFR 1610)

RESULTS

Sample Description: GRANZILLA AIRCARE 3 Ply Disposable Face Mask

Type of Fabric: Plain Surface Textile Fabrics

Flammability Test / 16 CFR 1610			Requirement / ASTM	Inferred	
S/N	Test Surface	Test Code	Test Classification	F2100 Clause 6.2	Result
1	Front	DNI	Class 1	Shall meet the	
2	Front	DNI	Class 1	requirements for Class 1, normal	
3	Front	DNI	Class 1	flammability	Pass
4	Back	DNI	Class 1	specified in 16 CFR	1
5	Back	DNI	Class 1	Part 1610	

REMARKS

- 1. The above test results relate to the samples as received.
- 2. Specimens were tested in their original state.
- 3. DNI = Did not ignite
- 4. Both directions of the fabrics were observed with same burning characteristic.

William Wei

Associate Engineer

Wong Bee Hui Product Manager

Medical Health Services (NAM)

Test Report No. 7191244814-EEC20-04-WBH dated 02 Oct 2020



Please note that this Report is issued under the following terms:

- 1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
- The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
- 3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
- 4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
- 5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.
- 6. The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

