FDC38 Chair Placement Template

firstar

ALERT SYMBOL KEY The safety notices associated with the following should be given special attention when they appear in maintenance, operating and emergency procedures in the guide.

⚠ CAUTION indicates that a particular procedure or precaution must be followed to avoid system irregularity or possible damage to the product.

ⓘ NOTE indicates special information to improve the ease of maintaining the product, or to clarify important information.

⚠ CAUTION: Emailed, faxed or copied templates may jeopardize precise full scale measurements.

⚠ CAUTION: Local regulation requires licensed plumbers and electricians to install the utilities. All plumbing and utilities must conform to prevailing local codes.

ⓘ NOTE: Distance from edge of baseplate to junction box: 0 to 3" (0 - 7.62 CM)



⚠ CAUTION: Emailed, faxed or copied templates may jeopardize precise full scale measurements.

FIRSTAR DENTAL CO. FIRSTARDENTAL.COM
18915 E VALLEY HWY KENT, WA 98032, USA

Questions? TECHNICAL@FIRSTARDENTAL.COM