

Rad Hemp Co.
 2185 E. 74th Place
 Denver, CO 80229
 sales@radhempco.com
 973-960-6579

Sample: 03-21-2023-31455
 Sample Received: 03/21/2023;
 Report Created: 03/24/2023; Expires: 03/21/2024

Apple Tart
 Plant, Flower - Cured



15.722 %

Total THC

0.285 %

Δ-9 THC

19.172 %
 Total Cannabinoids

<LOQ %
 Total CBD

Cannabinoids

(Testing Method: HPLC, CON-P-3000)
 Date Tested: 03/21/2023

Complete

Analyte	LOD	LOQ	Mass	Mass	
	%	%	%	mg/g	
Δ-8-Tetrahydrocannabinol (Δ-8 THC)	0.0476	0.0714	ND	ND	
Δ-9-Tetrahydrocannabinol (Δ-9 THC)	0.0476	0.0714	0.285	2.848	
Δ-9-Tetrahydrocannabinolic Acid (THCA-A)	0.0476	0.0714	17.602	176.019	
Δ-9-Tetrahydrocannabinophorol (Δ-9-THCP)	0.0476	0.0714	ND	ND	
Δ-9-Tetrahydrocannabivarin (Δ-9-THCV)	0.0476	0.0714	ND	ND	
Δ-9-Tetrahydrocannabivarinic Acid (Δ-9-THCVA)	0.0476	0.0714	0.729	7.286	
R-Δ-10-Tetrahydrocannabinol (R-Δ-10-THC)	0.0476	0.0714	ND	ND	
S-Δ-10-Tetrahydrocannabinol (S-Δ-10-THC)	0.0476	0.0714	ND	ND	
9R-Hexahydrocannabinol (9R-HHC)	0.0476	0.0714	ND	ND	
9S-Hexahydrocannabinol (9S-HHC)	0.0476	0.0714	ND	ND	
Tetrahydrocannabinol Acetate (THCO)	0.0476	0.0714	ND	ND	
Cannabidivarin (CBDV)	0.0476	0.0714	ND	ND	
Cannabidivarinic Acid (CBDVA)	0.0476	0.0714	ND	ND	
Cannabidiol (CBD)	0.0476	0.0714	ND	ND	
Cannabidiolic Acid (CBDA)	0.0190	0.0714	<LOQ	<LOQ	
Cannabigerol (CBG)	0.0190	0.0714	<LOQ	<LOQ	
Cannabigerolic Acid (CBGA)	0.0476	0.0714	0.348	3.476	
Cannabinol (CBN)	0.0476	0.0714	ND	ND	
Cannabinolic Acid (CBNA)	0.0190	0.0714	<LOQ	<LOQ	
Cannabichromene (CBC)	0.0476	0.0714	ND	ND	
Cannabichromenic Acid (CBCA)	0.0476	0.0714	0.209	2.086	
Total			19.172	191.715	

Total THC = THCa * 0.877 + Δ9-THC; Total CBD = CBDa * 0.877 + CBD; LOQ = Limit of Quantitation; ND = Not Detected.

Total THC Measurement of Uncertainty: ± 0.040%
 Total CBD Measurement of Uncertainty: ± 2.000%
 THCO potency analysis does not designate quantitative specificity of Δ-8-THCO and Δ-9-THCO isomers



New Bloom Labs
 6121 Heritage Park Drive, A500
 Chattanooga, TN 37416
 (844) 837-8223
 TN DEA#: RN0563975
 ANAB Testing Laboratory (AT-2868): ISO/IEC
 17025:2017

Natalie Siracusa
 Natalie Siracusa
 Laboratory Director

Powered by
 reLIMS
 info@relims.com