AURELIA® IGNITE

7mil, Nitrile Powder Free Examination Gloves

AURELIA® GLOVES
A SUPERMAX HEALTHCARE BRAND
sales@supermax.co.uk
+441733 615215
Supermax Healthcare Ltd,
Unit 12-16 Titan Drive
Fengate, Peterborough
PE1 5XN
aureliagloves.co.uk

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1186: 2002. Materials and articles in contact with foodstuffs.

MD | UKCA 0120 | CE 0598 | AQL 1.5 | EN ISO 374-5: 2016 VIRUS | EN ISO 374-1: 2016/Type B JKPT | LATEX

10 x 100

*10 x 90

M	L	*XL	*XXL
97887	97888	97889	97880

Sizes & Dimensions

M	95+10 mm	min. 240 mm
L	110+10 mm	min. 240 mm
XL	120+10 mm	min. 240 mm
XXL	> 120 mm	min. 240 mm

Single wall thickness

① min. 0.25 mm
② min. 0.20 mm
③ min. 0.09 mm

Physical Properties

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N

Nitrile | Powder Free | 5 Years

Diamond | 8.4g +/- 0.2g | Orange

(c) SUPERMAX HEALTHCARE LIMITED 22nd September 2020
All Content provided is for information purposes only. All data and claims were correct at the time of writing. Supermax Healthcare Ltd makes no representations as to any product or testing updates that may occur after the data-sheet was written.