

## AffiCHECK® Influenza A H1 RNA PCR Qualitative Positive Control

AFG-CHK-0510

### PRODUCT DESCRIPTION:

**AffiCHECK® Influenza A H1 RNA - PCR Qualitative Control** is a high-quality, reliable standard containing known quantities of Influenza A H1 RNA. It's designed using modified and non-infectious virus to ensure stability, thereby contributing to accurate and reliable results in PCR applications.

### INTENDED USE:

**AffiCHECK® Influenza A H1 RNA - PCR Qualitative Control** serves as a comprehensive process control for the evaluation of the performance of nucleic acid tests for determination of the presence of Influenza A H1 nucleic acid. It's essential in validating the performance of PCR assays in clinical, research, and diagnostic settings.

### PRODUCT FEATURES:

Specificity: Tailored for Influenza A H1 RNA detection.

Formulation: Lyophilized for enhanced stability.

Convenience: Ready-to-use upon reconstitution.

Compatibility: Suitable for a wide range of PCR assays.

### CHARACTERISTICS:

Cultivation & Extraction: Cultivated in Vero cells and extracted using a reliable genomic RNA method.

Quality Control: Subjected to rigorous testing to ensure batch consistency and reliability.

Packaging: Provided in a protective, heat-sealed foil pouch with silica gel.

### PERFORMANCE CHARACTERISTICS:

Sensitivity: Optimized for high sensitivity in PCR detection.

Specificity: Exclusively detects Influenza A H1 RNA.

Reproducibility: Consistent results across different test runs and batches.

**PACK CONTENT :**

Components	Pack Size	Note
<b>RNA Control Lyophilized</b> (1.25x10 <sup>4</sup> to 2.00x10 <sup>4</sup> Copies/μL)	1 Vial	Quantified by Real-Time PCR.
<b>Buffer</b>	500 μL	Water, DNase, RNase Free.

**PRECAUTIONS:**

- For research use or diagnostic control purposes only.
- Follow standard laboratory practices and use PPE.
- Use [Sterile Filter Tips](#) to prevent contamination.
- Handle infectious specimens with precautions.
- Disinfect work surfaces.
- Maintain separate working areas for testing.
- Dispose of unused reagents and waste as per regulations.
- May contain genetic material, handle with care.
- No method ensures total absence of infectious agents.
- Follow local waste disposal rules.
- Dilute below 10<sup>3</sup> copies/μL immediately before use.
- Avoid freezing for low-copy dilutions to prevent RNA degradation.

**INSTRUCTIONS FOR USE:**

- Open the foil pouch of AffiCHECK® Influenza A H1 RNA.
- Centrifuge AffiCHECK® Influenza A H1 RNA Lyophilized Powder at 1000 g for 1 minute.
- Add 100 μL of the Buffer and thoroughly mix until fully reconstituted.
- The concentration will be 1.25x10<sup>4</sup> to 2.00x10<sup>4</sup> Copies/μL once reconstituted.
- Vortex for 1 min for complete dissolution and homogenization.

**RECOMMENDED STORAGE:**

- Before Reconstitution: Store at 2-8°C.
- After Reconstitution: Store at -20°C.
- Shipping: No special instructions
- Shelf Life: 12 months from production date.

**ORDERING INFORMATION :**

Catalog Number	Description	Size
AFG-CHK-0510	AffiCHECK® Influenza A H1 RNA PCR Qualitative Positive Control	1 vial + 500 μL Buffer