

TEST·INSPECTION REPORT

**Residual Sterilant on Tyvek® Validation**

**Manufacturer Name:** Plasmapp Co., Ltd.

**Representative:** Youbong, LIM

**Location:** BVC-111, 125, Gwahak-ro, Yuseong-gu, Daejeon, 34141, Republic of Korea

**Product Name:** Low temperature plasma sterilizer and sterilant

**Brand Name:** STERLINK™ and STERMATE™

**Model Name:** STERLINK mini and STERLOAD™ mini

**Serial Number:** M07BUH027A and SM21J002

**Test·Inspection Item:** Residual Sterilant on Tyvek® Validation

**Testing Laboratory:** Plasmapp Research Institute

**Location:** 372, Dongbu-daero, Osan-si, Gyeonggi-do, 18151, Republic of Korea

**Tested until:** 13 Oct. 2021

**Issued Date:** 15 Oct. 2021

**Test manager:** 전현경  
Hyun Jeong, JEON  
Senior researcher

**Approver:** [Signature]  
Youbong, LIM  
Chief Executive Officer

**Plasmapp Research Institute**



## Residual Sterilant on Tyvek® Validation

### 1. Test schedule

1.1 Date of test beginning: 06 Oct. 2021

1.2 Date of test completion: 13 Oct. 2021

### 2. Test article

Low temperature plasma sterilizer (STERLINK mini, S/N: M07BUH027A)

Sterilant (STERLOAD™ mini, Lot No.: SM21J002)

### 3. Test guideline

3.1 The tests were performed in accordance with the internal test standard.

#### 3.2 Information of testing materials

##### 3.2.1 Test sample

|                        |   |
|------------------------|---|
| <b>Product</b>         | Sterilization Roll Made with Tyvek®                               |
| <b>Manufacturer</b>    | SIGMA Medical Supplies Corp.                                      |
| <b>Lot Number</b>      | 912A  |
| <b>Expiration date</b> | Jan. 2023   |
| <b>Comments</b>        | The sterilization wrap which has sterilization process indicator. |

##### 3.2.2 Validation load\*

| Test sample<br>(Width × Length, cm) | Validation load          |                    |
|-------------------------------------|--------------------------|--------------------|
|                                     | Medical devices          | Total weight [lbs] |
| Six sealed test sample<br>(10 × 10) | Stainless steel scissors | 1.54               |

##### 3.2.3 Hydrogen peroxide test strip

| Item            | Details   |
|-----------------|---|
| Product         | WaterWorks™ Low Range Peroxide Check  |
| Manufacturer    | Industrial Test Systems, Inc.   |
| Lot Number      | 060321H   |
| Expiration date | May. 2023   |
| Comments        | It can detect the hydrogen peroxide from 0.05-4.0 ppm.<br>(0.05, 0.3, 0.5, 1.0, 2.0, 4.0 ppm) |

\*Figure of the validation load was referred to the Appendix 1.

### 3.2.4 Sterilant

|                        |                                      |
|------------------------|--------------------------------------|
| <b>Brand</b>           | STERMATE™                            |
| <b>Model</b>           | STERLOAD™ mini                       |
| <b>Lot Number</b>      | SM21J002                             |
| <b>Expiration date</b> | Sep. 2022                            |
| <b>Manufacturer</b>    | Plasmapp Co., Ltd.                   |
| <b>Comments</b>        | Sterilant cassette for STERLINK mini |

### 3.2.5 Measuring instrument

| <b>Equipment</b>             | <b>Manufacturer</b> | <b>Model</b> | <b>Internal S/N</b> | <b>Calibration date</b> |
|------------------------------|---------------------|--------------|---------------------|-------------------------|
| High temperature data logger | MADGETECH           | HiTemp140    | PO-C-025            | 16 Feb. 2021            |
| Pressure data logger         | MADGETECH           | PR140        | PQ-C-031            | 10 Feb. 2021            |

### 3.3 Test methods

- (1) The sterilization roll was cut into 6 pieces each 10 × 10 (Width × Length, cm) size and sealed as test samples.
- (2) The validation load was placed in the chamber as shown Figure 1.1 (a) of Appendix 1.
- (3) Six sealed test samples were put on the validation load in the chamber.
- (4) The prepared validation load and the test samples were processed with full cycle sterilization of the chamber mode. The temperature and pressure during the sterilization process was measured.
- (5) After sterilization cycle, each test sample was left in the chamber with the chamber lid open for 0, 10, 30, 60, 120 and 300 seconds, and then removed from the chamber and put in each bottle which has 250 ml distilled water.
- (6) The bottle was shaken for 10 seconds to elute the hydrogen peroxide remaining in test sample.
- (7) One hydrogen peroxide test strip was dipped into the water sample in the bottle and shaken gently back and forth for 5 seconds.
- (8) The test strip removed from the water sample and shaken once briskly to remove excess water.
- (9) After 30 seconds, the color of the test strip was match to the color chart, and the hydrogen peroxide concentration was recorded.
- (10) For the negative control, the test strip which was dipped in the distilled water in the bottle without test sample was used.
- (11) The validation test except the negative control test was repeated three times.

## 4. Test results\*\*

### 4.1 Results of the validation test

| Test number | Aeration time [s] | Hydrogen peroxide concentration [ppm, mg·L <sup>-1</sup> ] |
|-------------|-------------------|--|
|             | Negative control  | < 0.05   |
| 1           | 0                 | < 0.3  |
|             | 10                | < 0.3  |
|             | 30                | ≤ 0.05   |
|             | 60                | ≤ 0.05   |
|             | 120               | < 0.05   |
|             | 300               | < 0.05   |
|             | 0                 | ≤ 0.3  |
| 2           | 10                | ≤ 0.3  |
|             | 30                | < 0.3  |
|             | 60                | < 0.3  |
|             | 120               | ≤ 0.05   |
|             | 300               | ≤ 0.05   |
|             | 0                 | ≤ 0.3  |
|             | 10                | < 0.3  |
| 3           | 30                | < 0.3  |
|             | 60                | < 0.3  |
|             | 120               | ≤ 0.05   |
|             | 300               | < 0.05   |

### 4.2. Pressure parameter data during sterilization process [Torr]

| Test number | Sterilization phase 1           |   | Sterilization phase 2           |   | Purification                          |
|-------------|---------------------------------|---|---------------------------------|---|---------------------------------------|
|             | Base <sup>a</sup><br>(< 3 Torr) | Diffusion <sup>b</sup><br>(20 - 100 Torr) | Base <sup>a</sup><br>(< 3 Torr) | Diffusion <sup>b</sup><br>(20 - 100 Torr) | Final base <sup>c</sup><br>(< 3 Torr) |
| 1           | 0.78                            | 43.7                                      | 0.89                            | 46.5                                      | 0.58                                  |
| 2           | 0.58                            | 46.7                                      | 0.78                            | 50.2                                      | 0.70                                  |
| 3           | 0.69                            | 47.6                                      | 1.17                            | 46.5                                      | 1.23                                  |

\*\*The related figures were referred to the Appendix 2. The time evolution of pressures and temperature during sterilization process inside the chamber were described in the Appendix 3, as well.

<sup>a</sup>The base pressure just before injection of the sterilant.

<sup>b</sup>The diffusion pressure after diffusion of the sterilant which is complete.

<sup>c</sup>The base pressure after injection and purification.

#### 4.3. Temperature parameter data during sterilization process [°C]

| Test number | Sterilization phase 1          |                                   |                                       | Sterilization phase 2          |                                   |                                       |
|-------------|--------------------------------|-----------------------------------|---------------------------------------|--------------------------------|-----------------------------------|---------------------------------------|
|             | Load <sup>d</sup><br>(40-60°C) | Chamber <sup>e</sup><br>(55-60°C) | Vaporizer <sup>f</sup><br>(110-130°C) | Load <sup>d</sup><br>(40-60°C) | Chamber <sup>e</sup><br>(55-60°C) | Vaporizer <sup>f</sup><br>(110-130°C) |
| 1           | 49.6 – 58.2                    | 56.1 – 57.7                       | 112 – 124                             | 52.7 – 59.1                    | 57.6 – 59.3                       | 113 – 124                             |
| 2           | 45.8 – 56.8                    | 57.2 – 58.6                       | 115 – 122                             | 48.2 – 58.7                    | 57.2 – 58.4                       | 116 – 127                             |
| 3           | 47.3 – 57.2                    | 57.3 – 58.2                       | 119 – 123                             | 48.8 – 59.2                    | 56.9 – 59.1                       | 115 – 121                             |

#### 4.4. Time parameter data during sterilization process [s]

| Test number | Sterilization phase 1<br>(300 ± 1 s) | Sterilization phase 2<br>(300 ± 1 s) |
|-------------|--------------------------------------|--------------------------------------|
| 1           | 300                                  | 300                                  |
| 2           | 300                                  | 300                                  |
| 3           | 300                                  | 300                                  |

## 5. Conclusions

- (1) As a result of the validation test, when the aeration time was 2 min or more, the concentration of hydrogen peroxide remaining in the test sample was detected as 0.05 ppm or less.
- (2) According to the test results, the aeration time required after the chamber mode sterilization process is at least 2 min.

<sup>d</sup>The load temperature is measured by the temperature data logger described in 3.2.5.

<sup>e</sup>The chamber temperature is controlled by K-type thermocouple.

<sup>f</sup>The vaporizer temperature is controlled by K-type thermocouple.

## Appendix 1

### 1. Residual sterilant on Tyvek® validation test




Figure 1.1 (a) The position of validation load and data loggers in the chamber and (b) the test samples were placed on the validation load.

## Appendix 2

### 1. Results of the residual sterilant on Tyvek® validation test




Figure 2.1 Color change result of hydrogen peroxide strip in (a) negative control and (b-d) three validation tests. It is a test sample aeration for 0, 10, 30, 60, 120, and 300 seconds in order from the top in each validation test figure.

### Appendix 3

#### 1. Pressure and temperature curves of the residual sterilant on Tyvek® validation test




Figure 3.1 (a) The whole plot of pressure and temperature curve during the full cycle and (b) magnified plot of diffusion phase for test number 1.




Figure 3.2 (a) The whole plot of pressure and temperature curve during the full cycle and (b) magnified plot of diffusion phase for test number 2.




Figure 3.3 (a) The whole plot of pressure and temperature curve during the full cycle and (b) magnified plot of diffusion phase for test number 3.