


Analysis of free fatty acids

BP21 (FFAP)

Column part number	054474		
Phase	BP21 (FFAP)	Detector	FID, 300°C
Column	25 m x 0.53 mm x 0.5 µm	Injection mode	Split (100:1)
Initial temperature	90°C, 1 min	Carrier gas	He, 5 mL/min
Rate	8°C/min	Injection volume	1 µL
Final temperature	160°C		


 For more information visit www.trajanscimed.com or contact techsupport@trajanscimed.com