





Important

⚠ Caution

Particulate matter will cause the valve to become stiff or seize.

Do not soak entire valve for long period of time.

Instructions

Valve type	To fit syringe	Description	Needle replacement
Push-pull valve with needle <i>Valve shown in OPEN position</i>	25 µL - 2.5 mL	V25/2.5M-0.63C* V25/2.5M-0.47C*	0.63 cone 0.63 side-hole dome 0.47 cone
			
Push-pull valve with Luer lock <i>Valve shown in CLOSED position</i>	25 µL - 2.5 mL	VLL25/2.5M*	Luer lock needles
			
Push-button valve with Luer lock <i>Valve shown in OPEN position</i>	5 mL - 2 L	VLLMA	Luer lock needles
			
Push-button valve for Luer lock syringes <i>Valve shown in OPEN position</i>	50 µL - 2 L Luer tip and Luer lock	SLLV	Luer lock needles
			

* Replacement valve seal assembly available.

Clean and care

Clean valve immediately after use by rinsing with suitable solvent (usually acetone), then dry.

Warranty

Trajan valves are warranted to be free of material and workmanship defects. No other warranty is expressed or implied. Any valve failing due to such defects will be replaced free of charge provided proof is given.

Do not return any valve contaminated with radioactive, infectious or hazardous materials

Information and support

Visit www.trajanscimed.com or contact techsupport@trajanscimed.com

Specifications are subject to change without notice.