

Operational hints for BPX70

To fully utilize the high thermal stability of the BPX70 GC column, Trajan Scientific and Medical recommends the use of helium as the carrier gas when operating the column for extended periods above 220/230°C.

MN-0121-A_RevE 11/2016

TRAJAN

www.trajanscimed.com