



Operational hints for BPX90

To fully utilize the high thermal stability of the BPX90 GC column, Trajan Scientific and Medical recommends the use of helium as the carrier gas when operating the column for extended periods above 220/230°C.

MN-0559-C_RevD 11/2020