



Coverslips

Intended use	For in vitro diagnostic use (European Directive 98/79/EC – IVDD, dated 27th October 1998). Coverslips are intended to cover the specimen mounted on microscope slide for better viewing and to protect them from contamination by airborne particles or other substances.
Product	Coverslips – circular or rectangular. Provided in trays of 900, 1,000 or 2,000. Packaged in plastic boxes with lids for protection.
Precautions	<ul style="list-style-type: none"> • Single use – do not reuse. • Beware of sharp glass edges. • Trained personnel and professional use.
Storage instructions	<ul style="list-style-type: none"> • Store the product in original packaging, in a dry environment, preferably protected from direct sources of heat and light. • Store at room temperature – Recommended storage temperature between 2°C and 30°C. • Avoid large variations in temperature conditions during storage as excessive cooling can lead to condensation. • Oldest coverslip trays should be used first. • Do not use after use-before date. • Shelf-life is indicated by the use-before date, printed on the outer packaging. • Special tropical packaging provides additional protection against environmental factors, however if the product is stored under conditions other than those specified, these conditions must be verified by the user. • There are no apparent signs to suggest instability or diminished performance of the product. If suboptimal performance is noticed with no apparent reason and cannot be attributed to variations in laboratory procedures, please contact your regional Trajan representative.
Coverslip preparation	<ul style="list-style-type: none"> • Tissue section of approximately 2-5 µm floating on pre-heated (max 45°C) distilled water bath. • Mount the tissue section carefully on the slide, making sure it is properly spread. • Hold the coverslip at a 45° angle to the surface of the slide, and allow the bottom edge to touch the drop of mounting media. When the drop has spread along the edge of the coverslip, let go of the coverslip and allow the medium to spread slowly (20-30 seconds). • For automated coverslipping, follow directions for use with the automated coverslipping instrument.
Compatible instruments	Trajan coverslips for coverslippers are compatible with most automated instruments used for coverslipping.
Product conformity	ISO 8255/1:1986 Optics and optical instruments - Microscope cover glass - Part 1: Dimensional tolerances, thickness and optical properties.
Compliance certification	 
Quality Management System	Trajan selects manufacturers that operate a Quality Management System, independently certified to ISO 9001 and or ISO 13485.

Technical data	
Thickness	0.13-0.16 mm
Luminous Transmittance (d = 0.15 mm)	91.7
Refractive Index n_d	1.5230
Refractive Index n_e	1.5255
Density	2.51 g/cm ³

For more information about this product visit www.trajanscimed.com or contact techsupport@trajanscimed.com

Australia and New Zealand

Trajan Scientific Australia Pty Ltd
7 Argent Place, Ringwood
Victoria 3134, Australia
Toll Free: 1800 257 213
Tel Melbourne: +61 (0) 3 9874 8577
Tel Sydney: +61 (0) 2 9609 1755
Fax: +61 (0) 2 9609 1411
anz@trajanscimed.com

Americas

Trajan Scientific Americas Inc
Wells Branch Tech Center - Austin TX
1421 W Wells Branch Pkwy # 108
Pflugerville, Texas 78660, USA
Toll Free: 800 945 6154
Tel: 512 837 7190
Fax: 512 836 9159
usa@trajanscimed.com

Europe, Middle East, Africa and India

Trajan Scientific Europe Ltd
Crownhill Business Centre
14 Vincent Avenue
Crownhill, Milton Keynes
MK8 0AB, United Kingdom
Tel: +44 (0) 1908 568 844
Fax: +44 (0) 1908 566 790
europe@trajanscimed.com

Asia

Trajan Scientific Japan Inc
RK CUBE 3F, 6-85 Otamachi
Naka-ku, Yokohama-shi
Kanagawa 231-0011, Japan
Tel: +81 (0) 45 222 2885
Fax: +81 (0) 45 222 2887
japan@trajanscimed.com