

Fused silica nanospray emitters

Tapered Tips

| Intended use | Trajan emitters are optimized for high resolution with extended life in a wide variety of mass spectrometry applications and workflows. Trajan's nanospray emitters (ID: $10 - 20 \mu m$) are designed specifically for low-flow applications and have been tested extensively under typical proteomics conditions. | | | | | |
|------------------------------|--|--|--|--|--|--|
| Product | Trajan emitters are compatible with common mass spectrometry workflow conditions including operating voltages, flow-rates, and solvents. A variety of lengths, tip geometries, internal diameters, and metal-coating options are available to best suit your specific workflow requirements and instrument configurations. | | | | | |
| Precautions | Caution should be taken to avoid injury (especially to the eyes) when working with fused silica (quartz). It is recommended to use safety glasses and avoid unnecessary bending or breaking of the fused silica. In the event of an emitter breakage or emitter replacement, observe local requirements for disposal of 'broken glass'. Narrow-bore emitters can lead to higher back-pressure in the system if partial blockages occur; be careful not to over-pressurize the system as the emitter may dislodge from the fitting/connection. Overtightening of the fused silica emitter can cause fractures, leading to higher risk of dead-volume and instrument contamination. | | | | | |
| Storage instructions | Packaging should be stored at room temperature and ideally in a clean environment Recommended to close the packaging lid when not actively retrieving an emitter to avoid contamination of the emitters with airborne dust and debris. | | | | | |
| Slide preparation | Caution should be taken to avoid injury (especially to the eyes) when working with fused silica (quartz). It is recommended to use safety glasses and avoid unnecessary bending or breaking of the fused silica. In the event of an emitter breakage or emitter replacement, observe local requirements for disposal of 'broken glass'. Narrow-bore emitters can lead to higher back-pressure in the system if partial blockages occur; be careful not to over-pressurize the system as the emitter may dislodge from the fitting/connection. Overtightening of the fused silica emitter can cause fractures, leading to higher risk of dead-volume and instrument contamination. | | | | | |
| Compatible instruments | See product brochure for instrument compatibility information. It is recommended to only use emitter tips that correspond to the instrument to avoid instrument damage and/or poor results. | | | | | |
| Quality Management System | Manufactured in a certified ISO 9001 environment. | | | | | |



| Specifications | |
|-----------------|--|
| Material | Fused silica (quartz) with polyimide coating |
| Cutting Method | Precision cleaved |
| Coating Length | 40 mm |
| OD | 360 μm ± 10 μm |
| ID | 10 μm - 30 μm ± 2 μm 30 μm - 50 μm ± 3 μm |
| Length | Lengths available from 30 mm -150 mm |
| Units/Pack | 5 |
| Pack Dimensions | 78 mm x 75 mm x 17mm |
| Disposal | Refer to local regulations. Please recycle the packaging where possible. |

| Technical Data | | | | | | | |
|----------------|-------|--------|-----------------|------------|-----------------------|---|--|
| Part Number | Angle | OD | ID | L | Metalized Coating | Compatible With | |
| 063200000 | - 13° | 360 µm | 10 µm | 50 mm | Uncoated | PharmaFluidics µPAC[™] Flex iON Connect ESI-MS interface | |
| 063220000 | | | | | Tip-Coated (40 mm) | Thermo Scientific Nanospray Flex[™] ion Source PharmaFluidics µPAC Flex iON Connect ESI-MS interface | |
| 063240000 | | | | | Distal-Coated (40 mm) | Thermo Scientific Nanospray Flex ion Source Thermo Scientific FAIMS Pro[™] Interface PharmaFluidics µPAC Flex iON Connect ESI-MS interface | |
| 063200102 | 13° | 360 µm | 20 µm | 40 mm | Uncoated | Thermo Scientific Nanospray Flex ion Source | |
| 063200100 | - 13° | 360 µm |) μm 20 μm 50 ι | 50 mm | Uncoated | PharmaFluidics µPAC Flex iON Connect ESI-MS interface | |
| 063220100 | | | | | Tip Coated (40 mm) | Thermo Scientific Nanospray Flex ion Source PharmaFluidics µPAC Flex iON Connect ESI-MS interface | |
| 063240100 | | 500 μm | | | Distal Coated (40 mm) | Thermo Scientific Nanospray Flex ion Source Thermo Scientific FAIMS Pro Interface PharmaFluidics µPAC Flex iON Connect ESI-MS interface | |
| 063200101 | 13° | 360 µm | 20 µm | 62.5 mm | Uncoated | Waters NanoFlow [™] Source Waters NanoLockSpray [™] Source Waters Z-spray [™] Source Waters Micromass [™] Source | |

Contact us about custom emitter lengths and geometries for your specific workflows

For more information about this product visit www.trajanscimed.com or contact techsupport@trajanscimed.com



www.trajanscimed.com/emitter-tips



Page 2 of 02