



Manufacturing Chemists Since 1953  
eCAMCO.com

## CERTIFICATE OF ANALYSIS

**Product: Camco Quik Stain II, #211**

**Lot Number:** 6071      **Exp. Date:** 2019-12-12

### Analytical Report:

|                   | Specification                      | Value Found | Remarks           |
|-------------------|------------------------------------|-------------|-------------------|
| Specific Gravity: | 0.77 to 0.87                       | <u>0.83</u> | <u>Pass</u> /Fail |
| Performance:      | Practical laboratory test results: | <u>Pass</u> | <u>Pass</u> /Fail |

This product is manufactured according to the production standards and procedures of Cambridge Diagnostic Products, Inc. and supervised by qualified personnel.

  
Quality Assurance Department

12-22-16  
Date