

# PDT INSTRUMENT CLEANING SOLUTION RECOMMENDATIONS

The following solutions have been found to be either safe or destructive to resin-handled instruments, including those manufactured by PDT. Please follow all manufacturer guidelines for solution strength and instrument exposure times. All cleaning solutions must be thoroughly rinsed from the instruments prior to autoclaving.

Product Name	Manufacturer	Safe?
Sani-treet Green	Enzyme Industries, Inc.	Yes
Tri-Clean - Triple Enzyme Cleaner	BrandMax	Yes
BioSonic Super Concentrate	Coltene/Whaldent	Yes
Euro Sept Max	Henry Schein	Yes
IMS Daily Clean	Hu Friedy	Yes
ReSurge Instrument Cleaning Solution	Sultan Healthcare	Yes
Maxicide Plus	Henry Schein	No
Maxizyme Tabs	Henry Schein	No
VibraKleen E2	Heraeus Kulzer	No
Gigazyme	Schülke & Mayer	No
ProEZ Foam	Certol International	No
Diamond Zyme	Kerr Dental	No



**PARADISE**  
DENTAL TECHNOLOGIES